

**ZOOM ® View™****Report Created 07 Jul 2021****Settings Changes Report**

Date of Birth      N/R N/R N/R  
Device              ACCOLADE MRI L310/836133

Last Office Interrogation  
**N/R**  
Implant Date  
**N/R**

Status	Parameter	Old Value	New Value
Pending	Rate Hysteresis Offset	Off	-10 min <sup>-1</sup>
Pending	Rate Search Hysteresis	---	Off
Pending	Rate Smoothing Up	---	Off
Programmed	Normal Brady Amplitude	Trend 3.5 V	Trend 3.0 V
Programmed	Normal Brady V-Sensitivity	Fixed 2.5 mV	Fixed 2.0 mV

Clinician Signature: