

	ZOOM® View™	Report Created 11 Dec 2020
	Device Settings Report	Last Office Interrogation 10-JUL-2013 Implant Date N.R. N.R. N.R.
	Device ALTRUA 50 S508/576134	

Programming

Last Programmed 09-JUL-2013

Tachy

Tachy Detection On
Rate 170 min⁻¹
Duration 8 cycles

Tachy Response

Ventricular Rate Regulation Off
Maximum Pacing Rate -- min⁻¹

Brady

Settings

Mode SSI
Lower Rate Limit 80 min⁻¹
Maximum Tracking Rate -- min⁻¹
Refractory 250 ms
Magnet Response ASYNC

Rate Enhancements

Rate Smoothing
Up -- %
Down Off %
Maximum Pacing Rate -- min⁻¹
Rate Hysteresis
Hysteresis Offset Off min⁻¹
Search Hysteresis -- cycles

Output

5.0 V @ 0.60 ms
1.0 mV

Sensitivity

Leads

Bipolar

Pace --
Sense --
Safety Switch On

Sensor(s)

Setup

Arrhythmia Logbook Setup 7 EGMs @ 7/7
Magnet (Patient Triggered) Off
Trending Setup Rate
Recording Method Long Duration
Duration 68.2 Hours
Setup Method Continuous