



ZOOM® View™

QUICK NOTES® Report

Report Created 04 Feb 2019

Date of Birth N/R N/R N/R
Device COGNIS 100-D P107/009600
Tachy Mode Monitor + Therapy

Last Office Interrogation
26 Jun 2012
Implant Date
N/R

My Alerts

- ◆ 05 Oct 2018 Battery capacity depleted
▲ 04 Feb 2019 14:52 Check RV, Atrial, LV and Shock Leads

Battery

- ◆ Battery capacity depleted on 05 Oct 2018
Charge Time 13.7 s
Last Capacitor Re-form 16 May 2018 15:23

One Year Remaining

Explant



Settings

Ventricular Tachy Settings

VF 165 min⁻¹ 41Jx8

Brady/CRT Settings

Mode	VVI	Pacing Output	
Lower Rate Limit	50 min ⁻¹	Right Ventricular	2.5 V @ 0.4 ms
RV-Refractory (RVRP)	250 ms	Left Ventricular	2.5 V @ 0.4 ms
LV-Refractory (LVRP)	250 ms	Sensitivity	
Ventricular Pacing Chamber	BiV	Atrial	AGC 0.25 mV
LV Offset	0 ms	Right Ventricular	AGC 0.6 mV
		Left Ventricular	AGC 0.8 mV
		Leads Configuration (Pace/Sense)	
		Atrial	Bipolar
		Right Ventricular	Bipolar
		Left Ventricular	LVtip>>RV