



ZOOM ® View™

QUICK NOTES ® Report
HOVADÍK 510404, BronislavDate of Birth 4 Apr 1951
Device CHARISMA EL ICD D332/243151
Tachy Mode Monitor + Therapy

Report Created 12 Nov 2018

Last Office Interrogation
08 Nov 2018
Implant Date
8 Nov 2018**My Alerts**

There are no alerts to display.

Events Since Last Reset (08 Nov 2018)

There are no events to display.

Battery OK

Approximate time to explant: > 12 years

Charge Time N/R s

Last Capacitor Re-form N/R

One Year Remaining

Explant

**Leads Data****Implant**

8 Nov 2018

Previous Session**Most Recent****Ventricular**

Intrinsic Amplitude N/R mV

4.9 mV

5.6 mV @ 75 min⁻¹

Pace Impedance N/R Ω

505 Ω

459 Ω

Pace Threshold N/R V @ N/R ms

0.6 V @ 0.4 ms

0.7 V @ 0.4 ms

Shock Vector

Shock Impedance N/R Ω

46 Ω

48 Ω

Brady Counters Since Last Reset (08 Nov 2018)**Counters**

Ventricular <1% Paced

Settings**Ventricular Tachy Settings**VF 220 min⁻¹ ATP 36J, 41J, 41Jx6VT 170 min⁻¹ Burst Ramp/Scan 26J, 31J, 41Jx4VT-1 150 min⁻¹ Monitor Only**Brady Settings**

Mode VVI

Pacing Output

Lower Rate Limit 50 min⁻¹

Ventricular

2.6 V @ 0.4 ms

V-Refractory (VRP) 250 ms

Sensitivity

Ventricular

AGC 0.6 mV

Leads Configuration (Pace/Sense)

Ventricular

Bipolar