

LATITUDE™ Programming System Report Created 24 Jul 2021

**QUICK NOTES ® Report** 

Date of Birth N/R N/R N/R

Device ACCOLADE MRI L310/ 839486

Last Office Interrogation

N/R

Implant Date N/R

My Alerts

There are no alerts to display.

Events Since Last Reset (23 Jul 2021)

24 Jul 2021 02:39 RV Auto

Battery OK

Approximate time to explant: > 8 years
Magnet Rate 100 min<sup>-1</sup>

100 min⁻¹

One Year Remaining

Explant

Leads DataImplant<br/>N/R N/R N/RPrevious Session<br/>N/R N/R N/RMost RecentIntrinsic AmplitudeN/R N/R MVN/R mV13.3 mV@41 min^-1

Pace Impedance  $N/R \Omega$   $N/R \Omega$   $N/R \Omega$   $N/R \Omega$   $N/R \Omega$  Auto 0.7 V @ 0.4 ms

POST completed. Test started at 23 Jul 2021 09:42

Brady Counters Since Last Reset (23 Jul 2021)

Counters

% Paced 74

Settings

**Tachy Settings** 

Tachy EGM Storage On
Detection Rate 160 min<sup>-1</sup>

**Brady Settings** 

Mode VVIR Pacing Output

Lower Rate Limit 55 min<sup>-1</sup> Auto 1.2 V @ 0.4ms

Maximum Sensor Rate 130 min<sup>-1</sup> Sensitivity

Refractory 250 - 250 ms AGC 0.6 mV

Leads Configuration (Pace/Sense)

Bipolar

3869 Software Version: 1.05 L310 Firmware Version: F\_v1.00.00(1.10) © 2014-2017
Boston Scientific Corporation
or its affiliates. All rights reserved.
Page 1 of 1

Clinician Signature: