



ZOOM® View™

Device Settings Report

Device

ALTRUA 50 S501/590330

Report Created 18 May 2020

Last Office Interrogation
20-JAN-2020
Implant Date
N.R. N.R. N.R.

Programming

Last Programmed 21-JAN-2019

Tachy

Tachy Detection On
Rate 170 min⁻¹
Duration 8 cycles

Tachy Response

Ventricular Rate Regulation Off
Maximum Pacing Rate -- min⁻¹

Brady

Settings

Mode	SSIR	Output	AUTO 1.6 V @ 0.40 ms
Lower Rate Limit	55 min ⁻¹	Sensitivity	AUTO 1.48 mV
Maximum Tracking Rate	-- min ⁻¹	Leads	Bipolar
Maximum Sensor Rate	130 min ⁻¹	Pace	--
Refractory	250 ms	Sense	--
Magnet Response	ASYN	Safety Switch	On

Rate Enhancements

Rate Smoothing		Sensor(s)	On
Up	Off %	Accelerometer	On
Down	Off %	Response Factor	AUTO 11
Maximum Pacing Rate	-- min ⁻¹	Activity Threshold	Medium
Rate Hysteresis		Reaction Time	30 sec
Hysteresis Offset	-5 min ⁻¹	Recovery Time	5 min
Search Hysteresis	Off cycles	Age	60
		Gender	Male
		Auto Response	On
		ACC Initial Response Factor	8
		Sensor Rate Target	110 min ⁻¹

Setup

Arrhythmia Logbook Setup 7 EGMs @ 7/7
Magnet (Patient Triggered) Off
Trending Setup Rate
Recording Method Long Duration
Duration 34.1 Hours
Setup Method Continuous