

ZOOM ® View™			Report Created 08 Nov 2016	
	Selected Episodes Report			
KRÍZ, Karel			Last Office Interrogation	
	Date of Birth	30 Aug 1954	31 Aug 2016	
	Device	AUTOGEN X4 CRT-D G179/136339	Implant Date	

12 Oct 2015

Monitor + Therapy

Tachy Mode

Event V-43: 01 Sep 2016 17:07

VF Event Onset
Avg A Rate: 207 min⁻¹
Avg V Rate: 270 min⁻¹
Detection: Rhythm ID
Template: 27 Jul 2016 12:12

RhythmMatch™ Threshold: 94 %


At V-Detect

Avg A Rate: 166 min⁻¹ Avg V Rate: 303 min⁻¹ Rate Zone: VF Stability: (79 ms, Off) V>A Rate: (True, Off) AFib: (True, Off)

RhythmID Correlated: False RhythmMatch™: Too Fast SRD Met: (False, Off) ATP Timeout: False Attempt 1, 36 J V Shock Elapsed Time: 00:00:02 Shock Information: Charge Time: 5.9 s

Lead Impedance: 62 Ω Lead Polarity: Initial


Event Ended 00:00:42





2868 Software Version: 3.07 G179 Firmware Version: E_v1.02.00(2.01) © 2014


Boston Scientific Corporation
or its affiliates. All rights reserved.


Page 1 of 8

















\S |60

