Scientific

ZOOM ® Vie	w TM	Report Created 21 Mar 2017
QUICK NOTES ® Report		Last Office Interrogation
		13-JUL-2013
		Implant Date
Device	ALTRUA 50 S508/576134	N.R. N.R. N.R.

Events Since Last Reset

Tach at 190 min⁻¹ 09-JUL-2013 12:36 09-JUL-2013 12:29 Tach at 280 min⁻¹ See last page for full list of Events Since Last Reset.

Good

Longevity Remaining >5.0 years

@ Current Pacing Percentage(s)

Magnet Rate 100 min⁻¹ Elective Replacement (ERT) End of Life (EOL)

Beginning of Life (BOL)

Leads Data Implant **Previous Session Present Session** N.R. N.R. N.R. 13-JUL-2013 N.R. mV Intrinsic Amplitude 3.3 mV N.R. mV Pace Impedance N.R. Ω 560 Ω 490 Ω Pace Threshold N.R. V 1.3 V @ 0.6 ms

Brady Counters Since Last Reset

87 % Paced

Settings

Tachy Settings

Tachy Detection On Rate 170 min⁻¹ Duration 8 cycles

Brady Settings

Pacing Output 0.1 V @ 0.05 ms Mode SSI Sensitivity 2.0 mV Lower Rate Limit 30 min⁻¹ Lead Configuration Bipolar Maximum Tracking Rate -- min⁻¹ Pace Refractory 250 ms Sense

All Events Since Last Reset

09-JUL-2013 12:36 Tach at 190 min⁻¹ 09-JUL-2013 12:29 Tach at 280 min⁻¹ 09-JUL-2013 12:25 Tach at 240 min⁻¹ 09-JUL-2013 12:25 Tach at 260 min⁻¹ Tach at 220 min⁻¹ 09-JUL-2013 12:23

2892 Software Version: 1.00 S508 Firmware Version: 0a

© 2010 **Boston Scientific Corporation** or its affiliates. All rights reserved. Page 1 of 1

Clinician Signature: