



ZOOM® View™

## QUICK NOTES® Report

Device

ALTRUA 50 S501/590330

Report Created 06 Mar 2019

Last Office Interrogation  
21-JAN-2019  
Implant Date  
N.R. N.R. N.R.


## Events Since Last Reset (22-JAN-2018)

There are no events to display.

## Battery Good

Longevity Remaining &gt;5.0 years

@ Current Pacing Percentage(s)

Magnet Rate 100 min<sup>-1</sup>Elective Replacement (ERT)  
End of Life (EOL)Beginning of Life  
(BOL)

## Leads Data

## Implant

## Previous Session

## Present Session

N.R. N.R. N.R.

21-JAN-2019

Intrinsic Amplitude

N.R. mV

9.9 mV

7.4 mV

Pace Impedance

N.R. Ω

480 Ω

460 Ω

Pace Threshold

N.R. V

1 V @ 0.4 ms

## Brady Counters Since Last Reset (22-JAN-2018)

27 % Paced

## Settings

## Tachy Settings

Tachy Detection

On

Rate

170 min<sup>-1</sup>

Duration

8 cycles

## Brady Settings

Mode

SSIR

Pacing Output

AUTO 1.5V @ 0.40 ms

Lower Rate Limit

55 min<sup>-1</sup>

Sensitivity

AUTO 3.76 mV

Maximum Tracking Rate

-- min<sup>-1</sup>

Lead Configuration

Bipolar

Maximum Sensor Rate

130 min<sup>-1</sup>

Pace

--

Refractory

250 ms

Sense

--

Clinician Signature: