

ZOOM ® View[™] Report Created 13 Sep 2021

QUICK NOTES ® Report

Hucík, Stepán

Date of Birth

7 Jan 1944

Device INCEPTA CRT-D P162/119874

Tachy Mode Monitor + Therapy

Last Office Interrogation

03 May 2021 Implant Date 12 Mar 2013

My Alerts

▲ 13 Sep 2021 10:23 Approximate time to explant: 7 months

▲ 13 Sep 2021 10:23 Check Atrial and LV Leads

Events Since Last Reset (03 May 2021)

16 Jul 2021 12:09 NonSustV at 175 min⁻¹, Nonsustained 12 Jun 2021 16:37 NonSustV at 127 min⁻¹, Nonsustained

Battery

Approximate time to explant: 7 months
Charge Time 13.8 s

Last Capacitor Re-form 28 Jul 2021 08:41

One Year Remaining

Explant

Leads Data Implant Previous Session Most Recent 12 Mar 2013 Atrial Intrinsic Amplitude N/R mV Paced mV N/R mV Pace Impedance $N/R \Omega$ >2500 Ω >2500 Ω Pace Threshold N/R V @ N/R ms 4.0 V @ 0.4 ms Right Ventricular Intrinsic Amplitude 7.0 mV Paced mV 8.3 mV @ 47 min⁻¹ Pace Impedance 650Ω 469 Ω 467Ω Pace Threshold 0.5 V @ 0.4 ms 0.7 V @ 0.4 ms 0.9 V @ 0.4 ms **Left Ventricular** Intrinsic Amplitude 12.0 mV Paced mV 21.2 mV Pace Impedance 650Ω 584 Ω 584 Ω Pace Threshold 1.3 V @ 0.4 ms 1.7 V @ 0.4 ms 2.0 V @ 0.4 ms **Shock Vector** 45 Ω 55 Ω 55 Ω Shock Impedance

Brady/CRT Counters Since Last Reset (03 May 2021)

Counters

Atrial 0% Paced Right Ventricular 95% Paced Left Ventricular 100% Paced

Atrial Arrhythmia

% AT/AF 0

(One Year of Data)

Settings

Ventricular Tachy Settings

VF 230 min⁻¹ ATP 31J, 41J, 41Jx6 VT 205 min⁻¹ Scan Ramp 31J, 41J, 41Jx4 VT-1 160 min⁻¹ Scan Ramp 41J, 41J, 41Jx3

Brady/CRT Settings

Mode VVIR Pacing Output
Lower Rate Limit 65 min⁻¹ Right Ventricular
Maximum Sensor Rate 130 min⁻¹ Left Ventricular

RV-Refractory (RVRP) 230 - 250 ms Sensitivity

LV-Refractory (LVRP)

250 ms

Atrial

AGC 0.25 mV

Ventricular Pacing Chamber

BiV

Right Ventricular

AGC 0.6 mV

Page 1 of 2

2868 Software Version: 4.08

P162 Firmware Version: B_v1.02.00(4.01)

Boston Scientific Corporation or its affiliates. All rights reserved.

© 2016 Clinician Signature:

2.5 V @ 0.4 ms

3.0 V @ 0.4 ms

Settings (Continued)

Brady/CRT Settings

LV Offset 0 ms Sensitivity

Left Ventricular AGC 1.0 mV

Leads Configuration (Pace/Sense)

Atrial Bipolar/Off
Right Ventricular Bipolar
Left Ventricular LVring>>RV / LVtip>>LVring

2868 Software Version: 4.08 P162 Firmware Version: B_v1.02.00(4.01) © 2016

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 2 of 2

Clinician Signature: