



ZOOM® View™

Selected Episodes Report

KRÍZ, Karel

Date of Birth

30 Aug 1954

Device

AUTOGEN X4 CRT-D G179/136339

Tachy Mode

Monitor + Therapy

Report Created 08 Nov 2016

Last Office Interrogation

31 Aug 2016

Implant Date

12 Oct 2015

Event V-35: 01 Sep 2016 16:35

VF Event Onset

Avg A Rate: 99 min⁻¹Avg V Rate: 242 min⁻¹

Detection: Rhythm ID

Template: 27 Jul 2016 12:12

RhythmMatch™ Threshold: 94 %

At V-Detect

Avg A Rate: 115 min⁻¹Avg V Rate: 233 min⁻¹

Rate Zone: VF

Stability: (5 ms, Off)

V>A Rate: (True, Off)

AFib: (False, Off)

RhythmID Correlated: False

RhythmMatch™: 90 %

SRD Met: (False, Off)

ATP Timeout: False

Attempt 1, 36 J V Shock

Elapsed Time: 00:00:03

Aborted Attempt

VF ATP delivered, no shock attempted due to fail to reconfirm

At V-Detect

Avg A Rate: 109 min⁻¹Avg V Rate: 240 min⁻¹

Rate Zone: VF

ATP Timeout: False

Attempt 2, 36 J V Shock

Elapsed Time: 00:00:16

Shock Information:

Charge Time: 7.2 s

Lead Impedance: 49 Ω

Lead Polarity: Initial

At V-Detect

Avg A Rate: 233 min⁻¹Avg V Rate: 223 min⁻¹

Rate Zone: VT

Stability: 2 ms

V>A Rate: True

AFib: True

SRD Met: False

ATP Timeout: False

Attempt 3, 36 J V Shock

Elapsed Time: 00:00:39

Aborted Attempt

No shock attempted due to fail to reconfirm

Event Ended 00:01:16















