

Tachy Mode

ZOOM ® View™		Report Created 08 Nov 2016		
Selected Episo	des Report			
KRÍZ, Karel		Last Office Interrogation		
Date of Birth	30 Aug 1954	31 Aug 2016		
Device	AUTOGEN X4 CRT-D G179/136339	Implant Date		

Monitor + Therapy

Implant Date

12 Oct 2015

Event V-44: 01 Sep 2016 17:08

**VT Event Onset** Avg A Rate: 144 min<sup>-1</sup> Avg V Rate: 316 min<sup>-1</sup> Detection: Rhythm ID

Template: 27 Jul 2016 12:12 RhythmMatch™ Threshold: 94 %

At V-Detect

Avg A Rate: 231 min<sup>-1</sup> Avg V Rate: 221 min<sup>-1</sup>

Rate Zone: VF

Stability: (115 ms, Off) V>A Rate: (True, Off) AFib: (True, Off)

RhythmID Correlated: False RhythmMatch™: 0 % SRD Met: (False, Off) ATP Timeout: False

At V-Detect

Avg A Rate: 80 min<sup>-1</sup> Avg V Rate: 223 min<sup>-1</sup> Rate Zone: VF Stability: (61 ms, Off) V>A Rate: (True, Off) AFib: (False, Off) SRD Met: (False, Off) ATP Timeout: False

At V-Detect

Avg A Rate: 42 min<sup>-1</sup> Avg V Rate: 324 min<sup>-1</sup>

Rate Zone: VT Stability: 89 ms V>A Rate: True AFib: False SRD Met: False ATP Timeout: False

At V-Detect

Avg A Rate: 33 min<sup>-1</sup> Avg V Rate: 195 min<sup>-1</sup> Rate Zone: VT Stability: 35 ms

V>A Rate: True AFib: False SRD Met: False

2868 Software Version: 3.07

G179 Firmware Version: E\_v1.02.00(2.01)

Attempt 1, 36 J V Shock Elapsed Time: 00:00:06

VF ATP delivered prior to shock

**Shock Information:** Charge Time: 5.8 s Lead Impedance: 62  $\Omega$ Lead Polarity: Initial

Attempt 2, 41 J V Shock

Elapsed Time: 00:00:29 **Shock Information:** Charge Time: 7.3 s Lead Impedance: 61  $\Omega$ Lead Polarity: Initial

Attempt 3, 41 J Max V Shock

Elapsed Time: 00:01:03 **Shock Information:** Charge Time: 7.1 s Lead Impedance: 62  $\Omega$ Lead Polarity: Initial

Attempt 4, 41 J Max V Shock

Elapsed Time: 00:01:42 **Shock Information:** Charge Time: 7.1 s Lead Impedance:  $60 \Omega$ Lead Polarity: Initial

© 2014 **Boston Scientific Corporation** or its affiliates. All rights reserved. Page 1 of 24

# ATP Timeout: False

#### At V-Detect

Avg A Rate: 47 min<sup>-1</sup> Avg V Rate: 279 min<sup>-1</sup>

Rate Zone: VF

Stability: (56 ms, Off) V>A Rate: (True, Off) AFib: (False, Off) SRD Met: (False, Off) ATP Timeout: False

#### At V-Detect

Avg A Rate: 39 min<sup>-1</sup> Avg V Rate: 264 min<sup>-1</sup>

Rate Zone: VF

Stability: (78 ms, Off) V>A Rate: (True, Off) AFib: (False, Off) SRD Met: (False, Off) ATP Timeout: False

#### At V-Detect

Avg A Rate: 46 min<sup>-1</sup> Avg V Rate: 245 min<sup>-1</sup> Rate Zone: VF Stability: (44 ms, Off)

V>A Rate: (True, Off) AFib: (False, Off) SRD Met: (False, Off) ATP Timeout: False

#### At V-Detect

Avg A Rate: 42 min<sup>-1</sup> Avg V Rate: 225 min<sup>-1</sup> Rate Zone: VT

Stability: 152 ms V>A Rate: True AFib: False SRD Met: False ATP Timeout: False

# At V-Detect

Avg A Rate: 38 min<sup>-1</sup> Avg V Rate: 236 min<sup>-1</sup> Rate Zone: VF ATP Timeout: False

### At V-Detect

Avg A Rate: 43 min<sup>-1</sup> Avg V Rate: 195 min<sup>-1</sup> Rate Zone: VF

2868 Software Version: 3.07

G179 Firmware Version: E\_v1.02.00(2.01)

### Attempt 5, 41 J Max V Shock

Elapsed Time: 00:02:18 Shock Information: Charge Time: 7 s Lead Impedance:  $64 \Omega$  Lead Polarity: Initial

### Attempt 6, 41 J Max V Shock

Elapsed Time: 00:02:35 **Shock Information:** Charge Time: 7.1 s Lead Impedance: 57  $\Omega$  Lead Polarity: Initial

### Attempt 7, 41 J Max V Shock

Elapsed Time: 00:03:04 **Shock Information:** Charge Time: 7 s Lead Impedance: 61  $\Omega$  Lead Polarity: Initial

### Attempt 8, Non Therapy V

Elapsed Time: 00:03:22
Attempt Information:

All programmed therapy has been delivered.

#### Attempt 9, 41 J Max V Shock

Elapsed Time: 00:03:25 **Shock Information:** Charge Time: 7.1 sLead Impedance:  $63 \Omega$ Lead Polarity: Reversed

# Attempt 10, Non Therapy V Elapsed Time: 00:03:54

Attempt Information:

All programmed therapy has been delivered.

© 2014


Boston Scientific Corporation
or its affiliates. All rights reserved.


Page 2 of 24

Clinician Signature:	

Stability: (97 ms, Off) V>A Rate: (True, Off) AFib: (False, Off) SRD Met: (False, Off) ATP Timeout: False

# Event Ended 00:20:14








2868 Software Version: 3.07 G179 Firmware Version: E\_v1.02.00(2.01) © 2014


Boston Scientific Corporation
or its affiliates. All rights reserved.


Page 3 of 24




























2868 Software Version: 3.07 G179 Firmware Version: E\_v1.02.00(2.01) © 2014


Boston Scientific Corporation
or its affiliates. All rights reserved.


Page 11 of 24












2868 Software Version: 3.07 G179 Firmware Version: E\_v1.02.00(2.01) © 2014


Boston Scientific Corporation
or its affiliates. All rights reserved.


Page 13 of 24



















2868 Software Version: 3.07 G179 Firmware Version: E\_v1.02.00(2.01) © 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 16 of 24





VT LVS VF VF VF VT VF VF LVS 290 133 245 220190 26£173 15520 VF VF LVS LVS VF VF 170 238 40 105 165 148 V-Dur

