

Pacemaker Model: Medtronic Sensia SESR01 Serial Number: PWR612775

Date of Visit: 02/10/14

Patient Name:	ID:	Physician:
----------------------	------------	-------------------

Pacemaker Model: Sensia SESR01 PWR612775 Implanted: 12/05/07 2:07 PM
 Atrial Lead:
 Ventricular Lead:

Pacemaker Status: 02/10/14 12:45:42 PM

Estimated remaining longevity: 3 years, 1.5 - 3.5 years (Based on Past History)

Battery Status OK
 Voltage 2.73 V
 Current 11.55 μ A
 Impedance 2,022 ohms

Lead Status: 02/10/14 12:45:42 PM

Ventricular Lead	
Output Energy	2.78 μ J
Measured Current	3.63 mA
Measured Impedance	526 ohms
Pace Polarity	Bipolar

V. Capture Management - from 02/10/14 11:59 AM

Measured Threshold: 1.250 V at 0.40 ms

Sensing Assurance - week ending 02/10/14

Min. R-Wave Amplitude 11.2 mV
 Max. R-Wave Amplitude 22.4 mV
 Min. Safety Margin 2.8X

In-Office Threshold Test Results

Ventricular Sensing Threshold

R-wave 11.20 mV - 15.68 mV

Patient Name:	ID:	Physician:
----------------------	------------	-------------------

Permanent Parameters (> indicates changes)**Modes**

	Initial	Final
Mode	VVIR	VVIR

Rates

Lower Rate	55 ppm	55 ppm
Upper Sensor Rate	130 ppm	130 ppm
ADL Rate	95 ppm	95 ppm

Refractory/Blanking

Ventricular Refractory	330 ms	330 ms
------------------------	--------	--------

Rate Response

Optimization	On	On
ADL Response	3	3
Exertion Response	3	3
ADLR Percent	2.0%	2.0%
Activity Threshold	Medium/Low	Medium/Low
Activity Acceleration	30 sec	30 sec
Activity Deceleration	Exercise	Exercise
High Rate Percent	0.2%	0.2%
ADL Rate Setpoint	5	5
Upper Rate Setpoint	15	15

Ventricular Lead

Amplitude	1.875 V	1.875 V
Pulse Width	0.40 ms	0.40 ms
Sensitivity	5.60 mV	5.60 mV
Sensing Assurance	On	On
Pace Polarity	Bipolar	Bipolar
Sense Polarity	Bipolar	Bipolar
Lead Monitor	Monitor Only	Monitor Only
Maximum Impedance	4,000 ohms	4,000 ohms
Minimum Impedance	200 ohms	200 ohms
Monitor Sensitivity	8	8
Capture Management	Adaptive	Adaptive
Amplitude Margin	1.5x	1.5x
Min. Adapted Amplitude	1.250 V	1.250 V
Capture Test Frequency	Day at Rest	Day at Rest
Acute Phase	Off	Off
Acute Phase Complete		03/26/08
V. Sensing During Search	Adaptive	Adaptive

Additional Features

Sleep	On	On
Sleep Rate	50 ppm	50 ppm
Bed Time	9:00:00 PM	9:00:00 PM
Wake Time	6:00:00 AM	6:00:00 AM
Transtelephonic Monitor	Off	Off
Extended Telemetry	Off	Off
Extended Marker	Standard	Standard
Implant Detection	Off/Complete	Off/Complete

Ventricular High Rate Episodes

Detection Rate	180 ppm	180 ppm
Detection Beats	5 beats	5 beats
Termination Beats	5 beats	5 beats
Collection Method	Rolling	Rolling

Selectable Diagnostic (Final Settings)

Chronic Lead Trend	On
High Rate Detail	
Include Refractory Senses?	Include
EGM Type	EGM
EGM Allocation	4 for 2/2 secs
EGM Timeout	8 weeks