



ZOOM® View™

QUICK NOTES® Report

KRÍZ, Karel

Date of Birth

30 Aug 1954

Device

AUTOGEN X4 CRT-D G179/136339

Tachy Mode

Monitor + Therapy

Report Created 05 May 2016

Last Office Interrogation

23 Mar 2016

Implant Date

12 Oct 2015

My Alerts

⚠ 05 May 2016 19:59 Check LV Lead

Events Since Last Reset (12 Nov 2015)

25 Apr 2016 13:30 ATR at 127 min⁻¹, Avg V Rate in ATR: 124 min⁻¹08 Apr 2016 08:34 ATR at 73 min⁻¹, Avg V Rate in ATR: 79 min⁻¹

See last page for full list of Events Since Last Reset.

Battery OK

Approximate time to explant: 7 years

Charge Time 9.5 s

Last Capacitor Re-form 20 Jan 2016 10:08

One Year Remaining

Explant



Leads Data

Implant

12 Oct 2015

Previous Session

Most Recent

Atrial

Intrinsic Amplitude N/R mV 7.6 mV 4.6 mV

Pace Impedance N/R Ω 829 Ω 766 Ω

Pace Threshold N/R V @ N/R ms 1.0 V @ 0.4 ms No data collected

Right Ventricular

Intrinsic Amplitude N/R mV 13.9 mV 8.7 mV @ 101 min⁻¹

Pace Impedance N/R Ω 391 Ω 352 Ω

Pace Threshold N/R V @ N/R ms 0.8 V @ 0.4 ms No data collected

Left Ventricular

Intrinsic Amplitude N/R mV 4.4 mV 2.8 mV

Pace Impedance N/R Ω 731 Ω 558 Ω

Pace Threshold N/R V @ N/R ms 1.1 V @ 0.4 ms No data collected

Shock Vector

Shock Impedance N/R Ω 55 Ω 46 Ω

Brady/CRT Counters Since Last Reset (12 Nov 2015)

Atrial 40% Paced

Right Ventricular 95% Paced

Left Ventricular 98% Paced

Settings

Ventricular Tachy Settings

VF 230 min⁻¹ ATP 36J, 41J, 41Jx6VT 185 min⁻¹ Burst Ramp 31J, 36J, 41Jx4

Atrial Tachy Settings

ATR Mode Switch 170 min⁻¹ DDI

Brady/CRT Settings

Mode DDD

Lower Rate Limit 65 min⁻¹Maximum Tracking Rate 130 min⁻¹

Paced AV Delay 220 - 250 ms

Sensed AV Delay 165 - 190 ms

A-Refractory (PVARP) 240 - 280 ms

RV-Refractory (RVRP) 230 - 250 ms

LV-Refractory (LVRP) 250 ms

Ventricular Pacing Chamber BiV

LV Offset -60 ms

Pacing Output

Atrial 3.0 V @ 0.4 ms

Right Ventricular 3.0 V @ 0.4 ms

Left Ventricular 4.0 V @ 0.4 ms

Sensitivity

Atrial AGC 0.25 mV

Right Ventricular AGC 0.6 mV

Left Ventricular AGC 1.0 mV

Leads Configuration (Pace/Sense)

Atrial Bipolar

Right Ventricular Bipolar

Settings (Continued)

Brady/CRT Settings

Leads Configuration (Pace/Sense)

LV (Pace)

LVRing3>>Can

LV (Sense)

LVTip1>>LVRing2

All Events Since Last Reset (12 Nov 2015)

05 May 2016 16:24	PMT at 130 min ⁻¹
05 May 2016 16:17	PMT at 130 min ⁻¹
05 May 2016 16:16	PMT at 130 min ⁻¹
05 May 2016 09:44	PMT at 130 min ⁻¹
05 May 2016 09:27	PMT at 130 min ⁻¹
25 Apr 2016 19:40	NonSustV at 186 min ⁻¹ , Nonsustained
25 Apr 2016 13:30	ATR at 127 min ⁻¹ , Avg V Rate in ATR: 124 min ⁻¹
08 Apr 2016 08:34	ATR at 73 min ⁻¹ , Avg V Rate in ATR: 79 min ⁻¹
31 Mar 2016 06:12	ATR at 67 min ⁻¹ , Avg V Rate in ATR: 71 min ⁻¹
09 Mar 2016 08:45	ATR at 188 min ⁻¹ , Avg V Rate in ATR: 88 min ⁻¹
19 Jan 2016 10:50	ATR at 225 min ⁻¹ , Avg V Rate in ATR: 80 min ⁻¹
19 Jan 2016 10:40	ATR at 224 min ⁻¹ , Avg V Rate in ATR: 107 min ⁻¹
19 Jan 2016 10:18	ATR at 224 min ⁻¹ , Avg V Rate in ATR: 88 min ⁻¹
19 Jan 2016 09:56	ATR at 228 min ⁻¹ , Avg V Rate in ATR: 93 min ⁻¹
19 Jan 2016 06:33	ATR at 216 min ⁻¹ , Avg V Rate in ATR: 89 min ⁻¹
18 Jan 2016 22:05	ATR at 221 min ⁻¹ , Avg V Rate in ATR: 71 min ⁻¹