



LATITUDE™ Programming System

Device Follow-up Report

Report Created 17 Sep 2021

Date of Birth N/R N/R N/R
Device PROPONENT MRI EL L231/ 938901

Last Office Interrogation
17 Sep 2021
Implant Date
N/R

My Alerts

There are no alerts to display.

Events Since Last Reset (16 Sep 2021)

17 Sep 2021 08:51 RA Auto
16 Sep 2021 11:53 RV Auto

Battery OK

Approximate time to explant: > 11 years
Magnet Rate 100 min⁻¹



Leads Data

Implant
N/R N/R N/R

Previous Session

Most Recent

Atrial

Intrinsic Amplitude	N/R mV	7.9 mV	6.8 mV
Pace Impedance	N/R Ω	744 Ω	733 Ω
Pace Threshold	N/R V @ N/R ms	1.0 V @ 0.4 ms	Auto 0.7 V @ 0.4 ms

Ventricular

Intrinsic Amplitude	N/R mV	Paced mV	Paced mV @ N/R min ⁻¹
Pace Impedance	N/R Ω	620 Ω	625 Ω
Pace Threshold	N/R V @ N/R ms	4.0 V @ 0.6 ms	Auto 2.3 V @ 0.4 ms

POST completed. Test started at 16 Sep 2021 11:01

Settings

Ventricular Tachy Settings

Ventricular Tachy EGM Storage On
Detection Rate 160 min⁻¹

Brady Settings

Mode DDD
RHYTHMIQ™ Off
Lower Rate Limit 60 min⁻¹
Maximum Tracking Rate 130 min⁻¹
Paced AV Delay 80 - 180 ms
Sensed AV Delay 65 - 150 ms
A-Refractory (PVARP) 240 - 280 ms
V-Refractory (VRP) 230 - 250 ms

Atrial Tachy Settings

ATR Mode Switch 170 min⁻¹ DD1

Pacing Output

Atrial Trend 3.5 V @ 0.4ms
Ventricular 5.5 V @ 0.6ms

Sensitivity

Atrial Fixed 0.75 mV
Ventricular Fixed 2.5 mV

Leads Configuration (Pace/Sense)


Atrial Bipolar
Ventricular Bipolar

Rate Adaptive Pacing


Minute Ventilation Passive
Accelerometer Passive

Trends


P-Wave Amplitude




A Pace Impedance




A Pace Threshold




R-Wave (V) Amplitude



V Pace Impedance





V Pace Threshold



Ventricular Tachy Counters		Since Last Reset 16 Sep 2021	Device Totals
Ventricular Episode Counters			
Total Episodes		0	0
VT Episodes (V>A)		0	0
SVT Episodes (V<=A)		0	0
Nonsustained Episodes		0	0
Brady Counters		Reset Before Last N/R day(s) N/R to 16 Sep 2021	Since Last Reset 1 day(s) 16 Sep 2021 to Today
Counters			
% A Paced		0	95
% V Paced		0	97
Intrinsic Promotion			
AV Search +			
% Successful		0	0
Rate Hysteresis			
% Successful		0	0
Atrial Arrhythmia			
% AT/AF		N/R	N/R
Total Time in AT/AF		N/R	N/R
Episodes by Duration			
< 1 minute		0	0
1 min - < 1 hr		0	0
1 hr - < 24 hr		0	0
24 hr - < 48 hr		0	0
> 48 hr		0	0
Total PACs		0	330
Ventricular Counters			
Total PVCs		0	2.4K
Three or More PVCs		0	0

Histograms Reset Before Last (N/R)



Histograms Since Last Reset (16 Sep 2021)

