



ZOOM® View™

## Combined Follow-up Report

Hucík, Stepán

Date of Birth 7 Jan 1944  
Device INCEPTA CRT-D P162/119874  
Tachy Mode Monitor + Therapy

Report Created 10 Jun 2019

Last Office Interrogation

02 May 2019

Implant Date

12 Mar 2013

## My Alerts

▲ 10 Jun 2019 11:06 Check Atrial and LV Leads

## Events Since Last Reset (04 Jun 2018)

07 Mar 2019 21:06 NonSustV at 159 min<sup>-1</sup>, Nonsustained30 Nov 2018 10:37 NonSustV at 136 min<sup>-1</sup>, Nonsustained

See last page for full list of Events Since Last Reset.

## Battery OK

Approximate time to explant: 2.5 years

Charge Time 11.3 s

Last Capacitor Re-form 28 Apr 2019 09:28

One Year Remaining

Explant



## Leads Data

## Implant

12 Mar 2013

## Previous Session

## Most Recent

## Atrial

Intrinsic Amplitude N/R mV N/R mV N/R mV

Pace Impedance N/R Ω &gt;2500 Ω &lt;200 Ω

Pace Threshold N/R V @ N/R ms 4.0 V @ 0.4 ms

## Right Ventricular

Intrinsic Amplitude 7.0 mV 6.0 mV N/R mV @ N/R min<sup>-1</sup>

Pace Impedance 650 Ω 507 Ω 495 Ω

Pace Threshold 0.5 V @ 0.4 ms 1.1 V @ 0.4 ms

## Left Ventricular

Intrinsic Amplitude 12.0 mV 20.7 mV N/R mV

Pace Impedance 650 Ω 545 Ω 494 Ω

Pace Threshold 1.3 V @ 0.4 ms 1.2 V @ 0.4 ms

## Shock Vector

Shock Impedance 45 Ω 56 Ω 52 Ω

## Settings

## Ventricular Tachy Settings

VF 230 min<sup>-1</sup> ATP 31J, 41J, 41Jx6VT 205 min<sup>-1</sup> Scan Ramp 31J, 41J, 41Jx4VT-1 160 min<sup>-1</sup> Scan Ramp 41J, 41J, 41Jx3

## Brady/CRT Settings

Mode VVIR

Lower Rate Limit 65 min<sup>-1</sup>Maximum Sensor Rate 130 min<sup>-1</sup>

RV-Refractory (RVRP) 230 - 250 ms

LV-Refractory (LVRP) 250 ms

Ventricular Pacing Chamber BiV

LV Offset 0 ms

## Pacing Output

Right Ventricular 2.5 V @ 0.4 ms

Left Ventricular 3.0 V @ 0.4 ms

## Sensitivity

Atrial AGC 0.25 mV

Right Ventricular AGC 0.6 mV

Left Ventricular AGC 1.0 mV

## Leads Configuration (Pace/Sense)

Atrial Bipolar/Off

Right Ventricular Bipolar

Left Ventricular LVring&gt;&gt;RV / LVtip&gt;&gt;LVring

## Sensors


Accelerometer On

## Rhythm ID Update

Last Successful Completion


N/R

## Trends




## Trends (Continued)


### P-Wave Amplitude




### A Pace Impedance




### R-Wave (RV) Amplitude




### RV Pace Impedance




### R-Wave (LV) Amplitude



### LV Pace Impedance




### Shock Impedance




Ventricular Tachy Counters		Since Last Reset	Device Totals
		04 Jun 2018	
<b>Ventricular Episode Counters</b>			
Total Episodes		9	7.2K
Treated			
VF Therapy		0	0
VT Therapy		0	4
VT-1 Therapy		0	17
Commanded Therapy		0	0
Non-Treated			
No Therapy Programmed		0	0
Nonsustained Episodes		9	7.2K
Other Untreated Episodes		0	7
<b>Ventricular Therapy Counters</b>			
ATP Delivered		0	29
ATP % Successful		0	67
Shocks Delivered		0	6
First Shock % Successful		0	100
Shocks Diverted		0	2
Brady/CRT Counters		Reset Before Last	Since Last Reset
		1907 day(s)	371 day(s)
		14 Mar 2013 to 04 Jun 2018	04 Jun 2018 to Today
<b>Counters</b>			
% A Paced		0	0
% RV Paced		90	93
% LV Paced		97	100
<b>Intrinsic Promotion</b>			
Rate Hysteresis			
% Successful		0	0
<b>Atrial Arrhythmia</b>			
% AT/AF		N/R	0
(One Year of Data)			
Total Time in AT/AF (min)		N/R	0.0
(One Year of Data)			
Episodes by Duration			
< 1 minute		0	0
1 min - < 1 hr		0	0
1 hr - < 24 hr		0	0
24 hr - < 48 hr		0	0
> 48 hr		0	0
Total PACs		0	0
<b>Ventricular Counters</b>			
Total PVCs		97	0
Three or More PVCs		9	0

### Histograms Reset Before Last (14 Mar 2013)



### Histograms Since Last Reset (04 Jun 2018)



## Heart Rate Variability

### Last Measured:

Last Measured Data unavailable

### Reference:

Reference Data Unavailable

### All Events Since Last Reset (04 Jun 2018)

07 Mar 2019 21:06	NonSustV at 159 min <sup>-1</sup> , Nonsustained
30 Nov 2018 10:37	NonSustV at 136 min <sup>-1</sup> , Nonsustained
05 Nov 2018 13:35	NonSustV at 143 min <sup>-1</sup> , Nonsustained
25 Oct 2018 00:58	NonSustV at 128 min <sup>-1</sup> , Nonsustained
02 Sep 2018 05:41	NonSustV at 163 min <sup>-1</sup> , Nonsustained
03 Aug 2018 22:41	NonSustV at 145 min <sup>-1</sup> , Nonsustained
02 Aug 2018 20:20	NonSustV at 171 min <sup>-1</sup> , Nonsustained
25 Jul 2018 17:13	NonSustV at 166 min <sup>-1</sup> , Nonsustained
24 Jul 2018 12:34	NonSustV at 154 min <sup>-1</sup> , Nonsustained