	ZOOM® View™	Report Created 06 Mar 2019
	Combined Follow-up Report	Last Office Interrogation 21-JAN-2019 Implant Date N.R. N.R. N.R.
Device ALTRUA 50 S501/590330		

Events Since Last Reset (22-JAN-2018)

There are no events to display.


Battery Good

Longevity Remaining >5.0 years

@ Current Pacing Percentage(s)

Magnet Rate 100 min⁻¹

Elective Replacement (ERT)
End of Life (EOL)



Beginning of Life (BOL)

Leads Data

Implant

N.R. N.R. N.R.

Previous Session

Present Session

21-JAN-2019

Intrinsic Amplitude

N.R. mV

9.9 mV

7.4 mV

Pace Impedance

N.R. Ω

480 Ω

460 Ω

Pace Threshold

N.R. V

1 V @ 0.4 ms

Settings

Tachy Settings

Tachy Detection

On

Rate

170 min⁻¹

Duration

8 cycles

Brady Settings

Mode

SSIR

Pacing Output

AUTO 1.5V @ 0.40 ms

Lower Rate Limit

55 min⁻¹

Sensitivity

AUTO 3.76 mV

Maximum Tracking Rate

-- min⁻¹

Lead Configuration

Bipolar

Maximum Sensor Rate

130 min⁻¹

Pace

--

Refractory

250 ms

Sense

--

Daily Measurement

Date	Amplitude (mV)	Impedance (Ω)	Threshold (V)
20-JAN-2019	>12.0	460	1.1
19-JAN-2019	N.R.	440	1.0
18-JAN-2019	>12.0	440	0.9
17-JAN-2019	N.R.	460	0.8
16-JAN-2019	9.3	460	0.9
15-JAN-2019	N.R.	460	0.9
14-JAN-2019	>12.0	470	1.1
11-JAN-2019	9.2	440	1.1
04-JAN-2019	5.9	440	1.1
28-DEC-2018	6.3	430	1.1
21-DEC-2018	6.2	440	1.1
14-DEC-2018	6.7	440	1.1
07-DEC-2018	6.5	430	1.1
30-NOV-2018	6.7	430	1.1
23-NOV-2018	7.1	460	1.0
16-NOV-2018	6.2	430	1.1

Daily Measurement (Continued)

Date	Amplitude (mV)	Impedance (Ω)	Threshold (V)
09-NOV-2018	6.5	440	1.0
02-NOV-2018	6.5	430	1.1
26-OCT-2018	6.4	430	1.1
19-OCT-2018	1.9	440	1.0
12-OCT-2018	6.0	440	1.0
05-OCT-2018	2.3	430	1.0
28-SEP-2018	5.9	430	1.0
21-SEP-2018	5.9	440	1.1
14-SEP-2018	7.0	430	1.0
07-SEP-2018	6.0	440	1.0
31-AUG-2018	5.7	440	1.1
24-AUG-2018	5.5	440	1.0
17-AUG-2018	6.3	430	1.0
10-AUG-2018	5.9	460	1.1
03-AUG-2018	6.7	440	1.1
27-JUL-2018	6.3	440	1.1
20-JUL-2018	6.4	440	1.0
13-JUL-2018	7.4	440	1.0
06-JUL-2018	6.7	460	1.1
29-JUN-2018	6.8	460	1.0
22-JUN-2018	5.4	440	1.1
15-JUN-2018	7.3	430	1.0
08-JUN-2018	6.5	440	1.0
01-JUN-2018	5.8	440	1.0
25-MAY-2018	5.2	460	1.1
18-MAY-2018	6.7	440	1.1
11-MAY-2018	4.7	460	1.1
04-MAY-2018	5.9	440	1.0
27-APR-2018	5.4	470	1.1
20-APR-2018	4.6	460	1.1
13-APR-2018	3.8	440	1.0
06-APR-2018	4.9	440	1.0
30-MAR-2018	4.1	460	1.1
23-MAR-2018	4.4	460	1.1
16-MAR-2018	4.3	460	1.0
09-MAR-2018	4.0	460	1.0
02-MAR-2018	5.1	440	1.1
23-FEB-2018	5.1	460	1.1
16-FEB-2018	4.7	440	1.0
09-FEB-2018	5.1	470	1.0
02-FEB-2018	7.4	460	1.0
26-JAN-2018	5.0	470	1.1

Activity Log

Date	Percent Active	Exercise Intensity
20-JAN-2019	1	3
19-JAN-2019	1	2
18-JAN-2019	1	2
17-JAN-2019	1	2
16-JAN-2019	2	3
15-JAN-2019	1	2
14-JAN-2019	1	2
12-JAN-2019	1	4
05-JAN-2019	1	3
29-DEC-2018	1	3
22-DEC-2018	1	4
15-DEC-2018	1	3
08-DEC-2018	1	3
01-DEC-2018	1	3
24-NOV-2018	1	3
17-NOV-2018	1	3
10-NOV-2018	1	4
03-NOV-2018	1	3
27-OCT-2018	1	4
20-OCT-2018	1	3
13-OCT-2018	1	3
06-OCT-2018	1	3
29-SEP-2018	1	3
22-SEP-2018	1	3
15-SEP-2018	1	4
08-SEP-2018	1	4
01-SEP-2018	1	3
25-AUG-2018	1	4
18-AUG-2018	1	4
11-AUG-2018	1	3
04-AUG-2018	1	4
28-JUL-2018	1	3
21-JUL-2018	1	4
14-JUL-2018	1	4
07-JUL-2018	1	4
30-JUN-2018	1	3
23-JUN-2018	1	3
16-JUN-2018	1	4
09-JUN-2018	1	4
02-JUN-2018	1	4
26-MAY-2018	1	3
19-MAY-2018	1	4
12-MAY-2018	1	3
05-MAY-2018	1	4
28-APR-2018	1	4

Activity Log (Continued)


Date	Percent Active	Exercise Intensity
21-APR-2018	1	4
14-APR-2018	1	3
07-APR-2018	1	3
31-MAR-2018	1	5
24-MAR-2018	1	3
17-MAR-2018	1	3
10-MAR-2018	1	3
03-MAR-2018	1	4
24-FEB-2018	1	3
17-FEB-2018	1	3
10-FEB-2018	1	3
03-FEB-2018	1	3
27-JAN-2018	1	3
20-JAN-2018	1	4

Trends


Events

Feb 18	Mar 18	Apr 18	May 18	Jun 18	Jul 18	Aug 18	Sep 18	Oct 18	Nov 18	Dec 18	Jan 19


Percent Active



Exercise Intensity




Amplitude




Trends (Continued)

Impedance



Threshold



Tachy Counters

Since Last Reset
22-JAN-2018

Episode Counters

Tachy Detections

0

Brady Counters

Percentages

Since Last Reset
22-JAN-2018

Counters

Paced

27 %

9.0M

Sensed

73 %

24.6M

Intrinsic Promotion

Rate Hysteresis

Searches

0

Successful Searches

0

Counters

Interval Variation

Variation 0 <= 10%

17.7M

Variation 10 <= 20%

2.4M

Variation 20 <= 30%

3.1M

Variation > 30%

10.3M

Histograms

Since Last Reset(22-JAN-2018)

