

ZOOM ® View™	Report Created 26 Aug 2021

QUICK NOTES ® Report

Date of Birth N/R N/R N/R

Device ACCOLADE MRI L310/836133

Last Office Interrogation

07 Jul 2021 Implant Date N/R

My Alerts

There are no alerts to display.

Events Since Last Reset (02 Jul 2021)

26 Aug 2021 02:51 RV Auto

Battery OK

Approximate time to explant: 11 years
Magnet Rate 100 min⁻¹

One Year Remaining

Explant

Leads Data Implant Previous Session Most Recent
N/R N/R N/R

Brady Counters Since Last Reset (02 Jul 2021)

Counters

% Paced 77

Settings

Tachy Settings

Tachy EGM Storage On
Detection Rate 160 min⁻¹

Brady Settings

Mode VVI Pacing Output

Lower Rate Limit 60 min⁻¹ Trend 2.5 V @ 0.4 ms

Refractory 250 ms Sensitivity

Fixed 2.0 mV

Leads Configuration (Pace/Sense)

Bipolar

2869 Software Version: 2.06 L310 Firmware Version: F_v1.00.00(1.10) © 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 1 of 1

Clinician Signature: