

**LATITUDE™ Programming System****Report Created 24 Jul 2021****Device Settings Report**

Date of Birth N/R N/R N/R
Device ACCOLADE MRI L310/ 839486

Last Office Interrogation
N/R
Implant Date
N/R

Programming

Last Programmed 24 Jul 2021

Tachy

Tachy EGM Storage On
Detection Rate 160 min⁻¹ (375 ms)

Atrial Tachy**Therapy****Ventricular Regulation**

Vent Rate Regulation Off

Brady

Settings

Mode	VVIR
Lower Rate Limit	55 min ⁻¹
Maximum Sensor Rate	130 min ⁻¹
Refractory	250 - 250 ms
Magnet Response	Pace Async
Noise Response	VOO

Rate Enhancements

Rate Smoothing	
Up	Off %
Down	Off %

Output

Sensitivity

Leads

Pace
Sense
Safety Switch

Rate Adaptive Pacing

Minute Ventilation	On
Response Factor	8
Fitness Level	Active
Ventilatory Threshold	120 min ⁻¹
Ventilatory Thresh. Response	70 %
Accelerometer	Passive

Minute Ventilation Sensor Settings

Minute Ventilation Sensor	On
Excitation Current	320 µA
Vector Selection	V Only
Signal Artifact Monitor	On

Minute Ventilation Sensor Status

Status	Calibrated
Current Vector	V

Auto 1.2 V @ 0.4ms

AGC 0.6 mV

Bipolar
Bipolar
On

Setup

Telemetry

Enable use of ZIP™ telemetry On

Trending Setup

Recording Method 30 Second Average

Duration 25 hours

Data Storage Continuous

Sleep Schedule

Sleep Start Time 23:00 hh:mm

Sleep Duration 07 hours

Leads Status Setup

	Daily Intrinsic Amplitude	Daily Impedance	Impedance Limits Low High	Pace Threshold
Pace/Sense	On	On	200 - 2000 Ω	On