



## LATITUDE™ Programming System

## Sensor Trending and Replay Report

Report Created 24 Jul 2021

Date of Birth  
DeviceN/R N/R N/R  
ACCOLADE MRI L310/ 839486Last Office Interrogation  
N/R  
Implant Date  
N/R

## Sensor Trending and Replay Settings

	Present	Replay
Lower Rate Limit	55	55 min <sup>-1</sup>
Max Sensor Rate	130	130 min <sup>-1</sup>
Minute Ventilation	On	On
Response Factor	8	8
Fitness Level	Active	Active
Ventilatory Threshold	120	120 min <sup>-1</sup>
Ventilatory Threshold Response	70	70 %
Accelerometer	Passive	Passive
Response Factor	---	---
Activity Threshold	---	---
Reaction Time	---	---
Recovery Time	---	---

## Sensor Trending - 30 Second Average

