

LATITUDE™ Programming System Report Created 17 Sep 2021

QUICK NOTES ® Report

Date of Birth

Device PROPONENT MRI EL L231/938901 Last Office Interrogation

17 Sep 2021 Implant Date N/R

My Alerts

There are no alerts to display.

Events Since Last Reset (16 Sep 2021)

RA Auto 17 Sep 2021 08:51 16 Sep 2021 11:53 **RV** Auto

Battery OK

> Approximate time to explant: > 11 years Magnet Rate 100 min⁻¹

One Year Remaining Explant

Leads Data Implant **Previous Session** Most Recent N/R N/R N/R Atrial Intrinsic Amplitude N/R mV 7.9 mV 6.8 mV Pace Impedance N/R Ω 744 Ω 733 Ω Pace Threshold N/R V @ N/R ms 1.0 V @ 0.4 ms Auto 0.7 V @ 0.4 ms Ventricular N/R mV Intrinsic Amplitude Paced mV Paced mV@N/R min⁻¹ Pace Impedance $N/R\Omega$ 620 Ω 625 Ω N/R V @ N/R ms 4.0 V @ 0.6 ms Pace Threshold Auto 2.3 V @ 0.4 ms

POST completed. Test started at 16 Sep 2021 11:01

Brady Counters Since Last Reset (16 Sep 2021)

Counters

Atrial 95% Paced Ventricular 97% Paced

Atrial Arrhythmia

V-Refractory (VRP)

% AT/AF N/R

Settings

Ventricular Tachy Settings Atrial Tachy Settings Ventricular Tachy EGM Storage ATR Mode Switch 170 min⁻¹ DDI On **Detection Rate** 160 min⁻¹ **Brady Settings** Mode DDD Pacing Output RYTHMIQ™ Off Atrial Trend 3.5 V @ 0.4ms 60 min⁻¹ Ventricular 5.5 V @ 0.6ms Lower Rate Limit Maximum Tracking Rate 130 min⁻¹ Sensitivity Fixed 0.75 mV Paced AV Delay 80 - 180 ms Atrial Sensed AV Delay 65 - 150 ms Fixed 2.5 mV Ventricular A-Refractory (PVARP) 240 - 280 ms Leads Configuration (Pace/Sense)

Atrial

Ventricular

230 - 250 ms

3869 Software Version: 1.05 L231 Firmware Version: F_v1.00.00(1.10)

© 2014-2017 **Boston Scientific Corporation** or its affiliates. All rights reserved. Page 1 of 1

Clinician Signature:

Bipolar

Bipolar