

ZOOM ® View™ Report Created 08 Oct 2018

QUICK NOTES ® Report ARRAMENDIOVÁ, Jaroslava

Date of Birth N/R N/R N/R

Device AUTOGEN MINI ICD D044/218955 Tachy Mode Monitor + Therapy Last Office Interrogation

05 Oct 2018 Implant Date **5 Oct 2018**

My Alerts

There are no alerts to display.

Events Since Last Reset (05 Oct 2018)

05 Oct 2018 09:00 VF at 276 min⁻¹, Induced: 26J

Battery OK

Approximate time to explant: > 5 years
Charge Time N/R s
Last Capacitor Re-form N/R

One Year Remaining

Explant

Leads Data Implant Previous Session Most Recent 5 Oct 2018

Ventricular

Shock Vector

Shock Impedance N/R Ω 81 Ω 79 Ω

Brady Counters Since Last Reset (05 Oct 2018)

Counters

Ventricular <1% Paced

Settings

Ventricular Tachy Settings

VF 220 min⁻¹ ATP 36J, 41J, 41Jx6 VT 170 min⁻¹ Burst Ramp/Scan 31J, 36J, 41Jx4 VT-1 150 min⁻¹ ATP1 Off ATP2 Off 41J, 41J, 41Jx3

Brady Settings

Mode VVI Pacing Output

Lower Rate Limit 40 min⁻¹ Ventricular 2.6 V @ 0.4 ms

V-Refractory (VRP) 250 ms Sensitivity

Ventricular AGC 0.6 mV

Leads Configuration (Pace/Sense)

Ventricular Bipolar

2868 Software Version: 4.04 D044 Firmware Version: E_v1.02.00(2.01)

© 2016

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 1 of 1

Clinician Signature: