



ZOOM® View™

QUICK NOTES® Report

Date of Birth N/R N/R N/R
Device INLIVEN W275/113732

Report Created 27 Nov 2018

Last Office Interrogation
29 May 2018
Implant Date
N/R

My Alerts

⚠ 27 Nov 2018 10:34 Check RV, Atrial and LV Leads

Events Since Last Reset (17 Feb 2015)

05 Mar 2016 17:12 VT (V>A) at 168 min⁻¹11 Nov 2015 11:06 VT (V>A) at 172 min⁻¹

See last page for full list of Events Since Last Reset.

Battery OK

Approximate time to explant: 5.5 years

Magnet Rate 100 min⁻¹

One Year Remaining

Explant



Leads Data

Implant
N/R N/R N/R

Previous Session

Most Recent

Atrial

Intrinsic Amplitude N/R mV 3.4 mV 3.1 mV

Pace Impedance N/R Ω >2000 Ω >2000 Ω

Pace Threshold N/R V @ N/R ms Test Failed No data collected

Right Ventricular

Intrinsic Amplitude N/R mV 24.7 mV 23.3 mV @ 46 min⁻¹

Pace Impedance N/R Ω 669 Ω 656 Ω

Pace Threshold N/R V @ N/R ms Auto 0.9 V @ 0.4 ms 0.9 V @ 0.4 ms

Left Ventricular

Intrinsic Amplitude N/R mV 2.6 mV 2.8 mV

Pace Impedance N/R Ω 705 Ω 609 Ω

Pace Threshold N/R V @ N/R ms 1.0 V @ 0.4 ms 1.1 V @ 0.4 ms

Brady/CRT Counters Since Last Reset (17 Feb 2015)

Counters

Atrial <1% Paced

Right Ventricular 62% Paced

Left Ventricular 62% Paced

Atrial Arrhythmia

% AT/AF 0

(One Year of Data)

Settings

Ventricular Tachy Settings

Ventricular Tachy EGM Storage On

Detection Rate 160 min⁻¹

Brady/CRT Settings

Mode VVIR

Lower Rate Limit 70 min⁻¹Maximum Sensor Rate 115 min⁻¹

RV-Refractory (RVRP) 250 - 250 ms

LV-Refractory (LVRP) 250 ms

Ventricular Pacing Chamber BiV

LV Offset 0 ms

Pacing Output

Right Ventricular Auto 5.0 V @ 0.4 ms

Left Ventricular 2.8 V @ 0.4 ms

Sensitivity

Atrial Fixed 0.75 mV

Right Ventricular Fixed 2.5 mV

Left Ventricular Fixed 2.5 mV

Leads Configuration (Pace/Sense)

Atrial --/Unipolar

Settings (Continued)

Brady/CRT Settings

Leads Configuration (Pace/Sense)

Right Ventricular

Left Ventricular

Bipolar

LVtip>>LVring

All Events Since Last Reset (17 Feb 2015)

27 Nov 2018 05:47	NonSustV at 162 min ⁻¹
26 Nov 2018 13:54	NonSustV at 129 min ⁻¹
26 Nov 2018 11:25	NonSustV at 168 min ⁻¹
22 Nov 2018 03:06	NonSustV at 169 min ⁻¹
21 Nov 2018 15:07	NonSustV at 157 min ⁻¹
21 Nov 2018 02:53	NonSustV at 149 min ⁻¹
19 Nov 2018 23:46	NonSustV at 159 min ⁻¹
19 Nov 2018 16:35	NonSustV at 161 min ⁻¹
19 Nov 2018 16:34	NonSustV at 158 min ⁻¹
19 Nov 2018 16:33	NonSustV at 160 min ⁻¹
07 Feb 2018 12:40	SVT (V<=A) at 167 min ⁻¹
05 Mar 2016 17:12	VT (V>A) at 168 min ⁻¹
11 Nov 2015 11:06	VT (V>A) at 172 min ⁻¹
11 Nov 2015 07:27	VT (V>A) at 169 min ⁻¹
10 Nov 2015 14:12	VT (V>A) at 155 min ⁻¹
10 Nov 2015 11:28	VT (V>A) at 169 min ⁻¹