

	ZOOM ® View™ Arrhythmia Logbook Report		Report Created 05 May 2016	
	KRÍZ, Karel		Last Office Interrogation	
	Date of Birth	30 Aug 1954	23 Mar 2016	
	Device	AUTOGEN X4 CRT-D G179/136339	Implant Date	
	Tachy Mode	Monitor + Therapy	12 Oct 2015	

Event	Date/Time	Туре	Therapy	Duration hh:mm:ss
PMT-471	05 May 2016 16:24	PMT		
PMT-470	05 May 2016 16:17	PMT		
PMT-469	05 May 2016 16:16	PMT		
PMT-468	05 May 2016 09:44	PMT		
PMT-467	05 May 2016 09:27	PMT		
V-3	25 Apr 2016 19:40	NonSustV	Nonsustained	00:00:13
ATR-22	25 Apr 2016 13:30	ATR	Avg V Rate in ATR: 124 min ⁻¹	00:00:01
ATR-21	08 Apr 2016 08:34	ATR	Avg V Rate in ATR: 79 min ⁻¹	00:00:03
ATR-20	31 Mar 2016 06:12	ATR	Avg V Rate in ATR: 71 min ⁻¹	00:00:02
ATR-19	09 Mar 2016 08:45	ATR	Avg V Rate in ATR: 88 min⁻¹	00:00:09
ATR-18	19 Jan 2016 10:50	ATR	Avg V Rate in ATR: 80 min ⁻¹	03:56:45
ATR-17	19 Jan 2016 10:40	ATR	Avg V Rate in ATR: 107 min ⁻¹	00:09:25
ATR-16	19 Jan 2016 10:18	ATR	Avg V Rate in ATR: 88 min ⁻¹	00:21:58
ATR-15	19 Jan 2016 09:56	ATR	Avg V Rate in ATR: 93 min ⁻¹	00:22:07
ATR-14	19 Jan 2016 06:33	ATR	Avg V Rate in ATR: 89 min ⁻¹	03:22:45
ATR-13	18 Jan 2016 22:05	ATR	Avg V Rate in ATR: 71 min ⁻¹	08:27:33
V-2	23 Oct 2015 09:48	NonSustV	Nonsustained	00:00:12
V-1	23 Oct 2015 09:44	NonSustV	Nonsustained	00:00:15

2868 Software Version: 3.07 G179 Firmware Version: E_v1.02.00(2.01)

© 2014 Boston Scientific Corporation or its affiliates. All rights reserved.

Page 1 of 1

Clinician Signature:	