

**ZOOM** ® View™ Report Created 12 Oct 2015

**QUICK NOTES ® Report** 

Hlucík, Stepán

Date of Birth

Device

7 Jan 1944

INCEPTA CRT-D P162/119874

Tachy Mode Monitor + Therapy Last Office Interrogation

08 Sep 2014 Implant Date 12 Mar 2013

My Alerts

**12 Oct 2015 09:10** Check Atrial and LV Leads

**Events Since Last Reset (14 Mar 2013)** 

VT at 200 min<sup>-1</sup>, ATPx4, 17J 30 Aug 2014 13:14 VT at 181 min<sup>-1</sup>, ATPx4, 17J 09 Aug 2014 16:46

See last page for full list of Events Since Last Reset.

**Battery** OK

> Approximate time to explant: 6 years Charge Time 9.7 s

> Last Capacitor Re-form 26 Jul 2015 10:20

**One Year Remaining** 

**Explant** 

Leads Data	<b>Implant</b> 12 Mar 2013	Previous Session	Most Recent
Atrial			
Intrinsic Amplitude	N/R mV	N/R mV	N/R mV
Pace Impedance	N/R $\Omega$	>2000 Ω	>2000 Ω
Pace Threshold	N/R V @ N/R ms	4.0 V @ 0.4 ms	
Right Ventricular			
Intrinsic Amplitude	7.0 mV	PVC mV	8.1 mV @ 79 min <sup>-1</sup>
Pace Impedance	$650~\Omega$	522 Ω	518 Ω
Pace Threshold	0.5 V @ 0.4 ms	1.0 V @ 0.4 ms	1.1 V @ 0.4 ms
Left Ventricular			
Intrinsic Amplitude	12.0 mV	PVC mV	13.7 mV
Pace Impedance	$650~\Omega$	$508~\Omega$	512 Ω
Pace Threshold	1.3 V @ 0.4 ms	1.4 V @ 0.4 ms	1.3 V @ 0.4 ms
Shock Vector			
Shock Impedance	45 Ω	57 Ω	55 Ω
Brady/CRT Counters Since Las	t Reset (14 Mar 2013)		

0% Paced Atrial Right Ventricular 93% Paced Left Ventricular 98% Paced

**Settings** 

**Ventricular Tachy Settings** 

VF 230 min<sup>-1</sup> ATP 31J, 41J, 41Jx6 VT 205 min<sup>-1</sup> Scan Ramp 31J, 41J, 41Jx4 VT-1 160 min<sup>-1</sup> Scan 41J, 41J, 41Jx3 Ramp

**Brady/CRT Settings** 

Mode **VVIR Pacing Output** Lower Rate Limit 60 min<sup>-1</sup> Right Ventricular 3.0 V @ 0.4 ms 3.0 V @ 0.4 ms Maximum Sensor Rate 130 min<sup>-1</sup> Left Ventricular RV-Refractory (RVRP) 230 - 250 ms Sensitivity LV-Refractory (LVRP) 250 ms AGC 0.25 mV Atrial BiV Ventricular Pacing Chamber Right Ventricular AGC 0.6 mV LV Offset 0 ms Left Ventricular AGC 1.0 mV

Leads Configuration (Pace/Sense)

Atrial Bipolar/Off Right Ventricular **Bipolar** 

2868 Software Version: 3.05

P162 Firmware Version: B\_v1.02.00(3.01)

© 2014 **Boston Scientific Corporation** or its affiliates. All rights reserved. Page 1 of 2

Clinician Signature:

## **Settings (Continued)**

**Brady/CRT Settings** 

## Leads Configuration (Pace/Sense)

Left Ventricular LVring>>RV / LVtip>>LVring

	Left Veritification	LVIIIIg>>KV / LVIIp>>LVIIII
All Events Since Last Rese	•	
11 Oct 2015 16:43	NonSustV at 155 min <sup>-1</sup> , Nonsustained	
11 Oct 2015 16:40	NonSustV at 166 min <sup>-1</sup> , Nonsustained	
11 Oct 2015 16:40	NonSustV at 169 min <sup>-1</sup> , Nonsustained	
11 Oct 2015 16:39	NonSustV at 163 min <sup>-1</sup> , Nonsustained	
11 Oct 2015 16:22	NonSustV at 160 min <sup>-1</sup> , Nonsustained	
11 Oct 2015 16:15	NonSustV at 159 min <sup>-1</sup> , Nonsustained	
11 Oct 2015 16:14	NonSustV at 153 min <sup>-1</sup> , Nonsustained	
11 Oct 2015 16:13	NonSustV at 162 min <sup>-1</sup> , Nonsustained	
11 Oct 2015 16:08 11 Oct 2015 16:06	NonSustV at 164 min <sup>-1</sup> , Nonsustained NonSustV at 160 min <sup>-1</sup> , Nonsustained	
23 Sep 2015 16:09	VT-1 at 163 min <sup>-1</sup> , ATPx1	
23 Sep 2015 15:54	VT-1 at 103 min <sup>-1</sup> , ATPX1	
23 Sep 2015 14:11	VT-1 at 173 min <sup>-1</sup> , ATPx1	
23 Sep 2015 13:11	VT-1 at 162 min <sup>-1</sup> , <b>♦</b> ATPx4	
·	VT-1 at 102 min <sup>-1</sup> , <b>♦</b> ATPx5	
23 Sep 2015 13:10		
23 Sep 2015 13:07	VT-1 at 173 min <sup>-1</sup> , <b>♦</b> ATPx2	
21 Sep 2015 20:26	VT-1 at 178 min <sup>-1</sup> , ATPx1	
21 Sep 2015 19:19	VT-1 at 186 min <sup>-1</sup> , ATPx1	
17 Sep 2015 12:20	VT-1 at 181 min <sup>-1</sup> , <b>♦</b> ATPx8, 41J	
17 Sep 2015 12:18	VT-1 at 181 min <sup>-1</sup> , <b>♦</b> ATPx5	
17 Sep 2015 10:44	VT-1 at 196 min <sup>-1</sup> , ATPx1	
30 Aug 2015 17:02	VT-1 at 155 min <sup>-1</sup> , <b>♦</b> ATPx3	
30 Aug 2015 16:19	VT-1 at 165 min <sup>-1</sup> , No Therapy	
29 Aug 2015 14:33	VT-1 at 170 min <sup>-1</sup> , No Therapy	
29 Aug 2015 14:32	VT-1 at 186 min <sup>-1</sup> , ATPx1	
29 Aug 2015 14:20	VT-1 at 176 min <sup>-1</sup> , ATPx1	
18 May 2015 10:09	VT-1 at 180 min <sup>-1</sup> , No Therapy	
25 Oct 2014 11:49	VT-1 at 183 min <sup>-1</sup> , <b>♦</b> ATPx8, 41J	
25 Oct 2014 11:38	VT-1 at 155 min <sup>-1</sup> , ATPx1	
10 Oct 2014 12:13	VT-1 at 168 min <sup>-1</sup> , <b>♦</b> ATPx8, 41J	
30 Aug 2014 13:14	VT at 200 min⁻¹, <b>◈</b> ATPx4, 17J	
09 Aug 2014 16:46	VT at 181 min⁻¹, <b>ጭ</b> ATPx4, 17J	
16 May 2013 17:54	VT at 178 min⁻¹, <b>�</b> ATPx2	
17 Apr 2013 15:14	VT at 179 min⁻¹, <b>�</b> ATPx4, 17J	
-		

2868 Software Version: 3.05

P162 Firmware Version: B\_v1.02.00(3.01)

© 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 2 of 2

Clinician Signature: