

Brussel, 20 april 2026

Kwaliteit van Laboratoria

GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2026.1A

Target HbA1c = 47.4 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180 series IFCC mmol/mol	47.6	18	1.7	3.6	0.2	0.4
ADAMS HA-8190 series IFCC mmol/mol	47.9	5	0.9	1.9	0.5	1.1
Afinion series IFCC mmol/mol	47.7	3	1.1	2.4	0.3	0.6
Alinity IFCC mmol/mol	47.2	7	0.8	1.7	-0.2	-0.4
Capillarys 3 IFCC mmol/mol	47.4	19	1.0	2.1	0.0	0.0
Cobas b 101 IFCC mmol/mol	44.0	2	1.4	3.2	-3.4	-7.2
Cobas C 303/503 IFCC mmol/mol	45.6	1	*****	*****	-1.8	-3.8
Cobas c501/502 IFCC mmol/mol	43.2	1	*****	*****	-4.2	-8.9
D-10_A1c_series IFCC mmol/mol	49.0	1	****	****	1.6	3.4
D-100_A1c_series IFCC mmol/mol	45.3	7	0.8	1.8	-2.1	-4.4
TOSOH G11 IFCC mmol/mol	48.2	39	0.7	1.4	0.8	1.7
TOSOH G8 IFCC mmol/mol	47.9	15	0.7	1.5	0.5	1.1
Hb NEXT IFCC mmol/mol	48.2	5	1.4	2.9	0.8	1.7
Minicap flex-piercing IFCC mmol/mol	46.5	3	1.4	3.0	-0.9	-1.9

Uw individueel rapport kunt u opvragen via de site

<https://hba1c-wholeblood.com>

De enquêtecoördinatoren

Y. Lenga

Dr. K. Vernelen