

Brussel, 20 april 2026

Kwaliteit van Laboratoria

GLOBAL RAPPORT EKE HbA1c VERS BLOED -2026.1C

Target HbA1c = 77.2 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180 series IFCC mmol/mol	76.8	18	2.8	3.7	-0.4	-0.5
ADAMS HA-8190 series IFCC mmol/mol	78.2	5	1.3	1.7	1.0	1.3
Afinion series IFCC mmol/mol	77.1	3	1.9	2.4	-0.1	-0.1
Alinity IFCC mmol/mol	76.3	7	1.6	2.1	-0.9	-1.2
Capillarys 3 IFCC mmol/mol	76.8	19	1.9	2.5	-0.4	-0.5
Cobas b 101 IFCC mmol/mol	70.0	2	2.8	4.0	-7.2	-9.3
Cobas C 303/503 IFCC mmol/mol	74.0	1	*****	*****	-3.2	-4.1
Cobas c501/502 IFCC mmol/mol	72.7	1	*****	*****	-4.5	-5.8
D-10_A1c_series IFCC mmol/mol	79.0	1	****	****	1.8	2.3
D-100_A1c_series IFCC mmol/mol	77.3	7	1.5	2.0	0.1	0.1
TOSOH G11 IFCC mmol/mol	78.7	39	0.7	0.8	1.5	1.9
TOSOH G8 IFCC mmol/mol	79.0	15	0.5	0.6	1.8	2.3
Hb NEXT IFCC mmol/mol	81.3	5	1.6	2.0	4.1	5.3
Minicap flex-piercing IFCC mmol/mol	77.3	3	0.6	0.7	0.1	0.1

Uw individueel rapport kunt u opvragen via de site

<https://hba1c-wholeblood.com>

De enquêtecoördinatoren

Y. Lenga

Dr. K. Vernelen