

Brussel, 20 avril 2026

Kwaliteit van Laboratoria

GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2026.1D

Target HbA1c = 33.9 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180 series IFCC mmol/mol	33.5	18	1.4	4.2	-0.4	-1.2
ADAMS HA-8190 series IFCC mmol/mol	33.6	5	1.1	3.1	-0.3	-0.9
Afinion series IFCC mmol/mol	33.2	3	0.2	0.6	-0.7	-2.1
Alinity IFCC mmol/mol	33.7	7	0.7	2.0	-0.2	-0.6
Capillarys 3 IFCC mmol/mol	33.0	19	0.8	2.4	-0.9	-2.7
Cobas b 101 IFCC mmol/mol	30.0	2	2.8	9.4	-3.9	-11.5
Cobas C 303/503 IFCC mmol/mol	35.0	1	*****	*****	1.1	3.2
Cobas c501/502 IFCC mmol/mol	33.3	1	*****	*****	-0.6	-1.8
D-10_A1c_series IFCC mmol/mol	34.0	1	****	****	0.1	0.3
D-100_A1c_series IFCC mmol/mol	29.5	7	1.0	2.0	-4.4	-13.0
TOSOH G11 IFCC mmol/mol	33.3	39	0.6	0.8	-0.6	-1.8
TOSOH G8 IFCC mmol/mol	33.1	15	0.6	0.6	-0.8	-2.4
Hb NEXT IFCC mmol/mol	31.3	5	2.2	2.0	-2.6	-7.7
Minicap flex-piercing IFCC mmol/mol	33.0	2	1.4	4.3	-0.9	-2.7

Uw individueel rapport kunt u opvragen via de site

<https://hba1c-wholeblood.com>

De enquêtecoördinatoren

Y. Lenga

Dr. K. Vernelen