

Bruxelles, le 20 avril 2026

Qualité des Laboratoires

RAPPORT GLOBAL EEQ HbA1c SANG FRAIS - 2026.1B

Target HbA1c = 56.6 mmol/mol (Méthode de référence NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180 series IFCC mmol/mol	56.7	18	1.7	3.1	0.1	0.2
ADAMS HA-8190 series IFCC mmol/mol	56.8	5	1.0	1.8	0.2	0.4
Afinion series IFCC mmol/mol	57.6	3	1.2	2.0	1.0	1.8
Alinity IFCC mmol/mol	57.5	7	1.0	1.7	0.9	1.6
Capillarys 3 IFCC mmol/mol	57.0	19	1.0	1.7	0.4	0.7
Cobas b 101 IFCC mmol/mol	52.0	2	1.4	2.7	-4.6	-8.1
Cobas C 303/503 IFCC mmol/mol	54.1	1	*****	*****	-2.5	-4.4
Cobas c501/502 IFCC mmol/mol	55.2	1	*****	*****	-1.4	-2.5
D-10_A1c_series IFCC mmol/mol	58.0	1	****	****	1.4	2.5
D-100_A1c_series IFCC mmol/mol	54.7	7	0.8	1.4	-1.9	-3.4
TOSOH G11 IFCC mmol/mol	57.0	39	0.6	1.1	0.4	0.7
TOSOH G8 IFCC mmol/mol	57.2	14	0.6	1.1	0.6	1.1
Hb NEXT IFCC mmol/mol	58.0	5	1.8	3.1	1.4	2.5
Minicap flex-piercing IFCC mmol/mol	56.8	3	0.4	0.7	0.2	0.4

Votre rapport individuel est disponible sur le site

<https://hba1c-wholeblood.com>

Les coordinateurs d'enquête

Y. Lenga

Dr. K. Vernelen