

Certificate of Analysis


Part Number:	090-257	Serial Number:	
Product Name:	Maurice CE-SDS PLUS-65 Cartridge	Date of Manufacture:	

This product was tested on a Maurice system and passed all physical and functional test criteria.

Test Methods

Analysis of Maurice CE-SDS PLUS-65 Cartridge on Maurice or Maurice S. systems. All tests were performed using ProteinSimple's reduced IgG standard. Figure 1, representative electropherogram for Reduced IgG standard, Peaks: internal standard (IS, 10KD), light chain (LC), non-glycosylated heavy chain (NGHC), glycosylated heavy chain (HC).

Figure 1.



Acceptance Results

PARAMETER	ACCEPTANCE CRITERIA	PASS/FAIL
Resolution, HC (Reduced Sample)	≥ 1.50	Pass
Peak Height, HC (Reduced Sample)	≥ 20.0 mAU	Pass

Reviewed and Approved:

Hooi Tan
Sr. Quality Manager

Date:



Toll-free: (888) 607-9692
Tel: (408) 510-5500
info@proteinsimple.com
proteinsimple.com

3001 Orchard Parkway
San Jose, CA 95134
USA