

## Maurice Turbo CE-SDS™ Cartridge


<b>Part Number</b>	101-0058	<b>Serial Number</b>	
<b>Product Name</b>	Maurice Turbo CE-SDS Cartridge	<b>Date of Manufacture</b>	
		<b>Expiration Date</b>	
		<b>Batch No.</b>	

This product was tested on a Maurice system with system suitability and passed all physical and functional test criteria.

### Test Methods

Analysis of Maurice Turbo CE-SDS Cartridge on Maurice or Maurice S. system. All tests were performed using ProteinSimple's reduced IgG standard. **FIGURE 1**, representative electropherogram for Reduced IgG standard. Peaks: internal standard (IS, 10KD), light chain (LC), non-glycosylated heavy chain (NGHC), glycosylated heavy chain (HC).

FIGURE 1



### ACCEPTANCE RESULTS

PARAMETER	ACCEPTANCE CRITERIA	PASS/FAIL
Resolution, HC (Reduced Sample)	$\geq 1.00$	Pass
Peak Height, HC (Reduced Sample)	$\geq 10$ mAU	Pass

Reviewed and Approved:

Hooi Tan  
Sr. Quality Manager

Date:

**Bio-Techne®** | R&D Systems™ Novus Biologicals™ Tocris Bioscience™ ProteinSimple™ ACD™