

cIEF Fractionation Cartridge


Part Number	101-0059	Serial Number	
Product Name	cIEF Fractionation Cartridge	Date of Manufacture	
		Expiration Date	
		Batch No.	

This product was tested on a MauriceFlex system with system suitability and passed all physical and functional test criteria.

Test Methods

Analysis of cIEF Fractionation Cartridge on a MauriceFlex system. All tests were performed using ProteinSimple's Maurice cIEF System Suitability. **FIGURE 1**, representative electropherogram collected in fluorescence mode, for pI markers 3.38, 5.85, 6.14, 9.99, and 10.17.

Figure 1



ACCEPTANCE RESULTS

PARAMETER	ACCEPTANCE CRITERIA	PASS/FAIL
pI Range	pI 3.38, 5.85, 6.14, 9.99, and 10.17 markers observed	Pass
Resolution 6.14 (Fluorescence, 10s exposure)	≥ 1.10	Pass
Resolution 10.17 (Fluorescence, 10s exposure)	≥ 0.90	Pass

Reviewed and Approved:

Hooi Tan
Sr. Quality Manager

Date:

Bio-Techne® | R&D Systems™ Novus Biologicals™ Tocris Bioscience™ ProteinSimple™ ACD™