

## **Certificate of Analysis**



## **Maurice cIEF Cartridge**

Part Number	090-101	Serial Number	
Product Name	Maurice cIEF Cartridge	Date of Manufacture	
		Expiration Date	

This product was tested on a Maurice system with system suitability and passed all physical and functional test criteria.

## **Test Methods**

Analysis of Maurice cIEF Cartridge on Maurice or Maurice C. systems. All tests were performed using ProteinSimple's Maurice cIEF System Suitability. FIGURE 1, representative electropherogram collected in absorbance mode, for pl markers 3.38, 5.85, 6.14, 9.99, and 10.17. FIGURE 2, representative electropherogram collected in fluorescence mode, for the same set of pl markers.

Figure 1. Absorbance





Figure 2. Fluorescence



## **ACCEPTANCE RESULTS**

PARAMETER	ACCEPTANCE CRITERIA	PASS/FAIL
pl Range	pl 3.38, 5.85, 6.14, 9.99, and 10.17 markers observed	Pass
Resolution 6.14 (Fluorescence, 10s exposure)	≥ 2.10	Pass
Resolution 10.17 (Fluorescence, 10s exposure)	≥ 1.70	Pass

Reviewed and Approved:

Hooi Tan

Sr. Quality Manager

Date:

**Bio-Techne®** | R&D Systems™ Novus Biologicals™ Tocris Bioscience™ ProteinSimple™ ACD™

ProteinSimple Proprietary & Confidential Information