



PROTOCOLCENTRAL ELECTRONICS PRIVATE LIMITED

526, 2nd Floor, 7th Cross, 10th Main, Jeevanbhima Nagar, HAL 3rd Stage, Bengaluru 560075, India
CIN: U74900KA2015PTC084745

U.K. Declaration of Conformity

Products Covered by this document:

1. HealthyPi 6 Kit (Part No: HEALTHYPI-6-KIT)
2. HealthyLink Edge AI Module for HealthyPi 6 (Part No: HEALTHYLINK-EDGE-AI)
3. HealthyLink EEG Module for HealthyPi 6 (Part No: HEALTHYLINK-EEG-8CH)

We, the undersigned, hereby declare that the aforementioned products are in conformity with the provisions of the following legislations of the UK:

- **Restriction of Hazardous Substance (RoHS) 2012 No. 3032**, presumption of conformity granted from compliance with the following set of harmonized standards:
 - RoHS: EN/IEC 63000: 2018 Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances
 - EN/IEC 62368-1: 2018 Audio/video, information and communication technology equipment - Part 1: Safety requirements
- **Radio Equipment (RED) 2017 No. 1206, The Radio Equipment Regulations 2017**, presumption of conformity granted from compliance with the following set of harmonized standards:
 - ETSI EN 300 328 V2.2.2:2019 (Data transmission equipment operating in the 2.4 GHz band; Harmonised Standard for access to radio spectrum)
 - ETSI EN 301 489-1 V2.2.3: 2019 (Data transmission equipment operating in the 2.4 GHz band, Harmonised Standard for ElectroMagnetic Compatibility)

Signed by : Ashwin K. Whitchurch, Managing Director, Protocentral Electronics Pvt. Ltd.

Date : 28/Jul/2026

Place: Bengaluru