

A Novel Passive Ferrofluid One-way (Check) Valve

Veronica Stuckey

Biomedical Engineer, University of Texas at Austin
Email: stuckey002@gmail.com

Robert L. Read

Founder, Public Invention
Email: read.robert@gmail.com

Small pumps and valves enable flow management in microfluidic systems. This paper presents a novel passive ferrofluid check valve. The valve consists of only a unique channel-and-chamber geometry, ferrofluid, and a stationary magnetic field. The flow is determined only by the inlet and output pressure, and the magnetic field is completely static. The prototype valve and experimental setup are explained and performance of the valves cracking and collapse pressure reported. This initial design can be used for microfluid handling and lab-on-a-chip applications.

1 INTRODUCTION

Ferrofluid can be manipulated by electronically controlled magnetic fields to exert force on fluids[1, 2, 3]. This makes it possible to build pneumatic or hydraulic devices, perhaps on very small scales, such as a single chip[4, 5], to miniaturize fluid handling. This has been proposed for biomedical purposes[6] that would use water or body fluids, although this paper reports only on experiments done with air. Miniature pumps and valves could be used to make a “lab on a chip” (LOC) or even to heat or cool different areas. A fundamental component of such devices is the check or one-way valve. Two check valves on either side of a chamber whose volume can vary creates a positive displacement pump. A perfect check valve opens or *cracks* with minimal pressure on the inlet side and sustains maximal pressure on the outlet side before *collapse*, allowing fluid to flow in only one direction. Following[7] we call the maximum pressure differential the valve can resist in the direction it is intended to check (from outlet to inlet) the *sustainable* or *collapse* pressure.

This article is a brief report on an initial but functioning design of a passive ferrofluid check valve (PFCV) that has no moving parts except for the ferrofluid bolus it-

self, which is stationary in normal operation. By passive, the authors mean a check valve that functions without changes to the magnetic field affecting the bolus, whether that field is induced by a permanent magnet or an electromagnet. That is, the flow is determined purely by the difference between the inlet port pressure and the outlet port pressure. (This terminology is not completely standardized, for example [7] uses the term *active* to simply mean a valve that is not a permanent seal.) To our knowledge, no passive ferrofluid check valve has been previously reported, despite being an active area of research and despite such a valve having significant advantages for operation but especially fabrication over valves with moving parts.

2 RELATED RESEARCH

A number of papers report on ferrofluid pumps, focusing in particular on micropump and lab-on-a-chip applications[3, 8]. Many of these papers use a version of mechanical valve not based on passive ferrofluid, even though they move a ferrofluid bolus with a magnetic field. For example, a corrugated silicone micro valve[4, 9] has been reported. Other researchers use active valves, which require synchronization with the ferrofluid plug to form a pump, such as [10], which describes an active T-Valve with a moving ferrofluid plug, and [11] describes a complete fluid pump with valves that use active control of a ferrofluid bolus. Active control is possible because the action of the plunger or bolus may be synchronized with the opening and closing of the valves. Nonetheless a passive valve would be simpler and less expensive, and would not require knowledge of the timing of the plunger. At least two additional kinds of active valves, a *well valve* and *Y valve*, have been described[7].

A passive ferrofluid two-way valve with tunable opening and closing pressure based on magnetic field




Fig. 1. The passive ferrofluid check valve components

strength[12] has been tested, but could not be passively used to make a pump.

This paper has not studied the closing pressure of the PFCV, but reports on the opening (or cracking) pressure (for flow from inlet to outlet) and sustainable (or collapse) pressure when the outlet pressure is higher than the inlet side. An interesting functional micropump in which the moving ferrofluid bolus merges with a fixed ferrofluid valve and then separates on each pumping cycle has been described[5], but is not a one-way valve.

3 PASSIVE FERROFLUID CHECK VALVE (PFCV) DESIGN

The PFCV depicted in Figure 1 is a simple asymmetric volume centered in a magnetic field which holds a ferrofluid bolus in place. In the center of a radially symmetric magnetic field a narrow channel meets a larger open chamber at a right angle. The ferrofluid bolus is large enough that at rest in the field it forms a semi-circle in the open chamber. The narrow channel is longer than the radius of the bolus at rest. The broad chamber is the outlet side of the valve. The narrow chamber opens onto a recovery chamber on the inlet side of the valve. This design allows the bolus to be recovered from the recovery chamber when the pressure is equalized if the outlet pressure is raised above the collapse pressure, driving the bolus away from the magnetic field. The PFCV does not resist pressure as well as a valve of the same size made out of moving, solid parts. That is, the sustainable pressure it can resist on the outlet side before failing is relatively low, and pressure required to crack it open and allow flow is relatively high. However, it may operate reliably within a range of known pressures, and thus be sufficient to build a pump-on-a-chip. Furthermore, the PFCV reported here is a preliminary design which can probably be significantly improved. The authors found the existence of the PFCV exciting enough to report on it immediately.



Fig. 2. Equipment set up

4 METHOD

The valve depicted Figure 1 was designed using Solidworks 2016. The model consists of a 15mm long, 2mm wide and 2mm high channel, two female luers, a magnet holder ring and two legs to provide room for the magnet. All volumes are 2mm high. The 3D shape of the chambers can be thought as a 2mm extrusion of a 2D shape. Viewed from the top, one end opens up to a recovery chamber of circular profile 30mm in diameter and the other an outlet chamber with a flat wall. A magnet holder ring 12.7mm inner diameter (1/2") was created centered on the channel-chamber junction, below where the bolus is placed, to hold a permanent magnet in place at the center of the valve. When two magnets are used, the magnet on top naturally stays in the same position due to attraction to the magnet below. On the inlet side the channel opens into a recovery chamber shaped to allow the ferrofluid to be passively drawn back into the channel by the magnetic field after a collapse of the bolus. The model was printed on a Projet MJP 2500 (3D Systems, Rock Hill, SC), using Visijet M2G-CL and VisiJet M2 SUP as material and support respectively (3D Systems, Rock Hill, SC). Support material was removed by using an EasyClean system (3D Systems, Rock Hill, SC) and Dawn dish soap (Procter & Gamble, Cincinnati, OH) to remove residuals.

As shown in Figure 2, a basixCOMPAK 30atm pressurizing syringe (Merit Medical, South Jordan, UT) is connected to the model via a two-way stopcock (Qosina, Ronkonkoma, NY), tubing (Natvar, City of Industry, CA), and male (Injectech, Fort Collins, CO) and female luer (Qosina, Ronkonkoma, NY), allowing integration of a manometer (General Tools, Secaucus, NJ) to measure pressure. A 12.7mm x 25.4mm (1/2" x 1") cylindrical neodymium magnet (Apex Magnets, Petersburg, WV) with a pull force of 14.6 kg (32.24 pounds) was placed inside the magnet channel by means of a tight fit and 0.4

Table 1. Result pressures

Magnet configuration	Cracking Pressure kPa (mmHg)	Collapse Pressure kPa (mmHg)	Pressure Difference kPa (mmHg)	Approx. Ratio: Cracking to Collapse Pressure
Single	1.1 (8)	5.5 (41)	4.4 (33)	1:5
Dual	8.5 (64)	17.5 (131)	8.9 (67)	1:2

mL of ferrofluid (Apex Magnets, Petersburg, WV) was injected into the model using a 3mL syringe (BH Supplies, Jackson, NJ).

To obtain values, pressure was applied through the pressurizing syringe. Pressure was first applied from the outlet side of the model. The maximum pressure difference from the outlet side the valve can withstand before collapsing, will be referred to as the sustainable pressure or *collapse* pressure. Pressure applied from the inlet needed to initiate flow will be referred to as the *cracking* pressure.

The cracking pressure was first measured by increasing the pressure difference on the inlet side until flow is initiated (at which point pushing the syringe plunger faster simply increases flow without increasing the pressure.) Then the collapse pressure was measured by increasing the pressure difference on the outlet side. At pressures below the sustainable pressure, the valve holds pressure well with no observable leaks of air in the short time (a few minutes) of our experiment. When the sustainable pressure is exceeded, the bolus explodes violently into the recovery chamber. When the pressure difference is equalized, the bolus may passively recover into the central position, or it may need to be actively “combed” with a magnet back into the central position.

The procedure was performed first with one magnet, named the “Single Magnet” configuration, placed below the channel-chamber connection. The “Dual Magnet” configuration was performed with the magnet in the same position as the Single Magnet case in the same position, but with a second magnet of equal one placed vertically on top of the model, arranged to be strongly attracted to the lower magnet.

5 RESULTS

The final pressures obtained demonstrate a clear difference between the inlet cracking pressure and outlet sustainable pressure, creating an effective passive check valve.

The ferrofluid had observable differences in behavior between the two configurations. After the pressure equal-

ized following a collapse of the bolus due to exceeding the sustainable pressure, the single magnet configuration often repaired itself by drawing the fluid back into a centered bolus passively. After a collapse with two magnets, fluid further from the bolus remained stationary while the fluid closer was pulled back to the center. Following the removal of the top magnet, the stationary fluid then began to return to the bolus. This is consistent with the localization of the magnetic field between two magnets, and the weakening of the magnetic field further from the channel-chamber juncture in the dual magnet configuration.

Although the dual magnet configuration demonstrated a larger absolute pressure difference due to magnet strength between the cracking and the collapse pressure, the single magnet configuration granted a larger ratio of collapse pressure to cracking pressure due to the much lower cracking pressure. The authors conjecture that the low cracking pressure may have been not only to the weaker magnetic field, but the weakening at the top of the 2mm high channel, which was further away from the magnet in the single magnet configuration.

6 CONCLUSIONS

This paper demonstrates an apparently novel passive ferrofluid one-way valve or check valve (PFCV). This valve is completely passive in that it depends entirely on the pressure at the inlet port and the outlet port. The valve has no moving parts (except for the ferrofluid, which is almost stationary), and a remarkably simple design, consisting of nothing but a channel, an inlet chamber, and outlet chamber, and a bolus of ferrofluid in a static magnetic field.

Although no effort has been made to optimize the design, the pressure difference between the cracking pressure and the sustainable back pressure appear great enough to make an effective micropump. The performance of this one-way valve may improve with additional design effort; the authors sought to publish this result as soon as it was observed. Obvious future research possibilities are:

1. To improve the performance by varying the geometry of the passive design or shape and strength of the magnetic field.
2. Utilizing this design to make a micro-pump similar to earlier micro-pumps but with this simpler check valve design.
3. To provide an explanatory and predictive theory of operation, for example based on magnetic field strength as per [11].
4. Studying the ability of the valve to recover after a

collapse automatically when high outlet pressure is removed, which would increase robustness in some applications.

REFERENCES

- [1] Torres-Díaz, I., and Rinaldi, C., 2014, “Recent progress in ferrofluids research: novel applications of magnetically controllable and tunable fluids,” *Soft matter*, **10**(43), pp. 8584–8602.
- [2] Kole, M., and Khandekar, S., 2021, “Engineering applications of ferrofluids: A review,” *Journal of Magnetism and Magnetic Materials*, p. 168222.
- [3] Özbey, A., Karimzadehkhoei, M., Yalçın, S. E., Gozuacik, D., and Koşar, A., 2015, “Modeling of ferrofluid magnetic actuation with dynamic magnetic fields in small channels,” *Microfluidics and Nanofluidics*, **18**(3), pp. 447–460.
- [4] Yamahata, C., Chastellain, M., Hofmann, H., and Gijs, M. A., 2003, “A ferrofluid micropump for lab-on-a-chip applications,” In Techn. Digest Eurosensors XVII, The 17th Europ. Conf. On Solid State Transducers, no. CONF.
- [5] Hatch, A., Kamholz, A. E., Holman, G., Yager, P., and Bohringer, K. F., 2001, “A ferrofluidic magnetic micropump,” *Journal of Microelectromechanical systems*, **10**(2), pp. 215–221.
- [6] Michelson, T., Rudnick, J., Baxter, J., and Rashidi, R., 2019, “A novel ferrofluid-based valve-less pump,” In ASME International Mechanical Engineering Congress and Exposition, Vol. 59445, American Society of Mechanical Engineers, p. V007T08A009.
- [7] Hartshorne, H., Backhouse, C. J., and Lee, W. E., 2004, “Ferrofluid-based microchip pump and valve,” *Sensors and Actuators B: Chemical*, **99**(2–3), pp. 592–600.
- [8] Hsu, M.-C., Alfadhel, A., Forouzandeh, F., and Borkholder, D. A., 2018, “Biocompatible magnetic nanocomposite microcapsules as microfluidic one-way diffusion blocking valves with ultra-low opening pressure,” *Materials & design*, **150**, pp. 86–93.
- [9] Yamahata, C., Chastellain, M., Parashar, V. K., Petri, A., Hofmann, H., and Gijs, M. A., 2005, “Plastic micropump with ferrofluidic actuation,” *Journal of microelectromechanical systems*, **14**(1), pp. 96–102.
- [10] Menz, A., Benecke, W., Perez-Castillejos, R., Plasza, J., Esteve, J., Garcia, N., Higuero, J., and Diez-Caballero, T., 2000, “Fluidic components based on ferrofluids,” In 1st Annual International IEEE-EMBS Special Topic Conference on Microtechnologies in Medicine and Biology. Proceedings (Cat. No. 00EX451), IEEE, pp. 302–306.
- [11] Ando, B., Ascia, A., Baglio, S., and Pitrone, N., 2009, “Ferrofluidic pumps: a valuable implementation without moving parts,” *IEEE Transactions on Instrumentation and Measurement*, **58**(9), pp. 3232–3237.
- [12] Paschalis, E. I., Chodosh, J., Sperling, R. A., Salvador-Culla, B., and Dohlman, C., 2013, “A novel implantable glaucoma valve using ferrofluid,” *PloS one*, **8**(6), p. e67404.