



Medical Physics Department

قسم الفيزياء الطبية

Annual Inspection Test for Radiation Protection Tools by Fluoroscopy Machine – KFHM 2018

Unit	Date of Test	Location of Test	Item	N.O	Code	Result	Action Plan
General X-Ray	11-8-2018	Fluoroscopy – Room 4	Lead Apron	9	50800010152	Pass	
					26502532	Pass	
					650230	Pass	
					14402800196	Pass	
					13679000029	Pass	
					14402800001	Pass	
					65021TC-22	Fail	Remove
					14402800187	Pass	
					SG058	Pass	
			Vest	7	76608222	Pass	
					09727	Pass	
					492114	Pass	
					26608322	Pass	
					4423216	Pass	
					96608222	Pass	
					966082-22	Pass	
			Thyroid Collar	6	XRT1 to XRT6	Pass	
			Gonadal Shield	8 + Set	XRG1 to XRG8	Pass	



Tested by Medical Physicist, Khaled Al-Raddadi

Special Thanks for

Technical Assistant, Mr. Bandar Mahmoud