



Medical Physics Department

Unit	Date of Test	Location of Test	Item	N.O	Code	Result	Action Plan
Angio	13-11-2019	Fluoroscopy – Room 4	Lead Apron	8	S49370	Pass	
					256021-224	Pass	
					SE09074058	Pass	
					SE09074053	Pass	
					22060955	Pass	
					14402800121	Pass	
					510250462	Pass	
					3165025-22	Pass	
			Vest	3	524773	Pass	
					524732	Pass	
					5247331	Pass	
			Skirt	2	66084-2233	Pass	
					5247255	Pass	
			Thyroid Collar	11	AG1 to AG9	Pass	
			Gonadal Shield	3	AG1 to AG3	Pass	
			Lead Glasses	8	Without Codes	Pass	

Medical Physicist

Mukhtar Al-Ansari

Medical physicist

Khaled Al-Radadi

Radiation Safety Officer

Faleh AL-Zaidi

Special Thanks for

Technical Assistant, Mr. Abdullah Al-Sharif



Medical Physics Department

Unit	Date of Test	Location of Test	Item	N.O	Code	Result	Action Plan
CT	1-9-2019	Fluoroscopy Room 4	Lead Apron	4	14178700003	Pass	
					13679100041	Pass	
					13679000049	Pass	
					13679200014	Pass	
			Thyroid Collar	3	CT1	Pass	
					CT2	Pass	
					CT3	Pass	

Medical Physicist

Mukhtar Al-Ansari

Medical physicist

Khaled Al-Radadi

Radiation Safety Officer

Faleh AL-Zaidi

Special Thanks for

Technical Assistant, Mr. Bandar Mahmoud



Medical Physics Department

Unit	Date of Test	Location of Test	Item	N.O	Code	Result	Action Plan
Endoscopy	17-6-2019	Fluoroscopy Room 4	Lead Apron	11	0-103	Pass	
					0093	Pass	
					0-041	Pass	
					0-092	Pass	
					67090-22	Pass	
					65021-22	Pass	
					68098LW-22	Pass	
					15468200007	Pass	
					367089-22	Pass	
					50800010152-2	Pass	
					467089-22	Pass	
			Thyroid Collar	4	E1	Pass	
					E2	Pass	
					E3	Pass	
					E4	Pass	
			Lead Glasses	4	Without Codes	Pass	

Medical Physicist

Mukhtar Al-Ansari

Medical physicist

Khaled Al-Radadi

Radiation Safety Officer

Faleh AL-Zaidi

Special Thanks for

Technical Assistant, Mr. Bandar Mahmoud



Medical Physics Department

Unit	Date of Test	Location of Test	Item	N.O	Code	Result	Action Plan
Fluoroscopy	2-9-2019	Fluoroscopy Room 4	Lead Apron	5	11100011327	Pass	
					11100011325	Pass	
					14402800128	Pass	
					14402800184	Pass	
					14402200040	Pass	
			Thyroid Collar	8	TO1 to TO8	Pass	
			Gonadal Shield	1	Set	Pass	

Medical Physicist

Mukhtar Al-Ansari

Medical physicist

Khaled Al-Radadi

Radiation Safety Officer

Faleh AL-Zaidi

Special Thanks for

Technical Assistant, Mr. Bandar Mahmoud



Medical Physics Department

Unit	Date of Test	Location of Test	Item	N.O	Code	Result	Action Plan
General X-Ray	2-9-2019	Fluoroscopy Room 4	Lead Apron	7	26502532	Pass	
					14402800196	Pass	
					65021-30	Pass	
					X1-102321	Pass	
					14402800080	Pass	
					14402800119	Pass	
					X110111132	Pass	
			Vest	6	76608222	Pass	
					092114	Pass	
					492114	Pass	
					02363	Pass	
					26608322	Pass	
					250415	Pass	
			Thyroid Collar	8	TO1 to TO8	Pass	
			Gonadal Shield	Mix	GO1 to GO5	Pass	
					2 Set		
					X11P2311		
					X11P2322		
					69096-2235		

Medical Physicist

Mukhtar Al-Ansari

Medical physicist

Khaled Al-Radadi

Radiation Safety Officer

Faleh AL-Zaidi

Special Thanks for

Technical Assistant, Mr. Bandar Mahmoud



Medical Physics Department

Unit	Date of Test	Location of Test	Item	N.O	Code	Result	Action Plan
Mammo	30-6-2019	Fluoroscopy Room 4	Lead Apron	3	SE09074060	Pass	
					SE09074059	Pass	
					SE09074066	Pass	
			Thyroid Collar	3	M1	Pass	
					M2	Pass	
					M3	Pass	
			Lead Glasses	1	Without Codes	Pass	

Medical Physicist

Mukhtar Al-Ansari

Medical physicist

Khaled Al-Radadi

Radiation Safety Officer

Faleh AL-Zaidi

Special Thanks for

Technical Assistant, Mr. Bandar Mahmoud



Medical Physics Department

Unit	Date of Test	Location of Test	Item	N.O	Code	Result	Action Plan
OR	10-4-2019	Fluoroscopy Room 4	Lead Apron	27	425321	Pass	
					527942	Pass	
					222956	Pass	
					7952234	Pass	
					245431	Pass	
					14178600018	Pass	
					542111	Pass	
					3992132	Pass	
					7213652	Pass	
					523472	Pass	
					S48622	Pass	
					621356	Pass	
					2015060159	Pass	
					S23476	Pass	
					3322550	Pass	
					552466	Pass	
					S48615	Pass	
					S48379	Pass	
					521431	Pass	
					S58749	Pass	
					522221	Pass	
					556662	Pass	
					6662132	Pass	
					7253612	Pass	
					14402800158	Fail	Remove
					14178700022	Fail	Remove
					13679200056	Fail	Remove
			Thyroid Collar	4	TO1 to TO4	Pass	
			Gonadal Shield	3	GO1 to GO3	Pass	

Medical Physicist

Mukhtar Al-Ansari

Medical physicist

Khaled Al-Radadi

Radiation Safety Officer

Faleh AL-Zaidi

Special Thanks for

Technical Assistant, Mr. Bandar Mahmoud



Medical Physics Department

Unit	Date of Test	Location of Test	Item	N.O	Code	Result	Action Plan
Urology	28-8-2019	Fluoroscopy Room 4	Lead Apron	5	306739	Pass	
					68096-22	Pass	
					68996-2252	Pass	
					306738	Pass	
					67089-22	Pass	
			Thyroid Collar	2	TU1	Pass	
					TU2	Pass	

Medical Physicist

Mukhtar Al-Ansari

Medical physicist


Khaled Al-Radadi

Radiation Safety Officer

Faleh AL-Zaidi

Special Thanks for

Technical Assistant, Mr. Bandar Mahmoud



Medical Physics Department

قسم الفيزياء الطبية

Statistics about Radiation Protections Tools

2019								
Department		Lead Apron	Vest	Skirt	Gonadal	Thyroid	Cover	Glasses
Radiology	General X-Ray Mobile X-Ray	7	6	0	8 + 2 set	8	0	0
	CT	4	0	0	0	3	0	0
	Fluoroscopy	5	0	0	1 set	8	0	0
	Angio	8	3	2	3	11	0	8
	Mammo	3	0	0	0	3	0	1
OR		27	0	0	3	4	0	0
Endoscopy		10	0	0	0	4	0	4
Urology		5	0	0	0	2	0	0
Total		67	9	2	14 + set	43	0	13



وزارة الصحة
Ministry of Health

المدينة المنورة مستشفى الملك فهد (الحماية من الإشعاع والفيزياء الطبية)

سعادة مدير ادارة صحة البيئة والصحة المهنية بالمدينة المنورة المحترم

السلام عليكم ورحمة الله وبركاته وبعد

إشارة الى ما ورد في خطابكم رقم 1439-1275492 وتاريخ 2018 - 19-03 بشأن حصر المرايل المرخصة
التالفة والرجيع والرفع لنا وتزويدنا بأعدادها ليتسنى لنا اكمال اللازم

عليه نفيد سعادتك بعدد المرايل المرخصة الرجيع والتالفة بمجمع مستشفى الملك فهد .

Item	N.O	Code	Result	السبب
Lead Apron	12	SE09074061	Fail	
		13679200056	Fail	
		14402800158	Fail	
		14178700022	Fail	
		09074053	Fail	
		09074051	Fail	
		09074072	Fail	
		09074070	Fail	
		306738	Fail	
		401120	Fail	
		67213562	Fail	
		05021TC-22	Fail	
Vest	00	00	00	
Skirt	00	00	00	
Thyroid Collar	1	A5	Fail	
Conadal Shield	00	00	00	
Lead Glasses	1	00	Fail	

Medical Physicist

Medical Physicist

Radiation Safety Officer

Mukhtar Al-Ansari


Khaled Al-Radadi

Faleh AL-Zaidi

2019
26/11

K

26-11-2019



Medical Physics Department

قسم الفيزياء الطبية

Date of Radiation Protection Tools Tests – 2019

Department	Tested in	Next Due Test
OR	10-4-2019	10-4-2020
Endoscopy	17-6-2019	17-6-2020
Mammo	30-6-2019	30-6-2020
Mobile X-Ray General X-Ray Fluoroscopy	2-9-2019	2-9-2020
Angio	13-11-2019	13-11-2020
ESWL	28-8-2019	28-8-2020
CT Scan	1-9-2019	1-9-2020