



Certificate No. : L0372-170120

財團法人全國認證基金會  
Taiwan Accreditation Foundation

## Certificate of Accreditation

This is to certify that

**TUNG PEI INDUSTRIAL CO., LTD.**

**Length Calibration Laboratory**

No. 7, Songjiang N. Rd., Zhongli Dist., Taoyuan City 320, Taiwan

**is accredited in respect of laboratory**

**Accreditation Criteria** : ISO/IEC 17025:2005  
**Accreditation Number** : 0372  
**Originally Accredited** : January 01, 1999  
**Effective Period** : January 20, 2017 to January 19, 2020  
**Accredited Scope** : Calibration Field, see described in the Appendix

*Chung-Lin Wang*

Chung-Lin Wang  
President, Taiwan Accreditation Foundation  
Date : January 20, 2017





證書編號：L0372-170120

財團法人全國認證基金會  
Taiwan Accreditation Foundation

## 認 證 證 書

茲證明

東培工業股份有限公司

長度校正實驗室

桃園市中壢區松江北路 7 號(中壢工業區)

為本會認證之實驗室

認 證 依 據：ISO/IEC 17025：2005

認 證 編 號：0372

初次認證日期：八十八年一月一日

認證有效期間：一百零六年一月二十日至一百零九年一月十九日止

認 證 範 圍：校正領域，如續頁

董事長

王聰麟

中華民國一百零六年一月二十日



證書編號：L0372-170120

財團法人全國認證基金會  
Taiwan Accreditation Foundation

認證編號：0372

實驗室主管：林鶴年

長度

項目代碼/ 校正件	最高 工作標準件	校正方法	校正範圍				量測條件	最小 不確定度	
	廠牌/型號		最小 範圍	單位	最大 範圍	單位		數值	單位
KA1001 塊規 (CNS 8092 / 112 片) (鋼質)	P.T.W B-1 Grade 00	自訂塊規校 正程序指導 書 (文件編號: Q10IT901A)	0.5	mm	100	mm		$[(91)^2 + (1.8 L)^2]^{1/2}$	nm
								L: mm, length of gauge block	

註：最小不確定度係以約 95 %信賴水準之擴充不確定度表示

簽署人姓名	簽署範圍
林鶴年	KA1001
蔡永傑	KA1001

(以下空白)

