

RAISE LAB EQUIPMENT

TAP DENSITY - TEST REPORT

RAM

Model No	THYY	Serial No	THY
Location	dd	Instrument No	qq
Last Validation Date	18-05-2026	Next Validation Due	18-08-2026

VALIDATION DETAILS

USP 1 validation			
Date / Time	18/05/2026 17:23:07		
USP	USP 1	Taps/Min	300
Drop Height (mm)	14	Status	Pass
Expected Tap Count	300 (+/- 15)	Actual Tap Count	302
USP 2 validation			
Date / Time	18/05/2026 17:24:26		
USP	USP 2	Taps/Min	250
Drop Height (mm)	3	Status	Pass
Expected Tap Count	250 (+/- 15)	Actual Tap Count	250

REMARKS

N/A			
Operated By	rahul	Employee ID	Rahul
Approved By	Factory (factory)	Pass / Fail	PASS
Approval remarks	gyvyh		