

RAISE LAB EQUIPMENT

TAP DENSITY - TEST REPORT

RAM

Model No	THYY	Serial No	THY
Location	dd	Instrument No	qq
Last Validation Date	20-05-2026	Next Validation Due	20-08-2026

VALIDATION DETAILS

Date / Time	20/05/2026 18:21:31		
USP	USP 2	Taps/Min	250
Drop Height (mm)	3	Status	Pass
Expected Tap Count	250 (+/- 50)	Actual Tap Count	250

REMARKS

N/A

Operated By	Test (comma)	Employee ID	,
Approved By	--	Pass / Fail	--
Approval remarks	N/A		

