

Awaiting approval. A reviewer must verify Pass/Fail on this screen before you can leave or print. Logging out will abort this report.

RAISE LAB EQUIPMENT
TAP DENSITY - TEST REPORT
RAM

Model No	THYY	Serial No	THY
Location	dd	Instrument No	qq
Last Validation Date	20-05-2026	Next Validation Due	20-08-2026

VALIDATION DETAILS

USP 1 validation			
Date / Time	20/05/2026 19:16:05		
USP	USP 1	Taps/Min	300
Drop Height (mm)	14	Status	Fail
Expected Tap Count	300 (+/- 15)	Actual Tap Count	272
USP 2 validation			
Date / Time	20/05/2026 19:13:02		
USP	USP 2	Taps/Min	250
Drop Height (mm)	3	Status	Pass
Expected Tap Count	250 (+/- 15)	Actual Tap Count	249

REMARKS

N/A			
Operated By	Test@1234	Employee ID	Test@1234
Approved By	--	Pass / Fail	--
Approval remarks	N/A		