

RapidCTQA Clinical Report

Generated on: 2026-05-07 10:27:11

Patient Information

Patient Name: VARGA^ANDRÁSNE
Series UID: 1.3.12.2.1107.5.1.7.128248.30000026050409150055800000439
Protocol: RTP Head and Neck Fej-Nyak
Final Status: CONDITIONAL

Agent Findings & Flags

Agent	Status	Message
Integrity	CONDITIONAL	Slice thickness exceeds preferred limit (3mm)

Quantitative Metrics

series_uid: 1.3.12.2.1107.5.1.7.128248.30000026050409150055800000439VARGA^ANDRÁSNE	
protocol: RTP Head and Neck Fej-Nyak	slice_count: 215
slice_thickness: 5.000	slice_spacing_var: 0.000
monotonic_z: True	duplicate_slices: False
gantry_tilt: 0.000	truncation_detected: False
background_air_sd: 0.000	center_noise_std: 205.319
air_hu_estimate: -1010.000	water_hu_estimate: -43.000
fluid_median_hu: 34.000	gas_volume_cc: 0.000
rescale_slope: 1.000	