

RapidCTQA Clinical Report

Generated on: 2026-05-28 14:40:33

Patient Information

Patient Name:	SOLTÉSZ^JÓZSEF
Series UID:	1.3.12.2.1107.5.1.7.128248.30000026042213345632400000244
Protocol:	RTP Head 120KV Koponya
Final Status:	CONDITIONAL

Agent Findings & Flags

Agent	Status	Message
ImplantAuditor	CONDITIONAL	METAL_IMPLANT: High-density metal detected (4.76 cc). Verify implant/cardiac de

Quantitative Metrics

series_uid: 1.3.12.2.1107.5.1.7.128248.30000026042213345632400000244SOLTÉSZ^JÓZSEF	
protocol: RTP Head 120KV Koponya	slice_count: 180
slice_thickness: 2.000	slice_spacing_var: 0.000
monotonic_z: True	duplicate_slices: False
gantry_tilt: 0.000	truncation_detected: False
background_air_sd: 0.000	center_noise_std: 296.016
air_hu_estimate: -1015.000	water_hu_estimate: 33.000
fluid_median_hu: 35.000	gas_volume_cc: 0.000
metal_detected: True	metal_volume_cc: 4.763
pediatric_mismatch: False	pediatric_mismatch_message:
rescale_slope: 1.000	