

RapidCTQA Clinical Report

Generated on: 2026-06-02 14:30:40

Patient Information

Patient Name:	KÁDÁR^SÁNDOR
Series UID:	1.3.12.2.1107.5.1.7.128248.30000026042409102746200000290
Protocol:	RTP Pelvis Medence
Final Status:	CONDITIONAL

Agent Findings & Flags

Agent	Status	Message
AlignmentAuditor	CONDITIONAL	CRITICAL_TILT_ALERT: Patient roll exceeds tolerance (15.1°) (Slices 1-5, 12-13,
AlignmentAuditor	CONDITIONAL	TWIST_YAW_ALERT: Significant spinal twist detected (>1.0° over 10cm) (Slices 1

Quantitative Metrics

series_uid: 1.3.12.2.1107.5.1.7.128248.30000026042409102746200000290KÁDÁR^SÁNDOR	
protocol: RTP Pelvis Medence	slice_count: 172
slice_thickness: 3.000	slice_spacing_var: 0.000
monotonic_z: True	duplicate_slices: False
gantry_tilt: 0.000	truncation_detected: False
background_air_sd: 0.000	center_noise_std: 290.636
air_hu_estimate: -1014.000	water_hu_estimate: 33.000
fluid_median_hu: 34.000	gas_volume_cc: 0.000
gas_slices: []	metal_detected: False
metal_volume_cc: 0.074	metal_slices: []
truncated_slices: []	max_roll_deg: 15.102
roll_slices: [1, 2, 3, 4, 5, 12, 13, 19, 20, 21, 22, 23, 24, 25, 27, 28, 29, 30, 31, 32, 33, 34, 35, 36, 37, 38, 39, 40, 41, 42, 43, 44, 45, 46, 47, 48, 49, 50, 51, 52, 53, 54, 55, 56, 57, 58, 59, 60, 61, 62, 63, 64, 65, 66, 67, 68, 69, 70, 71, 72, 73, 74, 75, 76, 77, 82, 83, 84, 85, 86, 87, 8	roll_deg: 15.102
twist_slices: [1, 2, 3, 4, 5, 6, 7, 8, 9, 10, 11, 12, 13, 14, 15, 16, 17, 18, 19, 20, 21, 22, 23, 24, 25, 26, 27, 28, 29, 30, 31, 32, 33, 3	twist_deg: 1.0
pediatric_mismatch_message:	rescale_slope: 1.000