

## DH medical lab WO Event Results

E45-WO-1

\* Peer Data \*

| Target                     | Reported Result | Expected Result | Performance    | Avg CT | SD   | Your CT |
|----------------------------|-----------------|-----------------|----------------|--------|------|---------|
| Staphylococcus aureus      | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Acinetobacter baumannii    | Detected        | Detected        | Acceptable     | 16.77  | 0.91 | 16.58   |
| Enterococcus faecalis      | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Escherichia coli           | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Klebsiella pneumoniae      | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Pseudomonas aeruginosa     | Detected        | Detected        | Acceptable     | 19.35  | 0.6  | 19.63   |
| Klebsiella oxytoca         | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Enterobacter cloacae       | Detected        | Detected        | Acceptable     | 17.31  | 1.71 | 17.99   |
| Proteus mirabilis          | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Morganella morganii        | Detected        | Detected        | Acceptable     | 21.19  | 2.36 | 17.74   |
| Proteus vulgaris           | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Citrobacter freundii       | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Enterococcus faecium       | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Bacteroides fragilis       | Detected        | Detected        | Acceptable     | 20.09  | 1.24 | 18.78   |
| Staphylococcus epidermidis | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| Clostridium novyi          | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Clostridium perfringens    | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Clostridium septicum       | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Kingella kingae            | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Group B Strep              | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| Group A Strep              | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Group C&G; Strep           | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| ampC                       | Not Detected    | Detected        | Not Acceptable | 18.3   | 0.94 | -       |
| mecA                       | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| femA                       | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| aph2                       | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| QnrB                       | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| QnrA                       | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| vanA1, vanA2               | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |

## E45-WO-1

\* Peer Data \*

| Target                 | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|------------------------|-----------------|-----------------|-------------|--------|------|---------|
| vanB                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| NDM                    | Detected        | Detected        | Acceptable  | 24.4   | 1.08 | 22.98   |
| KPC                    | Detected        | Detected        | Acceptable  | 18.29  | 1.76 | 16.31   |
| OXA 48                 | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| VIM, IMP-7             | Detected        | Detected        | Acceptable  | 20.01  | 2.18 | 18.95   |
| SHV                    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| TEM                    | Detected        | Detected        | Acceptable  | 18.28  | 1.16 | 16.8    |
| CTX-M group 1          | Detected        | Detected        | Acceptable  | 29.89  | 1.24 | 31.35   |
| CTX-M group 2          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| mefA                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ErmA                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ErmB                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ant2                   | Not Tested      | Detected        | Not Graded  | 19.43  | 2.65 | -       |
| DfrA                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| DfrA1                  | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| DfrA5                  | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetO                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Sul1                   | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| Sul2                   | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| GyrA                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| ParC                   | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| TetM                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetB                   | Not Tested      | Detected        | Not Graded  | 19.61  | 1.97 | -       |
| Antla                  | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| aph3                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Enterobacter aerogenes | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |

## E45-WO-2

\* Peer Data \*

| Target                | Reported Result | Expected Result | Performance | Avg CT | SD | Your CT |
|-----------------------|-----------------|-----------------|-------------|--------|----|---------|
| Staphylococcus aureus | Not Detected    | Not Detected    | Acceptable  | -      | -  | -       |

## E45-WO-2

\* Peer Data \*

| Target                     | Reported Result | Expected Result | Performance    | Avg CT | SD   | Your CT |
|----------------------------|-----------------|-----------------|----------------|--------|------|---------|
| Acinetobacter baumannii    | Detected        | Detected        | Acceptable     | 29.1   | 1.84 | 27.51   |
| Enterococcus faecalis      | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Escherichia coli           | Detected        | Detected        | Acceptable     | 17.03  | 0.71 | 17.02   |
| Klebsiella pneumoniae      | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Pseudomonas aeruginosa     | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Klebsiella oxytoca         | Detected        | Detected        | Acceptable     | 19.66  | 1.17 | 19.46   |
| Enterobacter cloacae       | Detected        | Detected        | Acceptable     | 17.44  | 1.85 | 18.27   |
| Proteus mirabilis          | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Morganella morganii        | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Proteus vulgaris           | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Citrobacter freundii       | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Enterococcus faecium       | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Bacteroides fragilis       | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Staphylococcus epidermidis | Not Tested      | Detected        | Not Graded     | -      | -    | -       |
| Clostridium novyi          | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Clostridium perfringens    | Detected        | Detected        | Acceptable     | 21.46  | 0.93 | 20.78   |
| Clostridium septicum       | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Kingella kingae            | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Group B Strep              | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| Group A Strep              | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Group C&G; Strep           | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| ampC                       | Not Detected    | Detected        | Not Acceptable | 18.11  | 0.89 | -       |
| mecA                       | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| femA                       | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| aph2                       | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| QnrB                       | Detected        | Detected        | Acceptable     | 18.91  | 0.77 | 18.68   |
| QnrA                       | Detected        | Detected        | Acceptable     | 20.17  | 2.19 | 19.41   |
| vanA1, vanA2               | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| vanB                       | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| NDM                        | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| KPC                        | Detected        | Detected        | Acceptable     | 16.66  | 1.31 | 14.96   |

## E45-WO-2

\* Peer Data \*

| Target                 | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|------------------------|-----------------|-----------------|-------------|--------|------|---------|
| OXA 48                 | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| VIM, IMP-7             | Detected        | Detected        | Acceptable  | 18.61  | 1.11 | 17.61   |
| SHV                    | Detected        | Detected        | Acceptable  | 17.07  | 1.68 | 15.09   |
| TEM                    | Detected        | Detected        | Acceptable  | 16.33  | 1.21 | 14.79   |
| CTX-M group 1          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| CTX-M group 2          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| mefA                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ErmA                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ErmB                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ant2                   | Not Tested      | Detected        | Not Graded  | 17.41  | 2.67 | -       |
| DfrA                   | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| DfrA1                  | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| DfrA5                  | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| TetO                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Sul1                   | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| Sul2                   | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| GyrA                   | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| ParC                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetM                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetB                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Antla                  | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| aph3                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Enterobacter aerogenes | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |

## E45-WO-3

\* Peer Data \*

| Target                  | Reported Result | Expected Result | Performance | Avg CT | SD | Your CT |
|-------------------------|-----------------|-----------------|-------------|--------|----|---------|
| Staphylococcus aureus   | Not Detected    | Not Detected    | Acceptable  | -      | -  | -       |
| Acinetobacter baumannii | Not Detected    | Not Detected    | Acceptable  | -      | -  | -       |
| Enterococcus faecalis   | Not Detected    | Not Detected    | Acceptable  | -      | -  | -       |
| Escherichia coli        | Not Detected    | Not Detected    | Acceptable  | -      | -  | -       |

## E45-WO-3

\* Peer Data \*

| Target                     | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|----------------------------|-----------------|-----------------|-------------|--------|------|---------|
| Klebsiella pneumoniae      | Detected        | Detected        | Acceptable  | 19.04  | 1.25 | 19.44   |
| Psuedomonas aeruginosa     | Detected        | Detected        | Acceptable  | 20.01  | 0.52 | 20.63   |
| Klebsiella oxytoca         | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterobacter cloacae       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Proteus mirabilis          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Morganella morganii        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Proteus vulgaris           | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Citrobacter freundii       | Detected        | Detected        | Acceptable  | 21.35  | 0.74 | 21.3    |
| Enterococcus faecium       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Bacteroides fragilis       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Staphylococcus epidermidis | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Clostridium novyi          | Detected        | Detected        | Acceptable  | 23.44  | 2.39 | 22.59   |
| Clostridium perfringens    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Clostridium septicum       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Kingella kingae            | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Group B Strep              | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Group A Strep              | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Group C&G; Strep           | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| ampC                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| mecA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| femA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| aph2                       | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| QnrB                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| QnrA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| vanA1, vanA2               | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| vanB                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| NDM                        | Detected        | Detected        | Acceptable  | 24.16  | 1.13 | 23.22   |
| KPC                        | Detected        | Detected        | Acceptable  | 17.91  | 1.94 | 16.62   |
| OXA 48                     | Detected        | Detected        | Acceptable  | 16.8   | 0.84 | 16.82   |
| VIM, IMP-7                 | Detected        | Detected        | Acceptable  | 23.22  | 4.37 | 24.73   |
| SHV                        | Detected        | Detected        | Acceptable  | 17.31  | 1.48 | 15.73   |

## E45-WO-3

\* Peer Data \*

| Target                 | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|------------------------|-----------------|-----------------|-------------|--------|------|---------|
| TEM                    | Detected        | Detected        | Acceptable  | 17.89  | 1.14 | 17.0    |
| CTX-M group 1          | Detected        | Detected        | Acceptable  | 22.04  | 1.29 | 24.05   |
| CTX-M group 2          | Detected        | Detected        | Acceptable  | 21.1   | 0.95 | 21.03   |
| mefA                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ErmA                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ErmB                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ant2                   | Not Tested      | Detected        | Not Graded  | 24.33  | 0.32 | -       |
| DfrA                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| DfrA1                  | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| DfrA5                  | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetO                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Sul1                   | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| Sul2                   | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| GyrA                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| ParC                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetM                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetB                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Antla                  | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| aph3                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Enterobacter aerogenes | Detected        | Detected        | Acceptable  | 19.32  | 1.95 | 20.4    |

## E45-WO-4

\* Peer Data \*

| Target                  | Reported Result | Expected Result | Performance    | Avg CT | SD   | Your CT |
|-------------------------|-----------------|-----------------|----------------|--------|------|---------|
| Staphylococcus aureus   | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Acinetobacter baumannii | Detected        | Not Detected    | Not Acceptable | -      | -    | 33.2    |
| Enterococcus faecalis   | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Escherichia coli        | Detected        | Detected        | Acceptable     | 18.14  | 1.76 | 17.81   |
| Klebsiella pneumoniae   | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Pseudomonas aeruginosa  | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Klebsiella oxytoca      | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |

## E45-WO-4

\* Peer Data \*

| Target                     | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|----------------------------|-----------------|-----------------|-------------|--------|------|---------|
| Enterobacter cloacae       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Proteus mirabilis          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Morganella morganii        | Detected        | Detected        | Acceptable  | 22.09  | 2.13 | 18.96   |
| Proteus vulgaris           | Detected        | Detected        | Acceptable  | 23.36  | 1.37 | 24.69   |
| Citrobacter freundii       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterococcus faecium       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Bacteroides fragilis       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Staphylococcus epidermidis | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Clostridium novyi          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Clostridium perfringens    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Clostridium septicum       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Kingella kingae            | Detected        | Detected        | Acceptable  | 26.09  | 0.72 | 26.08   |
| Group B Strep              | Not Tested      | Detected        | Not Graded  | 20.13  | 1.09 | -       |
| Group A Strep              | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Group C&G; Strep           | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| ampC                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| mecA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| femA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| aph2                       | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| QnrB                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| QnrA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| vanA1, vanA2               | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| vanB                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| NDM                        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| KPC                        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| OXA 48                     | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| VIM, IMP-7                 | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| SHV                        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| TEM                        | Detected        | Detected        | Acceptable  | 15.32  | 1.62 | 13.95   |
| CTX-M group 1              | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| CTX-M group 2              | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |

## E45-WO-4

\* Peer Data \*

| Target                 | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|------------------------|-----------------|-----------------|-------------|--------|------|---------|
| mefA                   | Detected        | Detected        | Acceptable  | 13.78  | 1.47 | 14.86   |
| ErmA                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ErmB                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ant2                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| DfrA                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| DfrA1                  | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| DfrA5                  | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetO                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Sul1                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Sul2                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| GyrA                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| ParC                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetM                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetB                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Antla                  | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| aph3                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Enterobacter aerogenes | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |

## E45-WO-5

\* Peer Data \*

| Target                  | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|-------------------------|-----------------|-----------------|-------------|--------|------|---------|
| Staphylococcus aureus   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Acinetobacter baumannii | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterococcus faecalis   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Escherichia coli        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Klebsiella pneumoniae   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Pseudomonas aeruginosa  | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Klebsiella oxytoca      | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterobacter cloacae    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Proteus mirabilis       | Detected        | Detected        | Acceptable  | 17.25  | 1.33 | 17.47   |
| Morganella morganii     | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |



## E45-WO-5

\* Peer Data \*

| Target                     | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|----------------------------|-----------------|-----------------|-------------|--------|------|---------|
| Proteus vulgaris           | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Citrobacter freundii       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterococcus faecium       | Detected        | Detected        | Acceptable  | 17.43  | 1.84 | 17.97   |
| Bacteroides fragilis       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Staphylococcus epidermidis | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Clostridium novyi          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Clostridium perfringens    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Clostridium septicum       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Kingella kingae            | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Group B Strep              | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Group A Strep              | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Group C&G; Strep           | Not Tested      | Detected        | Not Graded  | 21.31  | 1.25 | -       |
| ampC                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| mecA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| femA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| aph2                       | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| QnrB                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| QnrA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| vanA1, vanA2               | Detected        | Detected        | Acceptable  | 18.51  | 1.94 | 19.4    |
| vanB                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| NDM                        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| KPC                        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| OXA 48                     | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| VIM, IMP-7                 | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| SHV                        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| TEM                        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| CTX-M group 1              | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| CTX-M group 2              | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| mefA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ErmA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ErmB                       | Detected        | Detected        | Acceptable  | 17.81  | 1.89 | 16.11   |

## E45-WO-5

\* Peer Data \*

| Target                 | Reported Result | Expected Result | Performance | Avg CT | SD | Your CT |
|------------------------|-----------------|-----------------|-------------|--------|----|---------|
| ant2                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| DfrA                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| DfrA1                  | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| DfrA5                  | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| TetO                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| Sul1                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| Sul2                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| GyrA                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| ParC                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| TetM                   | Not Tested      | Detected        | Not Graded  | -      | -  | -       |
| TetB                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| Antla                  | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| aph3                   | Not Tested      | Detected        | Not Graded  | -      | -  | -       |
| Enterobacter aerogenes | Not Detected    | Not Detected    | Acceptable  | -      | -  | -       |

## E45-WO-6

\* Peer Data \*

| Target                  | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|-------------------------|-----------------|-----------------|-------------|--------|------|---------|
| Staphylococcus aureus   | Detected        | Detected        | Acceptable  | 18.92  | 1.48 | 17.58   |
| Acinetobacter baumannii | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterococcus faecalis   | Detected        | Detected        | Acceptable  | 18.1   | 2.2  | 16.49   |
| Escherichia coli        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Klebsiella pneumoniae   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Pseudomonas aeruginosa  | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Klebsiella oxytoca      | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterobacter cloacae    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Proteus mirabilis       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Morganella morganii     | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Proteus vulgaris        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Citrobacter freundii    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterococcus faecium    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |

## E45-WO-6

\* Peer Data \*

| Target                     | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|----------------------------|-----------------|-----------------|-------------|--------|------|---------|
| Bacteroides fragilis       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Staphylococcus epidermidis | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Clostridium novyi          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Clostridium perfringens    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Clostridium septicum       | Detected        | Detected        | Acceptable  | 24.32  | 1.01 | 24.28   |
| Kingella kingae            | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Group B Strep              | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Group A Strep              | Detected        | Detected        | Acceptable  | 19.37  | 0.69 | 18.74   |
| Group C&G; Strep           | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| ampC                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| mecA                       | Detected        | Detected        | Acceptable  | 20.34  | 2.09 | 17.82   |
| femA                       | Detected        | Detected        | Acceptable  | 19.39  | 1.32 | 18.01   |
| aph2                       | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| QnrB                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| QnrA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| vanA1, vanA2               | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| vanB                       | Detected        | Detected        | Acceptable  | 18.16  | 2.2  | 16.49   |
| NDM                        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| KPC                        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| OXA 48                     | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| VIM, IMP-7                 | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| SHV                        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| TEM                        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| CTX-M group 1              | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| CTX-M group 2              | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| mefA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ErmA                       | Detected        | Detected        | Acceptable  | 18.37  | 1.79 | 16.67   |
| ErmB                       | Detected        | Detected        | Acceptable  | 17.81  | 2.72 | 14.68   |
| ant2                       | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| DfrA                       | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| DfrA1                      | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |

## E45-WO-6

\* Peer Data \*

| Target                 | Reported Result | Expected Result | Performance | Avg CT | SD | Your CT |
|------------------------|-----------------|-----------------|-------------|--------|----|---------|
| DfrA5                  | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| TetO                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| Sul1                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| Sul2                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| GyrA                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| ParC                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| TetM                   | Not Tested      | Detected        | Not Graded  | -      | -  | -       |
| TetB                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| Antla                  | Not Tested      | Detected        | Not Graded  | -      | -  | -       |
| aph3                   | Not Tested      | Detected        | Not Graded  | -      | -  | -       |
| Enterobacter aerogenes | Not Detected    | Not Detected    | Acceptable  | -      | -  | -       |

## E45-WO-7

\* Peer Data \*

| Target                     | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|----------------------------|-----------------|-----------------|-------------|--------|------|---------|
| Staphylococcus aureus      | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Acinetobacter baumannii    | Detected        | Detected        | Acceptable  | 18.54  | 1.39 | 16.14   |
| Enterococcus faecalis      | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Escherichia coli           | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Klebsiella pneumoniae      | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Pseudomonas aeruginosa     | Detected        | Detected        | Acceptable  | 19.81  | 0.62 | 18.78   |
| Klebsiella oxytoca         | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterobacter cloacae       | Detected        | Detected        | Acceptable  | 18.97  | 1.66 | 18.77   |
| Proteus mirabilis          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Morganella morganii        | Detected        | Detected        | Acceptable  | 21.62  | 3.54 | 16.34   |
| Proteus vulgaris           | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Citrobacter freundii       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterococcus faecium       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Bacteroides fragilis       | Detected        | Detected        | Acceptable  | 21.13  | 1.65 | 19.17   |
| Staphylococcus epidermidis | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Clostridium novyi          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |

## E45-WO-7

\* Peer Data \*

| Target                  | Reported Result | Expected Result | Performance    | Avg CT | SD   | Your CT |
|-------------------------|-----------------|-----------------|----------------|--------|------|---------|
| Clostridium perfringens | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Clostridium septicum    | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Kingella kingae         | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Group B Strep           | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| Group A Strep           | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Group C&G; Strep        | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| ampC                    | Not Detected    | Detected        | Not Acceptable | 19.9   | 1.39 | -       |
| mecA                    | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| femA                    | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| aph2                    | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| QnrB                    | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| QnrA                    | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| vanA1, vanA2            | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| vanB                    | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| NDM                     | Detected        | Detected        | Acceptable     | 25.62  | 1.27 | 23.48   |
| KPC                     | Detected        | Detected        | Acceptable     | 20.56  | 1.76 | 18.14   |
| OXA 48                  | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| VIM, IMP-7              | Detected        | Detected        | Acceptable     | 20.59  | 2.67 | 18.96   |
| SHV                     | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| TEM                     | Detected        | Detected        | Acceptable     | 19.74  | 1.94 | 17.42   |
| CTX-M group 1           | Detected        | Detected        | Acceptable     | 30.99  | 0.39 | 31.52   |
| CTX-M group 2           | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| mefA                    | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| ErmA                    | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| ErmB                    | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| ant2                    | Not Tested      | Detected        | Not Graded     | 21.14  | 1.63 | -       |
| DfrA                    | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| DfrA1                   | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| DfrA5                   | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| TetO                    | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| Sul1                    | Not Tested      | Detected        | Not Graded     | -      | -    | -       |

## E45-WO-7

\* Peer Data \*

| Target                 | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|------------------------|-----------------|-----------------|-------------|--------|------|---------|
| Sul2                   | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| GyrA                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| ParC                   | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| TetM                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetB                   | Not Tested      | Detected        | Not Graded  | 20.98  | 1.17 | -       |
| Antla                  | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| aph3                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Enterobacter aerogenes | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |

## E45-WO-8

\* Peer Data \*

| Target                     | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|----------------------------|-----------------|-----------------|-------------|--------|------|---------|
| Staphylococcus aureus      | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Acinetobacter baumannii    | Detected        | Detected        | Acceptable  | 30.26  | 0.67 | 30.03   |
| Enterococcus faecalis      | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Escherichia coli           | Detected        | Detected        | Acceptable  | 19.25  | 0.63 | 19.03   |
| Klebsiella pneumoniae      | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Pseudomonas aeruginosa     | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Klebsiella oxytoca         | Detected        | Detected        | Acceptable  | 22.29  | 0.83 | 22.18   |
| Enterobacter cloacae       | Detected        | Detected        | Acceptable  | 19.4   | 2.06 | 20.81   |
| Proteus mirabilis          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Morganella morganii        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Proteus vulgaris           | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Citrobacter freundii       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterococcus faecium       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Bacteroides fragilis       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Staphylococcus epidermidis | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| Clostridium novyi          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Clostridium perfringens    | Detected        | Detected        | Acceptable  | 23.7   | 1.35 | 24.72   |
| Clostridium septicum       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Kingella kingae            | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |

## E45-WO-8

\* Peer Data \*

| Target           | Reported Result | Expected Result | Performance    | Avg CT | SD   | Your CT |
|------------------|-----------------|-----------------|----------------|--------|------|---------|
| Group B Strep    | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| Group A Strep    | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| Group C&G; Strep | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| ampC             | Not Detected    | Detected        | Not Acceptable | 19.95  | 1.44 | -       |
| mecA             | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| femA             | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| aph2             | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| QnrB             | Detected        | Detected        | Acceptable     | 21.31  | 0.97 | 21.63   |
| QnrA             | Detected        | Detected        | Acceptable     | 22.44  | 0.94 | 23.01   |
| vanA1, vanA2     | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| vanB             | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| NDM              | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| KPC              | Detected        | Detected        | Acceptable     | 19.57  | 0.81 | 18.37   |
| OXA 48           | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| VIM, IMP-7       | Detected        | Detected        | Acceptable     | 20.82  | 2.17 | 21.25   |
| SHV              | Detected        | Detected        | Acceptable     | 19.62  | 1.0  | 18.11   |
| TEM              | Detected        | Detected        | Acceptable     | 18.69  | 1.16 | 17.21   |
| CTX-M group 1    | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| CTX-M group 2    | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| mefA             | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| ErmA             | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| ErmB             | Not Detected    | Not Detected    | Acceptable     | -      | -    | -       |
| ant2             | Not Tested      | Detected        | Not Graded     | 20.27  | 1.73 | -       |
| DfrA             | Not Tested      | Detected        | Not Graded     | -      | -    | -       |
| DfrA1            | Not Tested      | Detected        | Not Graded     | -      | -    | -       |
| DfrA5            | Not Tested      | Detected        | Not Graded     | -      | -    | -       |
| TetO             | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |
| Sul1             | Not Tested      | Detected        | Not Graded     | -      | -    | -       |
| Sul2             | Not Tested      | Detected        | Not Graded     | -      | -    | -       |
| GyrA             | Not Tested      | Detected        | Not Graded     | -      | -    | -       |
| ParC             | Not Tested      | Not Detected    | Not Graded     | -      | -    | -       |

## E45-WO-8

\* Peer Data \*

| Target                 | Reported Result | Expected Result | Performance | Avg CT | SD | Your CT |
|------------------------|-----------------|-----------------|-------------|--------|----|---------|
| TetM                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| TetB                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| Antla                  | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| aph3                   | Not Tested      | Not Detected    | Not Graded  | -      | -  | -       |
| Enterobacter aerogenes | Not Detected    | Not Detected    | Acceptable  | -      | -  | -       |

## E45-WO-9

\* Peer Data \*

| Target                     | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|----------------------------|-----------------|-----------------|-------------|--------|------|---------|
| Staphylococcus aureus      | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Acinetobacter baumannii    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterococcus faecalis      | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Escherichia coli           | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Klebsiella pneumoniae      | Detected        | Detected        | Acceptable  | 19.88  | 0.59 | 19.44   |
| Pseudomonas aeruginosa     | Detected        | Detected        | Acceptable  | 18.85  | 0.81 | 17.81   |
| Klebsiella oxytoca         | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterobacter cloacae       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Proteus mirabilis          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Morganella morganii        | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Proteus vulgaris           | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Citrobacter freundii       | Detected        | Detected        | Acceptable  | 22.56  | 0.64 | 21.69   |
| Enterococcus faecium       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Bacteroides fragilis       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Staphylococcus epidermidis | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Clostridium novyi          | Detected        | Detected        | Acceptable  | 22.64  | 1.06 | 21.62   |
| Clostridium perfringens    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Clostridium septicum       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Kingella kingae            | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Group B Strep              | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Group A Strep              | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Group C&G; Strep           | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |



## E45-WO-9

\* Peer Data \*

| Target        | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|---------------|-----------------|-----------------|-------------|--------|------|---------|
| ampC          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| mecA          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| femA          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| aph2          | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| QnrB          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| QnrA          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| vanA1, vanA2  | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| vanB          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| NDM           | Detected        | Detected        | Acceptable  | 23.82  | 1.77 | 21.67   |
| KPC           | Detected        | Detected        | Acceptable  | 19.92  | 1.72 | 17.86   |
| OXA 48        | Detected        | Detected        | Acceptable  | 18.25  | 1.05 | 17.98   |
| VIM, IMP-7    | Detected        | Detected        | Acceptable  | 25.52  | 1.64 | 24.0    |
| SHV           | Detected        | Detected        | Acceptable  | 19.11  | 1.67 | 16.99   |
| TEM           | Detected        | Detected        | Acceptable  | 19.54  | 1.86 | 17.79   |
| CTX-M group 1 | Detected        | Detected        | Acceptable  | 23.98  | 1.29 | 25.88   |
| CTX-M group 2 | Detected        | Detected        | Acceptable  | 22.08  | 0.79 | 21.14   |
| mefA          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ErmA          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ErmB          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ant2          | Not Tested      | Detected        | Not Graded  | 22.76  | 0.06 | -       |
| DfrA          | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| DfrA1         | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| DfrA5         | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetO          | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Sul1          | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| Sul2          | Not Tested      | Detected        | Not Graded  | -      | -    | -       |
| GyrA          | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| ParC          | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetM          | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetB          | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Antla         | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |

## E45-WO-9

\* Peer Data \*

| Target                 | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|------------------------|-----------------|-----------------|-------------|--------|------|---------|
| aph3                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Enterobacter aerogenes | Detected        | Detected        | Acceptable  | 20.17  | 0.76 | 20.26   |

## E45-WO-10

\* Peer Data \*

| Target                     | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|----------------------------|-----------------|-----------------|-------------|--------|------|---------|
| Staphylococcus aureus      | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Acinetobacter baumannii    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterococcus faecalis      | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Escherichia coli           | Detected        | Detected        | Acceptable  | 19.97  | 1.27 | 18.62   |
| Klebsiella pneumoniae      | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Pseudomonas aeruginosa     | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Klebsiella oxytoca         | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterobacter cloacae       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Proteus mirabilis          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Morganella morganii        | Detected        | Detected        | Acceptable  | 21.71  | 4.39 | 15.33   |
| Proteus vulgaris           | Detected        | Detected        | Acceptable  | 23.66  | 1.19 | 23.57   |
| Citrobacter freundii       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Enterococcus faecium       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Bacteroides fragilis       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Staphylococcus epidermidis | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Clostridium novyi          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Clostridium perfringens    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Clostridium septicum       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Kingella kingae            | Detected        | Detected        | Acceptable  | 26.68  | 1.35 | 24.89   |
| Group B Strep              | Not Tested      | Detected        | Not Graded  | 23.12  | 0.59 | -       |
| Group A Strep              | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| Group C&G; Strep           | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| ampC                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| mecA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| femA                       | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |

## E45-WO-10

\* Peer Data \*

| Target                 | Reported Result | Expected Result | Performance | Avg CT | SD   | Your CT |
|------------------------|-----------------|-----------------|-------------|--------|------|---------|
| aph2                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| QnrB                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| QnrA                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| vanA1, vanA2           | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| vanB                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| NDM                    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| KPC                    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| OXA 48                 | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| VIM, IMP-7             | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| SHV                    | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| TEM                    | Detected        | Detected        | Acceptable  | 16.6   | 2.98 | 12.24   |
| CTX-M group 1          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| CTX-M group 2          | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| mefA                   | Detected        | Detected        | Acceptable  | 14.3   | 1.37 | 14.46   |
| ErmA                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ErmB                   | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |
| ant2                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| DfrA                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| DfrA1                  | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| DfrA5                  | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetO                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Sul1                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Sul2                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| GyrA                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| ParC                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetM                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| TetB                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Antla                  | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| aph3                   | Not Tested      | Not Detected    | Not Graded  | -      | -    | -       |
| Enterobacter aerogenes | Not Detected    | Not Detected    | Acceptable  | -      | -    | -       |