

Marquis Lab WO Event Results

E45-WO-1

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Staphylococcus aureus	Not Detected	Not Detected	Acceptable	-	-	-
Acinetobacter baumannii	Not Detected	Detected	Not Acceptable	16.77	0.91	-
Enterococcus faecalis	Not Detected	Not Detected	Acceptable	-	-	-
Escherichia coli	Not Detected	Not Detected	Acceptable	-	-	-
Klebsiella pneumoniae	Not Detected	Not Detected	Acceptable	-	-	-
Pseudomonas aeruginosa	Not Detected	Detected	Not Acceptable	19.35	0.6	-
Klebsiella oxytoca	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter cloacae	Not Detected	Detected	Not Acceptable	17.31	1.71	-
Proteus mirabilis	Not Detected	Not Detected	Acceptable	-	-	-
Morganella morganii	Not Detected	Detected	Not Acceptable	21.19	2.36	-
Proteus vulgaris	Not Detected	Not Detected	Acceptable	-	-	-
Citrobacter freundii	Not Detected	Not Detected	Acceptable	-	-	-
Enterococcus faecium	Not Detected	Not Detected	Acceptable	-	-	-
Bacteroides fragilis	Not Detected	Detected	Not Acceptable	20.09	1.24	-
Staphylococcus epidermidis	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium novyi	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium perfringens	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium septicum	Not Detected	Not Detected	Acceptable	-	-	-
Kingella kingae	Not Detected	Not Detected	Acceptable	-	-	-
Group B Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group A Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group C&G; Strep	Not Detected	Not Detected	Acceptable	-	-	-
ampC	Not Detected	Detected	Not Acceptable	18.3	0.94	-
mecA	Not Detected	Not Detected	Acceptable	-	-	-
femA	Not Detected	Not Detected	Acceptable	-	-	-
aph2	Not Detected	Not Detected	Acceptable	-	-	-
QnrB	Not Detected	Not Detected	Acceptable	-	-	-
QnrA	Not Detected	Not Detected	Acceptable	-	-	-
vanA1, vanA2	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-1

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
vanB	Not Detected	Not Detected	Acceptable	-	-	-
NDM	Not Detected	Detected	Not Acceptable	24.4	1.08	-
KPC	Not Detected	Detected	Not Acceptable	18.29	1.76	-
OXA 48	Not Detected	Not Detected	Acceptable	-	-	-
VIM, IMP-7	Not Detected	Detected	Not Acceptable	20.01	2.18	-
SHV	Not Detected	Not Detected	Acceptable	-	-	-
TEM	Not Detected	Detected	Not Acceptable	18.28	1.16	-
CTX-M group 1	Not Detected	Detected	Not Acceptable	29.89	1.24	-
CTX-M group 2	Not Detected	Not Detected	Acceptable	-	-	-
mefA	Not Detected	Not Detected	Acceptable	-	-	-
ErmA	Not Detected	Not Detected	Acceptable	-	-	-
ErmB	Not Detected	Not Detected	Acceptable	-	-	-
ant2	Not Detected	Detected	Not Acceptable	19.43	2.65	-
DfrA	Not Detected	Not Detected	Acceptable	-	-	-
DfrA1	Not Detected	Not Detected	Acceptable	-	-	-
DfrA5	Not Detected	Not Detected	Acceptable	-	-	-
TetO	Not Detected	Not Detected	Acceptable	-	-	-
Sul1	Not Detected	Detected	Not Acceptable	-	-	-
Sul2	Not Detected	Detected	Not Acceptable	-	-	-
GyrA	Not Detected	Not Detected	Acceptable	-	-	-
ParC	Not Detected	Detected	Not Acceptable	-	-	-
TetM	Not Detected	Not Detected	Acceptable	-	-	-
TetB	Not Detected	Detected	Not Acceptable	19.61	1.97	-
Antla	Not Detected	Not Detected	Acceptable	-	-	-
aph3	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter aerogenes	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-2

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Staphylococcus aureus	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-2

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Acinetobacter baumannii	Not Detected	Detected	Not Acceptable	29.1	1.84	-
Enterococcus faecalis	Not Detected	Not Detected	Acceptable	-	-	-
Escherichia coli	Not Detected	Detected	Not Acceptable	17.03	0.71	-
Klebsiella pneumoniae	Not Detected	Not Detected	Acceptable	-	-	-
Pseudomonas aeruginosa	Not Detected	Not Detected	Acceptable	-	-	-
Klebsiella oxytoca	Not Detected	Detected	Not Acceptable	19.66	1.17	-
Enterobacter cloacae	Not Detected	Detected	Not Acceptable	17.44	1.85	-
Proteus mirabilis	Not Detected	Not Detected	Acceptable	-	-	-
Morganella morganii	Not Detected	Not Detected	Acceptable	-	-	-
Proteus vulgaris	Not Detected	Not Detected	Acceptable	-	-	-
Citrobacter freundii	Not Detected	Not Detected	Acceptable	-	-	-
Enterococcus faecium	Not Detected	Not Detected	Acceptable	-	-	-
Bacteroides fragilis	Not Detected	Not Detected	Acceptable	-	-	-
Staphylococcus epidermidis	Not Detected	Detected	Not Acceptable	-	-	-
Clostridium novyi	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium perfringens	Not Detected	Detected	Not Acceptable	21.46	0.93	-
Clostridium septicum	Not Detected	Not Detected	Acceptable	-	-	-
Kingella kingae	Not Detected	Not Detected	Acceptable	-	-	-
Group B Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group A Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group C&G; Strep	Not Detected	Not Detected	Acceptable	-	-	-
ampC	Not Detected	Detected	Not Acceptable	18.11	0.89	-
mecA	Not Detected	Not Detected	Acceptable	-	-	-
femA	Not Detected	Not Detected	Acceptable	-	-	-
aph2	Not Detected	Not Detected	Acceptable	-	-	-
QnrB	Not Detected	Detected	Not Acceptable	18.91	0.77	-
QnrA	Not Detected	Detected	Not Acceptable	20.17	2.19	-
vanA1, vanA2	Not Detected	Not Detected	Acceptable	-	-	-
vanB	Not Detected	Not Detected	Acceptable	-	-	-
NDM	Not Detected	Not Detected	Acceptable	-	-	-
KPC	Not Detected	Detected	Not Acceptable	16.66	1.31	-

E45-WO-2

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
OXA 48	Not Detected	Not Detected	Acceptable	-	-	-
VIM, IMP-7	Not Detected	Detected	Not Acceptable	18.61	1.11	-
SHV	Not Detected	Detected	Not Acceptable	17.07	1.68	-
TEM	Not Detected	Detected	Not Acceptable	16.33	1.21	-
CTX-M group 1	Not Detected	Not Detected	Acceptable	-	-	-
CTX-M group 2	Not Detected	Not Detected	Acceptable	-	-	-
mefA	Not Detected	Not Detected	Acceptable	-	-	-
ErmA	Not Detected	Not Detected	Acceptable	-	-	-
ErmB	Not Detected	Not Detected	Acceptable	-	-	-
ant2	Not Detected	Detected	Not Acceptable	17.41	2.67	-
DfrA	Not Detected	Detected	Not Acceptable	-	-	-
DfrA1	Not Detected	Detected	Not Acceptable	-	-	-
DfrA5	Not Detected	Detected	Not Acceptable	-	-	-
TetO	Not Detected	Not Detected	Acceptable	-	-	-
Sul1	Not Detected	Detected	Not Acceptable	-	-	-
Sul2	Not Detected	Detected	Not Acceptable	-	-	-
GyrA	Not Detected	Detected	Not Acceptable	-	-	-
ParC	Not Detected	Not Detected	Acceptable	-	-	-
TetM	Not Detected	Not Detected	Acceptable	-	-	-
TetB	Not Detected	Not Detected	Acceptable	-	-	-
Antla	Not Detected	Not Detected	Acceptable	-	-	-
aph3	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter aerogenes	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-3

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Staphylococcus aureus	Not Detected	Not Detected	Acceptable	-	-	-
Acinetobacter baumannii	Not Detected	Not Detected	Acceptable	-	-	-
Enterococcus faecalis	Not Detected	Not Detected	Acceptable	-	-	-
Escherichia coli	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-3

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Klebsiella pneumoniae	Not Detected	Detected	Not Acceptable	19.04	1.25	-
Pseudomonas aeruginosa	Not Detected	Detected	Not Acceptable	20.01	0.52	-
Klebsiella oxytoca	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter cloacae	Not Detected	Not Detected	Acceptable	-	-	-
Proteus mirabilis	Not Detected	Not Detected	Acceptable	-	-	-
Morganella morganii	Not Detected	Not Detected	Acceptable	-	-	-
Proteus vulgaris	Not Detected	Not Detected	Acceptable	-	-	-
Citrobacter freundii	Not Detected	Detected	Not Acceptable	21.35	0.74	-
Enterococcus faecium	Not Detected	Not Detected	Acceptable	-	-	-
Bacteroides fragilis	Not Detected	Not Detected	Acceptable	-	-	-
Staphylococcus epidermidis	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium novyi	Not Detected	Detected	Not Acceptable	23.44	2.39	-
Clostridium perfringens	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium septicum	Not Detected	Not Detected	Acceptable	-	-	-
Kingella kingae	Not Detected	Not Detected	Acceptable	-	-	-
Group B Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group A Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group C&G; Strep	Not Detected	Not Detected	Acceptable	-	-	-
ampC	Not Detected	Not Detected	Acceptable	-	-	-
mecA	Not Detected	Not Detected	Acceptable	-	-	-
femA	Not Detected	Not Detected	Acceptable	-	-	-
aph2	Not Detected	Not Detected	Acceptable	-	-	-
QnrB	Not Detected	Not Detected	Acceptable	-	-	-
QnrA	Not Detected	Not Detected	Acceptable	-	-	-
vanA1, vanA2	Not Detected	Not Detected	Acceptable	-	-	-
vanB	Not Detected	Not Detected	Acceptable	-	-	-
NDM	Not Detected	Detected	Not Acceptable	24.16	1.13	-
KPC	Not Detected	Detected	Not Acceptable	17.91	1.94	-
OXA 48	Not Detected	Detected	Not Acceptable	16.8	0.84	-
VIM, IMP-7	Not Detected	Detected	Not Acceptable	23.22	4.37	-
SHV	Not Detected	Detected	Not Acceptable	17.31	1.48	-

E45-WO-3

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
TEM	Not Detected	Detected	Not Acceptable	17.89	1.14	-
CTX-M group 1	Not Detected	Detected	Not Acceptable	22.04	1.29	-
CTX-M group 2	Not Detected	Detected	Not Acceptable	21.1	0.95	-
mefA	Not Detected	Not Detected	Acceptable	-	-	-
ErmA	Not Detected	Not Detected	Acceptable	-	-	-
ErmB	Not Detected	Not Detected	Acceptable	-	-	-
ant2	Not Detected	Detected	Not Acceptable	24.33	0.32	-
DfrA	Not Detected	Not Detected	Acceptable	-	-	-
DfrA1	Not Detected	Not Detected	Acceptable	-	-	-
DfrA5	Not Detected	Not Detected	Acceptable	-	-	-
TetO	Not Detected	Not Detected	Acceptable	-	-	-
Sul1	Not Detected	Detected	Not Acceptable	-	-	-
Sul2	Not Detected	Detected	Not Acceptable	-	-	-
GyrA	Not Detected	Not Detected	Acceptable	-	-	-
ParC	Not Detected	Not Detected	Acceptable	-	-	-
TetM	Not Detected	Not Detected	Acceptable	-	-	-
TetB	Not Detected	Not Detected	Acceptable	-	-	-
Antla	Not Detected	Not Detected	Acceptable	-	-	-
aph3	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter aerogenes	Not Detected	Detected	Not Acceptable	19.32	1.95	-

E45-WO-4

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Staphylococcus aureus	Not Detected	Not Detected	Acceptable	-	-	-
Acinetobacter baumannii	Not Detected	Not Detected	Acceptable	-	-	-
Enterococcus faecalis	Not Detected	Not Detected	Acceptable	-	-	-
Escherichia coli	Not Detected	Detected	Not Acceptable	18.14	1.76	-
Klebsiella pneumoniae	Not Detected	Not Detected	Acceptable	-	-	-
Pseudomonas aeruginosa	Not Detected	Not Detected	Acceptable	-	-	-
Klebsiella oxytoca	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-4

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Enterobacter cloacae	Not Detected	Not Detected	Acceptable	-	-	-
Proteus mirabilis	Not Detected	Not Detected	Acceptable	-	-	-
Morganella morganii	Not Detected	Detected	Not Acceptable	22.09	2.13	-
Proteus vulgaris	Not Detected	Detected	Not Acceptable	23.36	1.37	-
Citrobacter freundii	Not Detected	Not Detected	Acceptable	-	-	-
Enterococcus faecium	Not Detected	Not Detected	Acceptable	-	-	-
Bacteroides fragilis	Not Detected	Not Detected	Acceptable	-	-	-
Staphylococcus epidermidis	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium novyi	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium perfringens	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium septicum	Not Detected	Not Detected	Acceptable	-	-	-
Kingella kingae	Not Detected	Detected	Not Acceptable	26.09	0.72	-
Group B Strep	Not Detected	Detected	Not Acceptable	20.13	1.09	-
Group A Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group C&G; Strep	Not Detected	Not Detected	Acceptable	-	-	-
ampC	Not Detected	Not Detected	Acceptable	-	-	-
mecA	Not Detected	Not Detected	Acceptable	-	-	-
femA	Not Detected	Not Detected	Acceptable	-	-	-
aph2	Not Detected	Not Detected	Acceptable	-	-	-
QnrB	Not Detected	Not Detected	Acceptable	-	-	-
QnrA	Not Detected	Not Detected	Acceptable	-	-	-
vanA1, vanA2	Not Detected	Not Detected	Acceptable	-	-	-
vanB	Not Detected	Not Detected	Acceptable	-	-	-
NDM	Not Detected	Not Detected	Acceptable	-	-	-
KPC	Not Detected	Not Detected	Acceptable	-	-	-
OXA 48	Not Detected	Not Detected	Acceptable	-	-	-
VIM, IMP-7	Not Detected	Not Detected	Acceptable	-	-	-
SHV	Not Detected	Not Detected	Acceptable	-	-	-
TEM	Not Detected	Detected	Not Acceptable	15.32	1.62	-
CTX-M group 1	Not Detected	Not Detected	Acceptable	-	-	-
CTX-M group 2	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-4

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
mefA	Not Detected	Detected	Not Acceptable	13.78	1.47	-
ErmA	Not Detected	Not Detected	Acceptable	-	-	-
ErmB	Not Detected	Not Detected	Acceptable	-	-	-
ant2	Not Detected	Not Detected	Acceptable	-	-	-
DfrA	Not Detected	Not Detected	Acceptable	-	-	-
DfrA1	Not Detected	Not Detected	Acceptable	-	-	-
DfrA5	Not Detected	Not Detected	Acceptable	-	-	-
TetO	Not Detected	Not Detected	Acceptable	-	-	-
Sul1	Not Detected	Not Detected	Acceptable	-	-	-
Sul2	Not Detected	Not Detected	Acceptable	-	-	-
GyrA	Not Detected	Not Detected	Acceptable	-	-	-
ParC	Not Detected	Not Detected	Acceptable	-	-	-
TetM	Not Detected	Not Detected	Acceptable	-	-	-
TetB	Not Detected	Not Detected	Acceptable	-	-	-
Antla	Not Detected	Not Detected	Acceptable	-	-	-
aph3	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter aerogenes	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-5

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Staphylococcus aureus	Not Detected	Not Detected	Acceptable	-	-	-
Acinetobacter baumannii	Not Detected	Not Detected	Acceptable	-	-	-
Enterococcus faecalis	Not Detected	Not Detected	Acceptable	-	-	-
Escherichia coli	Not Detected	Not Detected	Acceptable	-	-	-
Klebsiella pneumoniae	Not Detected	Not Detected	Acceptable	-	-	-
Pseudomonas aeruginosa	Not Detected	Not Detected	Acceptable	-	-	-
Klebsiella oxytoca	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter cloacae	Not Detected	Not Detected	Acceptable	-	-	-
Proteus mirabilis	Not Detected	Detected	Not Acceptable	17.25	1.33	-
Morganella morganii	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-5

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Proteus vulgaris	Not Detected	Not Detected	Acceptable	-	-	-
Citrobacter freundii	Not Detected	Not Detected	Acceptable	-	-	-
Enterococcus faecium	Not Detected	Detected	Not Acceptable	17.43	1.84	-
Bacteroides fragilis	Not Detected	Not Detected	Acceptable	-	-	-
Staphylococcus epidermidis	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium novyi	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium perfringens	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium septicum	Not Detected	Not Detected	Acceptable	-	-	-
Kingella kingae	Not Detected	Not Detected	Acceptable	-	-	-
Group B Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group A Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group C&G; Strep	Not Detected	Detected	Not Acceptable	21.31	1.25	-
ampC	Not Detected	Not Detected	Acceptable	-	-	-
mecA	Not Detected	Not Detected	Acceptable	-	-	-
femA	Not Detected	Not Detected	Acceptable	-	-	-
aph2	Not Detected	Detected	Not Acceptable	-	-	-
QnrB	Not Detected	Not Detected	Acceptable	-	-	-
QnrA	Not Detected	Not Detected	Acceptable	-	-	-
vanA1, vanA2	Not Detected	Detected	Not Acceptable	18.51	1.94	-
vanB	Not Detected	Not Detected	Acceptable	-	-	-
NDM	Not Detected	Not Detected	Acceptable	-	-	-
KPC	Not Detected	Not Detected	Acceptable	-	-	-
OXA 48	Not Detected	Not Detected	Acceptable	-	-	-
VIM, IMP-7	Not Detected	Not Detected	Acceptable	-	-	-
SHV	Not Detected	Not Detected	Acceptable	-	-	-
TEM	Not Detected	Not Detected	Acceptable	-	-	-
CTX-M group 1	Not Detected	Not Detected	Acceptable	-	-	-
CTX-M group 2	Not Detected	Not Detected	Acceptable	-	-	-
mefA	Not Detected	Not Detected	Acceptable	-	-	-
ErmA	Not Detected	Not Detected	Acceptable	-	-	-
ErmB	Not Detected	Detected	Not Acceptable	17.81	1.89	-

E45-WO-5

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
ant2	Not Detected	Not Detected	Acceptable	-	-	-
DfrA	Not Detected	Not Detected	Acceptable	-	-	-
DfrA1	Not Detected	Not Detected	Acceptable	-	-	-
DfrA5	Not Detected	Not Detected	Acceptable	-	-	-
TetO	Not Detected	Not Detected	Acceptable	-	-	-
Sul1	Not Detected	Not Detected	Acceptable	-	-	-
Sul2	Not Detected	Not Detected	Acceptable	-	-	-
GyrA	Not Detected	Not Detected	Acceptable	-	-	-
ParC	Not Detected	Not Detected	Acceptable	-	-	-
TetM	Not Detected	Detected	Not Acceptable	-	-	-
TetB	Not Detected	Not Detected	Acceptable	-	-	-
Antla	Not Detected	Not Detected	Acceptable	-	-	-
aph3	Not Detected	Detected	Not Acceptable	-	-	-
Enterobacter aerogenes	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-6

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Staphylococcus aureus	Not Detected	Detected	Not Acceptable	18.92	1.48	-
Acinetobacter baumannii	Not Detected	Not Detected	Acceptable	-	-	-
Enterococcus faecalis	Not Detected	Detected	Not Acceptable	18.1	2.2	-
Escherichia coli	Not Detected	Not Detected	Acceptable	-	-	-
Klebsiella pneumoniae	Not Detected	Not Detected	Acceptable	-	-	-
Pseudomonas aeruginosa	Not Detected	Not Detected	Acceptable	-	-	-
Klebsiella oxytoca	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter cloacae	Not Detected	Not Detected	Acceptable	-	-	-
Proteus mirabilis	Not Detected	Not Detected	Acceptable	-	-	-
Morganella morganii	Not Detected	Not Detected	Acceptable	-	-	-
Proteus vulgaris	Not Detected	Not Detected	Acceptable	-	-	-
Citrobacter freundii	Not Detected	Not Detected	Acceptable	-	-	-
Enterococcus faecium	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-6

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Bacteroides fragilis	Not Detected	Not Detected	Acceptable	-	-	-
Staphylococcus epidermidis	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium novyi	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium perfringens	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium septicum	Not Detected	Detected	Not Acceptable	24.32	1.01	-
Kingella kingae	Not Detected	Not Detected	Acceptable	-	-	-
Group B Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group A Strep	Not Detected	Detected	Not Acceptable	19.37	0.69	-
Group C&G; Strep	Not Detected	Not Detected	Acceptable	-	-	-
ampC	Not Detected	Not Detected	Acceptable	-	-	-
mecA	Not Detected	Detected	Not Acceptable	20.34	2.09	-
femA	Not Detected	Detected	Not Acceptable	19.39	1.32	-
aph2	Not Detected	Not Detected	Acceptable	-	-	-
QnrB	Not Detected	Not Detected	Acceptable	-	-	-
QnrA	Not Detected	Not Detected	Acceptable	-	-	-
vanA1, vanA2	Not Detected	Not Detected	Acceptable	-	-	-
vanB	Not Detected	Detected	Not Acceptable	18.16	2.2	-
NDM	Not Detected	Not Detected	Acceptable	-	-	-
KPC	Not Detected	Not Detected	Acceptable	-	-	-
OXA 48	Not Detected	Not Detected	Acceptable	-	-	-
VIM, IMP-7	Not Detected	Not Detected	Acceptable	-	-	-
SHV	Not Detected	Not Detected	Acceptable	-	-	-
TEM	Not Detected	Not Detected	Acceptable	-	-	-
CTX-M group 1	Not Detected	Not Detected	Acceptable	-	-	-
CTX-M group 2	Not Detected	Not Detected	Acceptable	-	-	-
mefA	Not Detected	Not Detected	Acceptable	-	-	-
ErmA	Not Detected	Detected	Not Acceptable	18.37	1.79	-
ErmB	Not Detected	Detected	Not Acceptable	17.81	2.72	-
ant2	Not Detected	Not Detected	Acceptable	-	-	-
DfrA	Not Detected	Detected	Not Acceptable	-	-	-
DfrA1	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-6

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
DfrA5	Not Detected	Not Detected	Acceptable	-	-	-
TetO	Not Detected	Not Detected	Acceptable	-	-	-
Sul1	Not Detected	Not Detected	Acceptable	-	-	-
Sul2	Not Detected	Not Detected	Acceptable	-	-	-
GyrA	Not Detected	Not Detected	Acceptable	-	-	-
ParC	Not Detected	Not Detected	Acceptable	-	-	-
TetM	Not Detected	Detected	Not Acceptable	-	-	-
TetB	Not Detected	Not Detected	Acceptable	-	-	-
Antla	Not Detected	Detected	Not Acceptable	-	-	-
aph3	Not Detected	Detected	Not Acceptable	-	-	-
Enterobacter aerogenes	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-7

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Staphylococcus aureus	Not Detected	Not Detected	Acceptable	-	-	-
Acinetobacter baumannii	Not Detected	Detected	Not Acceptable	18.54	1.39	-
Enterococcus faecalis	Not Detected	Not Detected	Acceptable	-	-	-
Escherichia coli	Not Detected	Not Detected	Acceptable	-	-	-
Klebsiella pneumoniae	Not Detected	Not Detected	Acceptable	-	-	-
Pseudomonas aeruginosa	Not Detected	Detected	Not Acceptable	19.81	0.62	-
Klebsiella oxytoca	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter cloacae	Not Detected	Detected	Not Acceptable	18.97	1.66	-
Proteus mirabilis	Not Detected	Not Detected	Acceptable	-	-	-
Morganella morganii	Not Detected	Detected	Not Acceptable	21.62	3.54	-
Proteus vulgaris	Not Detected	Not Detected	Acceptable	-	-	-
Citrobacter freundii	Not Detected	Not Detected	Acceptable	-	-	-
Enterococcus faecium	Not Detected	Not Detected	Acceptable	-	-	-
Bacteroides fragilis	Not Detected	Detected	Not Acceptable	21.13	1.65	-
Staphylococcus epidermidis	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium novyi	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-7

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Clostridium perfringens	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium septicum	Not Detected	Not Detected	Acceptable	-	-	-
Kingella kingae	Not Detected	Not Detected	Acceptable	-	-	-
Group B Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group A Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group C&G; Strep	Not Detected	Not Detected	Acceptable	-	-	-
ampC	Not Detected	Detected	Not Acceptable	19.9	1.39	-
mecA	Not Detected	Not Detected	Acceptable	-	-	-
femA	Not Detected	Not Detected	Acceptable	-	-	-
aph2	Not Detected	Not Detected	Acceptable	-	-	-
QnrB	Not Detected	Not Detected	Acceptable	-	-	-
QnrA	Not Detected	Not Detected	Acceptable	-	-	-
vanA1, vanA2	Not Detected	Not Detected	Acceptable	-	-	-
vanB	Not Detected	Not Detected	Acceptable	-	-	-
NDM	Not Detected	Detected	Not Acceptable	25.62	1.27	-
KPC	Not Detected	Detected	Not Acceptable	20.56	1.76	-
OXA 48	Not Detected	Not Detected	Acceptable	-	-	-
VIM, IMP-7	Not Detected	Detected	Not Acceptable	20.59	2.67	-
SHV	Not Detected	Not Detected	Acceptable	-	-	-
TEM	Not Detected	Detected	Not Acceptable	19.74	1.94	-
CTX-M group 1	Not Detected	Detected	Not Acceptable	30.99	0.39	-
CTX-M group 2	Not Detected	Not Detected	Acceptable	-	-	-
mefA	Not Detected	Not Detected	Acceptable	-	-	-
ErmA	Not Detected	Not Detected	Acceptable	-	-	-
ErmB	Not Detected	Not Detected	Acceptable	-	-	-
ant2	Not Detected	Detected	Not Acceptable	21.14	1.63	-
DfrA	Not Detected	Not Detected	Acceptable	-	-	-
DfrA1	Not Detected	Not Detected	Acceptable	-	-	-
DfrA5	Not Detected	Not Detected	Acceptable	-	-	-
TetO	Not Detected	Not Detected	Acceptable	-	-	-
Sul1	Not Detected	Detected	Not Acceptable	-	-	-

E45-WO-7

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Sul2	Not Detected	Detected	Not Acceptable	-	-	-
GyrA	Not Detected	Not Detected	Acceptable	-	-	-
ParC	Not Detected	Detected	Not Acceptable	-	-	-
TetM	Not Detected	Not Detected	Acceptable	-	-	-
TetB	Not Detected	Detected	Not Acceptable	20.98	1.17	-
Antla	Not Detected	Not Detected	Acceptable	-	-	-
aph3	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter aerogenes	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-8

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Staphylococcus aureus	Not Detected	Not Detected	Acceptable	-	-	-
Acinetobacter baumannii	Not Detected	Detected	Not Acceptable	30.26	0.67	-
Enterococcus faecalis	Not Detected	Not Detected	Acceptable	-	-	-
Escherichia coli	Not Detected	Detected	Not Acceptable	19.25	0.63	-
Klebsiella pneumoniae	Not Detected	Not Detected	Acceptable	-	-	-
Pseudomonas aeruginosa	Not Detected	Not Detected	Acceptable	-	-	-
Klebsiella oxytoca	Not Detected	Detected	Not Acceptable	22.29	0.83	-
Enterobacter cloacae	Not Detected	Detected	Not Acceptable	19.4	2.06	-
Proteus mirabilis	Not Detected	Not Detected	Acceptable	-	-	-
Morganella morganii	Not Detected	Not Detected	Acceptable	-	-	-
Proteus vulgaris	Not Detected	Not Detected	Acceptable	-	-	-
Citrobacter freundii	Not Detected	Not Detected	Acceptable	-	-	-
Enterococcus faecium	Not Detected	Not Detected	Acceptable	-	-	-
Bacteroides fragilis	Not Detected	Not Detected	Acceptable	-	-	-
Staphylococcus epidermidis	Not Detected	Detected	Not Acceptable	-	-	-
Clostridium novyi	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium perfringens	Not Detected	Detected	Not Acceptable	23.7	1.35	-
Clostridium septicum	Not Detected	Not Detected	Acceptable	-	-	-
Kingella kingae	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-8

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Group B Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group A Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group C&G; Strep	Not Detected	Not Detected	Acceptable	-	-	-
ampC	Not Detected	Detected	Not Acceptable	19.95	1.44	-
mecA	Not Detected	Not Detected	Acceptable	-	-	-
femA	Not Detected	Not Detected	Acceptable	-	-	-
aph2	Not Detected	Not Detected	Acceptable	-	-	-
QnrB	Not Detected	Detected	Not Acceptable	21.31	0.97	-
QnrA	Not Detected	Detected	Not Acceptable	22.44	0.94	-
vanA1, vanA2	Not Detected	Not Detected	Acceptable	-	-	-
vanB	Not Detected	Not Detected	Acceptable	-	-	-
NDM	Not Detected	Not Detected	Acceptable	-	-	-
KPC	Not Detected	Detected	Not Acceptable	19.57	0.81	-
OXA 48	Not Detected	Not Detected	Acceptable	-	-	-
VIM, IMP-7	Not Detected	Detected	Not Acceptable	20.82	2.17	-
SHV	Not Detected	Detected	Not Acceptable	19.62	1.0	-
TEM	Not Detected	Detected	Not Acceptable	18.69	1.16	-
CTX-M group 1	Not Detected	Not Detected	Acceptable	-	-	-
CTX-M group 2	Not Detected	Not Detected	Acceptable	-	-	-
mefA	Not Detected	Not Detected	Acceptable	-	-	-
ErmA	Not Detected	Not Detected	Acceptable	-	-	-
ErmB	Not Detected	Not Detected	Acceptable	-	-	-
ant2	Not Detected	Detected	Not Acceptable	20.27	1.73	-
DfrA	Not Detected	Detected	Not Acceptable	-	-	-
DfrA1	Not Detected	Detected	Not Acceptable	-	-	-
DfrA5	Not Detected	Detected	Not Acceptable	-	-	-
TetO	Not Detected	Not Detected	Acceptable	-	-	-
Sul1	Not Detected	Detected	Not Acceptable	-	-	-
Sul2	Not Detected	Detected	Not Acceptable	-	-	-
GyrA	Not Detected	Detected	Not Acceptable	-	-	-
ParC	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-8

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
TetM	Not Detected	Not Detected	Acceptable	-	-	-
TetB	Not Detected	Not Detected	Acceptable	-	-	-
Antla	Not Detected	Not Detected	Acceptable	-	-	-
aph3	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter aerogenes	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-9

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Staphylococcus aureus	Not Detected	Not Detected	Acceptable	-	-	-
Acinetobacter baumannii	Not Detected	Not Detected	Acceptable	-	-	-
Enterococcus faecalis	Not Detected	Not Detected	Acceptable	-	-	-
Escherichia coli	Not Detected	Not Detected	Acceptable	-	-	-
Klebsiella pneumoniae	Not Detected	Detected	Not Acceptable	19.88	0.59	-
Pseudomonas aeruginosa	Not Detected	Detected	Not Acceptable	18.85	0.81	-
Klebsiella oxytoca	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter cloacae	Not Detected	Not Detected	Acceptable	-	-	-
Proteus mirabilis	Not Detected	Not Detected	Acceptable	-	-	-
Morganella morganii	Not Detected	Not Detected	Acceptable	-	-	-
Proteus vulgaris	Not Detected	Not Detected	Acceptable	-	-	-
Citrobacter freundii	Not Detected	Detected	Not Acceptable	22.56	0.64	-
Enterococcus faecium	Not Detected	Not Detected	Acceptable	-	-	-
Bacteroides fragilis	Not Detected	Not Detected	Acceptable	-	-	-
Staphylococcus epidermidis	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium novyi	Not Detected	Detected	Not Acceptable	22.64	1.06	-
Clostridium perfringens	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium septicum	Not Detected	Not Detected	Acceptable	-	-	-
Kingella kingae	Not Detected	Not Detected	Acceptable	-	-	-
Group B Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group A Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group C&G; Strep	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-9

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
ampC	Not Detected	Not Detected	Acceptable	-	-	-
mecA	Not Detected	Not Detected	Acceptable	-	-	-
femA	Not Detected	Not Detected	Acceptable	-	-	-
aph2	Not Detected	Not Detected	Acceptable	-	-	-
QnrB	Not Detected	Not Detected	Acceptable	-	-	-
QnrA	Not Detected	Not Detected	Acceptable	-	-	-
vanA1, vanA2	Not Detected	Not Detected	Acceptable	-	-	-
vanB	Not Detected	Not Detected	Acceptable	-	-	-
NDM	Not Detected	Detected	Not Acceptable	23.82	1.77	-
KPC	Not Detected	Detected	Not Acceptable	19.92	1.72	-
OXA 48	Not Detected	Detected	Not Acceptable	18.25	1.05	-
VIM, IMP-7	Not Detected	Detected	Not Acceptable	25.52	1.64	-
SHV	Not Detected	Detected	Not Acceptable	19.11	1.67	-
TEM	Not Detected	Detected	Not Acceptable	19.54	1.86	-
CTX-M group 1	Not Detected	Detected	Not Acceptable	23.98	1.29	-
CTX-M group 2	Not Detected	Detected	Not Acceptable	22.08	0.79	-
mefA	Not Detected	Not Detected	Acceptable	-	-	-
ErmA	Not Detected	Not Detected	Acceptable	-	-	-
ErmB	Not Detected	Not Detected	Acceptable	-	-	-
ant2	Not Detected	Detected	Not Acceptable	22.76	0.06	-
DfrA	Not Detected	Not Detected	Acceptable	-	-	-
DfrA1	Not Detected	Not Detected	Acceptable	-	-	-
DfrA5	Not Detected	Not Detected	Acceptable	-	-	-
TetO	Not Detected	Not Detected	Acceptable	-	-	-
Sul1	Not Detected	Detected	Not Acceptable	-	-	-
Sul2	Not Detected	Detected	Not Acceptable	-	-	-
GyrA	Not Detected	Not Detected	Acceptable	-	-	-
ParC	Not Detected	Not Detected	Acceptable	-	-	-
TetM	Not Detected	Not Detected	Acceptable	-	-	-
TetB	Not Detected	Not Detected	Acceptable	-	-	-
Antla	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-9

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
aph3	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter aerogenes	Not Detected	Detected	Not Acceptable	20.17	0.76	-

E45-WO-10

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
Staphylococcus aureus	Not Detected	Not Detected	Acceptable	-	-	-
Acinetobacter baumannii	Not Detected	Not Detected	Acceptable	-	-	-
Enterococcus faecalis	Not Detected	Not Detected	Acceptable	-	-	-
Escherichia coli	Not Detected	Detected	Not Acceptable	19.97	1.27	-
Klebsiella pneumoniae	Not Detected	Not Detected	Acceptable	-	-	-
Pseudomonas aeruginosa	Not Detected	Not Detected	Acceptable	-	-	-
Klebsiella oxytoca	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter cloacae	Not Detected	Not Detected	Acceptable	-	-	-
Proteus mirabilis	Not Detected	Not Detected	Acceptable	-	-	-
Morganella morganii	Not Detected	Detected	Not Acceptable	21.71	4.39	-
Proteus vulgaris	Not Detected	Detected	Not Acceptable	23.66	1.19	-
Citrobacter freundii	Not Detected	Not Detected	Acceptable	-	-	-
Enterococcus faecium	Not Detected	Not Detected	Acceptable	-	-	-
Bacteroides fragilis	Not Detected	Not Detected	Acceptable	-	-	-
Staphylococcus epidermidis	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium novyi	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium perfringens	Not Detected	Not Detected	Acceptable	-	-	-
Clostridium septicum	Not Detected	Not Detected	Acceptable	-	-	-
Kingella kingae	Not Detected	Detected	Not Acceptable	26.68	1.35	-
Group B Strep	Not Detected	Detected	Not Acceptable	23.12	0.59	-
Group A Strep	Not Detected	Not Detected	Acceptable	-	-	-
Group C&G; Strep	Not Detected	Not Detected	Acceptable	-	-	-
ampC	Not Detected	Not Detected	Acceptable	-	-	-
mecA	Not Detected	Not Detected	Acceptable	-	-	-
femA	Not Detected	Not Detected	Acceptable	-	-	-

E45-WO-10

* Peer Data *

Target	Reported Result	Expected Result	Performance	Avg CT	SD	Your CT
aph2	Not Detected	Not Detected	Acceptable	-	-	-
QnrB	Not Detected	Not Detected	Acceptable	-	-	-
QnrA	Not Detected	Not Detected	Acceptable	-	-	-
vanA1, vanA2	Not Detected	Not Detected	Acceptable	-	-	-
vanB	Not Detected	Not Detected	Acceptable	-	-	-
NDM	Not Detected	Not Detected	Acceptable	-	-	-
KPC	Not Detected	Not Detected	Acceptable	-	-	-
OXA 48	Not Detected	Not Detected	Acceptable	-	-	-
VIM, IMP-7	Not Detected	Not Detected	Acceptable	-	-	-
SHV	Not Detected	Not Detected	Acceptable	-	-	-
TEM	Not Detected	Detected	Not Acceptable	16.6	2.98	-
CTX-M group 1	Not Detected	Not Detected	Acceptable	-	-	-
CTX-M group 2	Not Detected	Not Detected	Acceptable	-	-	-
mefA	Not Detected	Detected	Not Acceptable	14.3	1.37	-
ErmA	Not Detected	Not Detected	Acceptable	-	-	-
ErmB	Not Detected	Not Detected	Acceptable	-	-	-
ant2	Not Detected	Not Detected	Acceptable	-	-	-
DfrA	Not Detected	Not Detected	Acceptable	-	-	-
DfrA1	Not Detected	Not Detected	Acceptable	-	-	-
DfrA5	Not Detected	Not Detected	Acceptable	-	-	-
TetO	Not Detected	Not Detected	Acceptable	-	-	-
Sul1	Not Detected	Not Detected	Acceptable	-	-	-
Sul2	Not Detected	Not Detected	Acceptable	-	-	-
GyrA	Not Detected	Not Detected	Acceptable	-	-	-
ParC	Not Detected	Not Detected	Acceptable	-	-	-
TetM	Not Detected	Not Detected	Acceptable	-	-	-
TetB	Not Detected	Not Detected	Acceptable	-	-	-
Antla	Not Detected	Not Detected	Acceptable	-	-	-
aph3	Not Detected	Not Detected	Acceptable	-	-	-
Enterobacter aerogenes	Not Detected	Not Detected	Acceptable	-	-	-