

CERTIFICATE OF CALIBRATION

Issue Date: 4 August 2026

Certificate No: 2451

Certificate issued by:

ACR Instruments
Unit 2, Brick Kiln Way
Rugeley
WS15 1RD
01543 480993
www.acrsuk.com

Customer Name:

Address:

Order #:

Test Date:

4 August 2026

Ambient Temperature:

22°C +/- 2°C

Ambient Humidity:

65% RH

Temperature Scale:

International Temp Scale of 1990

Instrument Desc:

Fieldpiece SM480VINT

Sensor Desc:

K Type Thermocouple

Serial Number (if applicable):

21036287AABT

Instrument I.D:

6287

Calibration Range:

-18°C to +100°C / 0.1°C resolution

Procedure: The instrument was stabilised at ambient temperature and then calibrated by comparison from traceable probes in liquid baths.

Results:	Test Temperature °C	Instrument Reading °C
	-18.00	-18.00
	0.00	0.00
	40.00	40.01
	70.00	70.01
	100.00	100.03

Procedure: The instrument was stabilised at ambient temperature and then calibrated by comparison with nitrogen from traceable transducers.

Instrument Desc:	Fieldpiece SM480VINT
Sensor Desc:	Relative Pressure Sensor
Serial Number (if applicable):	21036287AABT
Instrument I.D:	6287
Calibration Range:	0 to 12 BAR(g) / 0.01 BAR(g) resolution

Results:	Test Pressure BAR(g)	Instrument Reading BAR(g)
	0.00	0.00
	1.00	1.00
	2.00	2.00
	3.00	3.00
	4.00	4.00
	5.00	5.00
	6.00	6.01
	7.00	7.00
	8.00	8.02
	9.00	9.00
	10.00	10.00
	11.00	11.01
	12.00	12.01

Uncertainty of Measurement +/- 0.01 BarG