

*** DEMO DATA — FOR TESTING AND EVALUATION PURPOSES ONLY ***

MINISTRY OF CONSUMER AFFAIRS, FOOD & PUBLIC DISTRIBUTION
DEPARTMENT OF LEGAL METROLOGY — GOVERNMENT OF INDIA

OIML R-76 TYPE-EVALUATION TEST REPORT FOR NON-AUTOMATIC WEIGHING INSTRUMENT (NAWI)

Report Certificate No: NAWI-R76-2026-DEMO-005 | Date: 2026-08-17

1. ISSUING LABORATORY DETAILS

Laboratory Name:	Delhi Central Legal Metrology Laboratory	Accreditation Ref:	NABL-LM-2026-001
Lab Code:	LAB-DEL-01	Contact Email:	delhi.lab@doca.gov.in
Address:	Block C, CGO Complex, Lodhi Road, New Delhi 110003	Contact Phone:	+91-11-24360001

2. APPLICANT / MANUFACTURER DETAILS

Manufacturer Name:	Avery India Ltd.	Country:	India
Address:	Sector 25, Faridabad, Haryana 121004	Contact:	contact@averyindia.co.in +91-129-2234000

3. INSTRUMENT IDENTIFICATION & TECHNICAL SPECIFICATIONS

Model Name:	Avery Weigh-Tronix E1205	Accuracy Class:	Class III
Serial Number:	SN-AVERY-2026-005	Max Capacity (Max):	30.0 kg
Year of Manufacture:	2026	Min Capacity (Min):	0.2 kg
Verification Interval (e):	0.01 kg	Actual Interval (d):	0.01 kg
Scale Intervals (n = Max/e):	3000	Temp Range:	-10.0°C to 40.0°C

4. TEST ENVIRONMENT & REFERENCE STANDARDS USED

Ambient Temperature: 20.5 °C | Relative Humidity: 55.0 % | Barometric Pressure: 1013.25 hPa

Equipment Type	Identifier	Cert No.	Due Date	Status
Standard Weight	W-E2-50K-01	CAL-2025-E2-001	2027-01-10	VALID

5. COMPLIANCE RULE EVALUATION AGAINST OIML R-76

Test Item	Observed Value	Limit (e)	Margin (e)	Result	Notes
TEST-WARMUP	0.200 e	±0.5 e	+0.300 e	PASS	Warm-up drift of 0.200 e evaluated against limit of ±0.5 e (Accuracy Class III). Margin: +0.300 e. [REQUIRES VERIFICATION AGAINST THE APPLICABLE OIML R-76 EDITION]

TEST-REPEATABILITY	0.500 e	≤ 1.0 e	+0.500 e	PASS	Repeatability maximum spread of readings P_max - P_min is 0.500 e. Evaluated against MPE limit of 1.0 e. Margin: +0.500 e. [REQUIRES VERIFICATION AGAINST THE APPLICABLE OIML R-76 EDITION]
TEST-ECCENTRICITY	0.500 e	±1.0 e	+0.500 e	PASS	Eccentric corner loading maximum error is 0.500 e. Evaluated against limit of ±1.0 e (Class III). Margin: +0.500 e. [REQUIRES VERIFICATION AGAINST THE APPLICABLE OIML R-76 EDITION]

6. OVERALL COMPLIANCE DETERMINATION & SIGN-OFF

OVERALL EVALUATION RESULT:	PASS
Inspector / Tester: Amit Sharma (Inspector 1) Date: 2026-08-17	Reviewer / Approver: Priya Verma (Reviewer 1) Decision Date: 2026-08-17

7. DIGITAL INTEGRITY & REPORT AUTHENTICITY

	SHA-256 Report Content Hash: 146a2d7034531bd09739b385c669f0808a1fd4770437e7a62bef29d1a09fc6f0 Verification Payload URL: http://localhost:5173/verify/NAWI-R76-2026-DEMO-005 Scan QR code or visit public verification URL to confirm legal validity of this report record.
---	--