



Republic of the Philippines  
Department of Health  
Metro Manila Center for Health Development  
**VALENZUELA MEDICAL CENTER**



**BID BULLETIN NO. 1**

**Supply and Delivery of Various Medical Supplies for Cy 2024 – Batch 2**  
**PB No. 2024-025**  
**19 January 2024**

Issued pursuant to Sec. 22.5 of the IRR of RA 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the **Pre-Bid** Conference held on **January 19, 2024**, likewise, respond to bidder's written queries received within the prescriptive period for filing.

ITEM NO.	UNIT	Posted in Bidding Documents:	Amendment in the Minutes of the Pre-Bid Conference:
11	pad	CTG paper, thermal paper (Fetal Monitoring)	CTG paper, thermal paper (Fetal Monitoring), compatible to fetal monitor, HUNTLEIGH CTG MACHINE
16	pc	Disposable Isolation Gown, non-woven security Protection Suit, non sterile, knitted arm sleeve, 40-60gsm, with FDA approved/ Certificate of Registration (CPR)	Disposable Isolation Gown, non-woven security Protection Suit, non sterile, knitted arm sleeve, 40-63gsm, with FDA approved/ Certificate of Registration (CPR)
21	pc	ECG Electrodes, foam, standard size, 50's/pack	ECG Electrodes, foam, standard size, adult, with gel, disposable 50's/pack
22	pc	Empty Gallon 15 liters (for sharps container) color: yellow, with label SHARP CONTAINER, printed in red ink	Empty Gallon 15 liters (for sharps container) color: yellow, with label SHARP CONTAINER, printed in red ink, thickness: 1-1.5mm
38	pc	Ligating Clip, Medium and Large (with free 1applier)	Ligating Clip, Medium and Large (with free 1applier) titanium, medium - large
50	gal	Soaking Solution, ( 2% Glutaraldehyde ) ,with material safety data sheet (MSDS); FDA approved/ with CPR with indicated date of manufacture and expiration	Soaking Solution, ( 2% - 2.5% Glutaraldehyde ) ,with material safety data sheet (MSDS); FDA approved/ with CPR with indicated date of manufacture and expiration
61	gal	Ultrasound Transonic gel, clear water based, scent free (Transonic); with indicated date of manufacture and expiration	Ultrasound gel, clear water based, scent free ; with indicated date of manufacture and expiration

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Noted by:

  
**SHIRLENE V. VIANZON**