



Republic of the Philippines  
Department of Health  
Metro Manila Center for Health Development  
**VALENZUELA MEDICAL CENTER**



BAGONG PILIPINAS

**BIDS AND AWARDS COMMITTEE**

June 3, 2025

**NOTICE OF AWARD**

Supply and Delivery of Various PM & R Medical Equipment – Batch 2  
Public Bidding No. VMC-2025 – 053 dated April 23, 2025

**MS. ELIZA VILLARAMA**

Authorized Representative

**FAIRBRIGHT ENTERPRISES, INC**

LG 29-32 Alfaro Place Bldg. 146 Alfaro St.,

Salcedo Village Bel-Air, Makati City

Contact No.: 09178902418

Email: [fairbright@gmail.com](mailto:fairbright@gmail.com)

**Dear Ms. Villarama:**

We are happy to notify your Bid dated April 23, 2025 for execution of Valenzuela Medical Center's IAEB Ref. No. 11940607 for contract price equivalent to **Two Hundred Eighty-Seven Thousand Two Hundred Twenty Pesos Only (P 287,220.00)** for the Supply and Delivery of Various PM & R Medical Equipment – Batch 2 as corrected and defined in accordance with the Instruction to Bidders is hereby accepted.

ITEM NO	QTY	UNIT	ITEM DESCRIPTION	UNIT AMOUNT	TOTAL AMOUNT
1	1	unit	<b>PARAFFIN WAX BATH TANK</b> <b>BRAND: ENRAF NONIUS</b> <b>ORIGIN: THE NETHERLANDS</b> <b>I. Equipment Specification:</b> - Stainless Steel 316 - With built-In thermostat - Can hold 30 Liters of wax - Melting time of wax not more than 3 hours - With heavy duty wheels and casterlock - Comes with 4 Block Pure Paraffin Wax <b>II. Electrical Specification:</b> - Equipment shall compatible with hospital power supply of 220 to 240VAC, single phase, 60Hz with grounding - With dedicated compatible Uninterruptible Power Supply (UPS) with voltage regulation function that can provide back-up power at least 30 minutes <b>III. Technical Specification:</b> <b>A. General Requirements</b>	<b>₱ 195,000.00</b>	<b>₱ 195,000.00</b>

*"PHIC Accredited Healthcare Provider"*

*"Valenzuela Medical Center...Where your health matters most"*



Address: Padrigal St., Karuhatan, Valenzuela City, 1441  
Telephone Nos: 8294-6711 to 17  
Director's Office Direct Line: 8291-4259  
Email Address: [valgen\\_hosp@yahoo.com](mailto:valgen_hosp@yahoo.com)  
Website: <https://vmc.doh.gov.ph/>

		<p>- FDA Certificate of Medical Device Notification (CMDN)</p> <p><b>B. Warranty and Preventive Maintenance Services:</b></p> <ul style="list-style-type: none"> <li>- One (1) year comprehensive warranty on parts and services for equipment</li> <li>- One (1) year warranty on parts and services for UPS/AVR</li> <li>- Certification that the 95% uptime of the product is within the warranty period and that any accumulated downtime in excess of 5% shall be added to the warranty period.</li> <li>- Preventive Maintenance and Calibration Schedule</li> <li>- Semi-Annual Preventive Maintenance Service and Annual Calibration during warranty period</li> <li>- Certification that the bidder will provide a Service Unit that the end-user can use in case the equipment or any system component will be pulled-out for repair or maintenance within the warranty period.</li> <li>- Any corrective action requiring replacement of part/component(s) shall be conducted and completed within at most 3 days, during the warranty period.</li> <li>- Service support shall cover 24 hours/day, 7 days/week during the warranty period. Remote service should be provided within 12 hours and on-site service within 24 hours must be provided during the warranty period.</li> </ul> <p><b>IV. Installation, Acceptance, Testing and Commissioning</b></p> <ul style="list-style-type: none"> <li>- Installation, Acceptance, Testing &amp; Commissioning of Equipment</li> <li>- Certification that the bidder shall be responsible for notification, transportation, delivery, installation and commissioning at no cost to the government.</li> <li>- Training of End-User (Operation) with Certificate of Training should be provided and should contain the following details: <ul style="list-style-type: none"> <li>a. Name of Trainee</li> <li>b. Modality, Brand, Model of Equipment</li> <li>c. Type of Training Conducted</li> <li>d. Inclusive Dates of the Training</li> <li>e. Name of Trainer, Date and Venue</li> </ul> </li> <li>- The unit to be delivered must show proof that its manufacturing date is not later than CY 2023 onwards</li> <li>- Electrical Safety Test</li> <li>- Provide two (2) sets of colored Technical Manual, Hardcopy and Softcopy (Flashdrive) in English Manual for all equipment including peripherals, etc. <ul style="list-style-type: none"> <li>a. User's Operational Manual</li> <li>b. Quality and Maintenance Manual</li> <li>c. Service and Technical Manual</li> </ul> </li> <li>- Hands-on Training for Biomedical, suppliers must perform an actual: <ul style="list-style-type: none"> <li>a. Operations</li> <li>b. Disassembly and Assembly</li> <li>c. Troubleshooting</li> <li>d. Recommended Maintenance as per manufacturer</li> </ul> </li> </ul> <p><b>V. Other Terms &amp; Condition:</b></p> <ul style="list-style-type: none"> <li>- Preferably packaging is made of recyclable materials</li> </ul>		
--	--	---	--	--

*"PHIC Accredited Healthcare Provider"*

*"Valenzuela Medical Center...Where your health matters most"*



Address: Padrigal St., Karuhatan, Valenzuela City, 1441  
Telephone Nos: 8294-6711 to 17  
Director's Office Direct Line: 8291-4259  
Email Address: valgen\_hosp@yahoo.com  
Website: <https://vmc.doh.gov.ph/>

			- Preferably with green eco-products specification (non-toxic materials, energy efficient, sustainable, etc.)		
3	2	unit	<p><b>ELECTRICAL STIMULATOR UNIT</b>  <b>BRAND/MODEL: ENRAF NONIUS/TENSMED S82 (NEW MODEL)</b>  <b>ORIGIN: THE NETHERLANDS</b></p> <p><b>I. Equipment Specification:</b></p> <ul style="list-style-type: none"> <li>- Intensity: 0-100 mA per channel</li> <li>- Frequency: 1-150 Hz; Total Pulse Width: 100-800</li> <li>- Stimulation Forms: Conventional / Continuous / Burst / Intermittent</li> <li>- Timer: maximum of 100 minutes</li> <li>- Power Supply: AC/DC and battery (with charger)</li> <li>- Dual-channel device with LCD screen</li> <li>- With 4 sets of cable connectors, and 8 electrodes</li> </ul> <p><b>III. Technical Specification:</b></p> <p><b>A. General Requirements</b></p> <ul style="list-style-type: none"> <li>- FDA Certificate of Medical Device Notification (CMDN)</li> </ul> <p><b>B. Warranty and Preventive Maintenance Services:</b></p> <ul style="list-style-type: none"> <li>- One (1) year comprehensive warranty on parts and services for equipment</li> <li>- One (1) year warranty on parts and services for UPS/AVR</li> <li>- Certification that the 95% uptime of the product is within the warranty period and that any accumulated downtime in excess of 5% shall be added to the warranty period.</li> <li>- Preventive Maintenance and Calibration Schedule</li> <li>- Semi-Annual Preventive Maintenance Service and Annual Calibration during warranty period</li> <li>- Certification that the bidder will provide a Service Unit that the end-user can use in case the equipment or any system component will be pulled-out for repair or maintenance within the warranty period.</li> <li>- Any corrective action requiring replacement of part/component(s) shall be conducted and completed within at most 3 days, during the warranty period.</li> <li>- Service support shall cover 24 hours/day, 7 days/week during the warranty period. Remote service should be provided within 12 hours and on-site service within 24 hours must be provided during the warranty period.</li> </ul> <p><b>IV. Installation, Acceptance, Testing and Commissioning</b></p> <ul style="list-style-type: none"> <li>- Certification that the bidder shall be responsible for notification, transportation, delivery, installation and commissioning at no cost to the government.</li> <li>- Training of End-User (Operation) with Certificate of Training should be provided and should contain the following details: <ul style="list-style-type: none"> <li>a. Name of Trainee</li> <li>b. Modality, Brand, Model of Equipment</li> <li>c. Type of Training Conducted</li> <li>d. Inclusive Dates of the Training</li> <li>e. Name of Trainer, Date and Venue</li> </ul> </li> </ul>	<b>₱ 46,110.00</b>	<b>₱ 92,220.00</b>

*"PHIC Accredited Healthcare Provider"*

*"Valenzuela Medical Center...Where your health matters most"*



Address: Padrigal St., Karuhatan, Valenzuela City, 1441  
Telephone Nos: 8294-6711 to 17  
Director's Office Direct Line: 8291-4259  
Email Address: valgen\_hosp@yahoo.com  
Website: <https://vmc.doh.gov.ph/>

		<ul style="list-style-type: none"> <li>- The unit to be delivered must show proof that its manufacturing date is not later than CY 2023 onwards</li> <li>- Electrical Safety Test</li> <li>- Provide two (2) sets of colored Technical Manual, Hardcopy and Softcopy (Flashdrive) in English Manual for all equipment including peripherals, etc. <ul style="list-style-type: none"> <li>a. User's Operational Manual</li> <li>b. Quality and Maintenance Manual</li> <li>c. Service and Technical Manual</li> </ul> </li> <li>- Hands-on Training for Biomedical, suppliers must perform an actual: <ul style="list-style-type: none"> <li>a. Operations</li> <li>b. Disassembly and Assembly</li> <li>c. Troubleshooting</li> <li>d. Recommended Maintenance as per manufacturer</li> </ul> </li> </ul> <p><b>V. Other Terms &amp; Condition:</b></p> <ul style="list-style-type: none"> <li>- Preferably packaging is made of recyclable materials</li> <li>- Preferably with green eco-products specification (non-toxic materials, energy efficient, sustainable, etc.)</li> </ul>		
		<b>TOTAL...</b>	<b>₱ 287,220.00</b>	

You are required to submit the performance security in the form and amount stipulated in the Instruction to Bidders. Failure to provide the performance security shall constitute sufficient grounds for cancellation of the award and failure of the bid security.

The Valenzuela Medical Center reserves the right to withdraw / terminate this Notice of Award in view of non-delivery on the time specified, delivery of items of poor quality / substandard and any other violation of the terms and conditions provided in the Instruction to Bidders.

Form of Performance Security	Amount of Performance Security (Equal to Percentage of the Total Contract Price)
a) Cash or Cashier's/manager (payable to <b>BUREAU OF TREASURY</b> ), check issued by a Universal or Commercial Bank	Goods and Consulting Services - Five percent (5%); Infrastructure Projects - Ten percent (10%)
b) Bank draft/guarantee or irrevocable letter of credit issued by a Universal or Commercial Bank: Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank.	
c) Surety bond callable upon demand issued by a surety or insurance company duly certified by the Insurance Commission as authorized to issue such security.	Thirty percent (30%)

#### INSTRUCTIONS:

If you have no corrections to the contents of this NOTICE OF AWARD (NOA), please submit Performance Security. The original copy of the NOA should be signed and claimed at the BAC Office, Valenzuela Medical Center. Further, please submit a certificate stating your Bank Account details.

*"PHIC Accredited Healthcare Provider"*  
*"Valenzuela Medical Center...Where your health matters most"*




Address: Padrigal St., Karuhatan, Valenzuela City, 1441  
Telephone Nos: 8294-6711 to 17  
Director's Office Direct Line: 8291-4259  
Email Address: valgen\_hosp@yahoo.com  
Website: <https://vmc.doh.gov.ph/>


Approved by:

  
**EDILBERTO V. CAVANEYRO, MD, MHA, FPCS, FPSGS**  
OIC- Medical Center Chief II

Conformed:

  
**MARLON AQUINO**  
Authorized Representative  
**FAIRBRIGHT ENTERPRISES, INC**  
Company

Date: 06-05-20

emailed  
6/5/20 

*"PHIC Accredited Healthcare Provider"*  
*"Valenzuela Medical Center...Where your health matters most"*



Address: Padrigal St., Karuhatan, Valenzuela City, 1441  
Telephone Nos: 8294-6711 to 17  
Director's Office Direct Line: 8291-4259  
Email Address: valgen\_hosp@yahoo.com  
Website: <https://vmc.doh.gov.ph/>